

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM101918\
 Data File : BM017250.D
 Acq On : 20 Oct 2018 04:48
 Operator : MJ/SJ
 Sample : J5585-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSSI-1017-S-DH-20

Quant Time: Oct 20 06:53:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 14:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	291462	20.00	ng	-0.01
21) Naphthalene-d8	10.44	136	1195749	20.00	ng	-0.01
39) Acenaphthene-d10	14.30	164	708079	20.00	ng	-0.01
64) Phenanthrene-d10	17.06	188	1505661	20.00	ng	-0.01
76) Chrysene-d12	21.25	240	1286253	20.00	ng	-0.01
87) Perylene-d12	23.46	264	1237601	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	1883090	109.27	ng	0.00
7) Phenol-d6	6.85	99	2637465	112.38	ng	-0.01
23) Nitrobenzene-d5	8.82	82	1609541	78.72	ng	-0.02
42) 2,4,6-Tribromophenol	15.80	330	1042555	108.87	ng	-0.01
45) 2-Fluorobiphenyl	12.92	172	3961691	74.44	ng	-0.01
79) Terphenyl-d14	19.69	244	4900925	85.88	ng	-0.01
Target Compounds						
10) Phenol	6.87	94	227038	8.988	ng	96
50) Dimethylphthalate	13.77	163	485627	8.890	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM101918\
Data File : BM017250.D
Acq On : 20 Oct 2018 04:48
Operator : MJ/SJ
Sample : J5585-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSSI-1017-S-DH-20

Quant Time: Oct 20 06:53:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 17 14:23:17 2018
Response via : Initial Calibration

