

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM101918\
 Data File : BM017262.D
 Acq On : 20 Oct 2018 13:14
 Operator : MJ/SJ
 Sample : J5538-16
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 NL 7

Manual Integrations
 APPROVED

Sohil
 10/22/2018 1:09:39 PM

Quant Time: Oct 22 04:46:36 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 14:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	278075	20.00	ng	-0.01
21) Naphthalene-d8	10.44	136	1172610	20.00	ng	-0.01
39) Acenaphthene-d10	14.30	164	707547	20.00	ng	-0.01
64) Phenanthrene-d10	17.06	188	1684771	20.00	ng	-0.01
76) Chrysene-d12	21.25	240	1881060	20.00	ng	-0.01
87) Perylene-d12	23.46	264	1670484	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	1850092	112.53	ng	0.00
7) Phenol-d6	6.85	99	2300486	102.74	ng	-0.01
23) Nitrobenzene-d5	8.82	82	1690414	84.31	ng	-0.02
42) 2,4,6-Tribromophenol	15.80	330	1221269	127.63	ng	-0.01
45) 2-Fluorobiphenyl	12.92	172	3877661	72.92	ng	-0.01
79) Terphenyl-d14	19.70	244	6764569	81.06	ng	0.00
Target Compounds						
10) Phenol	6.87	94	146543	6.081	ng	92
31) Naphthalene	10.49	128	119650	2.082	ng	97
32) Benzoic acid	9.70	122	58426m	5.349	ng	
50) Dimethylphthalate	13.77	163	242036	4.434	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM101918\
Data File : BM017262.D
Acq On : 20 Oct 2018 13:14
Operator : MJ/SJ
Sample : J5538-16
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
NL 7

Manual Integrations
APPROVED

Sohil
10/22/2018 1:09:39 PM

Quant Time: Oct 22 04:46:36 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 17 14:23:17 2018
Response via : Initial Calibration

