

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102122\
 Data File : BM037155.D
 Acq On : 21 Oct 2022 02:52
 Operator : CG/JU
 Sample : PB148326BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK326

Quant Time: Oct 21 03:26:23 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 20 02:11:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	5850	0.400	ng/u1	0.00
4) Naphthalene-d8	10.704	136	18212	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.534	164	10576	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.272	188	21855	0.400	ng/u1	0.00
17) Chrysene-d12	21.452	240	15776	0.400	ng/u1	0.00
23) Perylene-d12	23.826	264	14794	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.314	96	38915	5.456	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.294	152	7515	0.290	ng/u1	0.00
18) Fluoranthene-d10	19.299	212	16282	0.341	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102122\
Data File : BM037155.D
Acq On : 21 Oct 2022 02:52
Operator : CG/JU
Sample : PB148326BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK326

Quant Time: Oct 21 03:26:23 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 20 02:11:25 2022
Response via : Initial Calibration

