

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102319\
 Data File : BM023380.D
 Acq On : 24 Oct 2019 15:33
 Operator : JU
 Sample : K5322-05
 Misc : GCMS CONFIRMATION
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ESP13

Quant Time: Oct 24 16:29:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 24 13:46:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	537633	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.41	136	2027960	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.28	164	1156306	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.02	188	2166086	20.00	ng/ul	-0.01
78) Chrysene-d12	21.22	240	2163095	20.00	ng/ul	-0.01
86) Perylene-d12	23.42	264	2818268	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0d	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
58) Fluorene-d10	0.00	176	0d	0.00	ng/ul
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
71) Anthracene-d10	0.00	188	0d	0.00	ng/ul
79) Pyrene-d10	0.00	212	0d	0.00	ng/ul
90) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102319\
Data File : BM023380.D
Acq On : 24 Oct 2019 15:33
Operator : JU
Sample : K5322-05
Misc : GCMS CONFIRMATION
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ESP13

Quant Time: Oct 24 16:29:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 24 13:46:11 2019
Response via : Initial Calibration

