

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102323\
 Data File : BM042395.D
 Acq On : 23 Oct 2023 10:41
 Operator : MA/JU
 Sample : PB156324BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB156324BL

Quant Time: Oct 23 11:43:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 21 02:32:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	138494	20.000	ng	0.00
21) Naphthalene-d8	10.816	136	519325	20.000	ng	0.00
39) Acenaphthene-d10	14.639	164	246057	20.000	ng	0.00
64) Phenanthrene-d10	17.392	188	441387	20.000	ng	0.00
76) Chrysene-d12	21.568	240	299578	20.000	ng	-0.01
86) Perylene-d12	24.039	264	311312	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1517982	154.908	ng	0.00
7) Phenol-d6	7.146	99	1947621	148.309	ng	0.00
23) Nitrobenzene-d5	9.192	82	1205690	102.662	ng	0.00
42) 2,4,6-Tribromophenol	16.133	330	394725	144.175	ng	0.00
45) 2-Fluorobiphenyl	13.263	172	1982188	108.480	ng	0.00
79) Terphenyl-d14	20.003	244	2201887	130.351	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102323\
Data File : BM042395.D
Acq On : 23 Oct 2023 10:41
Operator : MA/JU
Sample : PB156324BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB156324BL

Quant Time: Oct 23 11:43:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration

