

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102323\
 Data File : BM042407.D
 Acq On : 23 Oct 2023 18:10
 Operator : MA/JU
 Sample : 04957-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MH-G

Quant Time: Oct 23 18:57:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 21 02:32:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	132450	20.000	ng	0.00
21) Naphthalene-d8	10.816	136	512828	20.000	ng	0.00
39) Acenaphthene-d10	14.639	164	265068	20.000	ng	0.00
64) Phenanthrene-d10	17.392	188	562559	20.000	ng	0.00
76) Chrysene-d12	21.574	240	582756	20.000	ng	0.00
86) Perylene-d12	24.039	264	643846	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	631432	67.377	ng	0.00
7) Phenol-d6	7.146	99	496304	39.518	ng	0.00
23) Nitrobenzene-d5	9.192	82	1101292	94.961	ng	0.00
42) 2,4,6-Tribromophenol	16.133	330	403835	137.388	ng	0.00
45) 2-Fluorobiphenyl	13.263	172	1883731	95.698	ng	0.00
79) Terphenyl-d14	20.004	244	3139605	95.547	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	18.333	149	6701	2.993	ng	Qvalue # 98
84) Bis(2-ethylhexyl)phtha...	21.439	149	6706	3.606	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102323\
Data File : BM042407.D
Acq On : 23 Oct 2023 18:10
Operator : MA/JU
Sample : 04957-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-G

Quant Time: Oct 23 18:57:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration

