

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102423\
 Data File : BM042420.D
 Acq On : 24 Oct 2023 15:27
 Operator : MA/JU
 Sample : 04978-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 NB-406-COMP-03

Quant Time: Oct 24 17:06:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 21 02:32:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	133325	20.000	ng	0.00
21) Naphthalene-d8	10.816	136	509870	20.000	ng	0.00
39) Acenaphthene-d10	14.639	164	243025	20.000	ng	0.00
64) Phenanthrene-d10	17.392	188	598402	20.000	ng	0.00
76) Chrysene-d12	21.574	240	498774	20.000	ng	0.00
86) Perylene-d12	24.039	264	152423	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1205489	127.788	ng	0.00
7) Phenol-d6	7.146	99	1361626	107.706	ng	0.00
23) Nitrobenzene-d5	9.192	82	1049059	90.982	ng	0.00
42) 2,4,6-Tribromophenol	16.133	330	534624	194.278	ng	0.00
45) 2-Fluorobiphenyl	13.263	172	1996357	110.619	ng	0.00
79) Terphenyl-d14	20.004	244	2756424	98.010	ng	0.00
Target Compounds						
74) Di-n-butylphthalate	18.333	149	4071	2.917	ng	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102423\
Data File : BM042420.D
Acq On : 24 Oct 2023 15:27
Operator : MA/JU
Sample : 04978-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
NB-406-COMP-03

Quant Time: Oct 24 17:06:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration

