




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102423\
Data File : BM042412.D
Acq On : 24 Oct 2023 10:33
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 24 20:52:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration



TIC: BM042412.D\data.ms

(70) Pentachlorophenol (C)




17.021min (+ 0.000) 0.00 ng

response 513925

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.30	63.51
264.00	61.70	62.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102423\
Data File : BM042412.D
Acq On : 24 Oct 2023 10:33
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 24 20:52:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration



TIC: BM042412.D\data.ms

(77) Benzidine**19.656min (-0.000) 239771.26 ng****response 2455424**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.13
183.00	11.50	11.42
0.00	0.00	0.00