


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_M\DATA\BM102518\  
Data File : BM017341.D  
Acq On : 25 Oct 2018 13:21  
Operator : MJ/SJ  
Sample : DFTPP56  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM102418MA.M  
Title : SVOA CALIBRATION  
Last Update : Fri Oct 26 15:57:14 2018



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2461

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 32.8      | 178090  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 3577    | PASS             |
| 69          | 198          | 0.00         | 100          | 35.1      | 190357  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 1032    | PASS             |
| 127         | 198          | 10           | 80           | 43.5      | 236181  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 3093    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 542506  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 36488   | PASS             |
| 275         | 198          | 10           | 60           | 27.6      | 149565  | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 18497   | PASS             |
| 441         | 198          | 0.01         | 100          | 11.8      | 64056   | PASS             |
| 442         | 198          | 50           | 100          | 79.7      | 432490  | PASS             |
| 443         | 442          | 15           | 24           | 19.3      | 83560   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 676.0     |
| 38.00  | 1849.0    |
| 39.00  | 11562.0   |
| 40.00  | 2549.0    |
| 41.10  | 321.0     |
| 44.00  | 6593.0    |
| 44.90  | 488.0     |
| 49.00  | 1190.0    |
| 50.10  | 39232.0   |
| 51.00  | 158656.0  |
| 52.00  | 7964.0    |
| 55.00  | 847.0     |
| 56.00  | 4809.0    |
| 57.00  | 10234.0   |
| 58.00  | 595.0     |
| 61.00  | 2333.0    |
| 62.10  | 2515.0    |
| 63.10  | 6334.0    |
| 64.00  | 1198.0    |
| 65.10  | 3755.0    |
| 68.00  | 3448.0    |
| 69.00  | 161216.0  |
| 69.90  | 1007.0    |
| 73.10  | 2527.0    |
| 74.10  | 15467.0   |
| 75.00  | 25224.0   |
| 76.10  | 8119.0    |
| 77.10  | 180480.0  |
| 78.10  | 12508.0   |
| 79.00  | 13283.0   |
| 80.00  | 8530.0    |
| 81.00  | 13135.0   |
| 82.00  | 3323.0    |
| 83.00  | 2904.0    |
| 83.90  | 321.0     |
| 85.00  | 2381.0    |
| 86.00  | 4515.0    |
| 87.00  | 1773.0    |
| 88.00  | 693.0     |
| 89.10  | 461.0     |
| 91.00  | 3316.0    |
| 92.00  | 4250.0    |
| 93.00  | 23024.0   |
| 93.90  | 1467.0    |
| 95.00  | 411.0     |
| 96.10  | 1565.0    |
| 96.80  | 327.0     |
| 97.20  | 352.0     |
| 98.00  | 16232.0   |
| 99.00  | 14458.0   |
| 100.00 | 1351.0    |
| 101.00 | 7514.0    |
| 102.10 | 491.0     |
| 103.00 | 2671.0    |
| 104.00 | 5093.0    |
| 105.00 | 4260.0    |
| 106.00 | 1499.0    |
| 107.00 | 54152.0   |
| 108.00 | 7679.0    |
| 109.00 | 1311.0    |
| 110.00 | 99096.0   |
| 111.00 | 14860.0   |
| 112.00 | 2298.0    |
| 113.00 | 589.0     |
| 116.00 | 2839.0    |
| 117.00 | 43872.0   |
| 118.10 | 3188.0    |
| 119.00 | 551.0     |
| 119.90 | 982.0     |
| 121.10 | 302.0     |
| 122.00 | 4090.0    |
| 123.00 | 6382.0    |
| 124.00 | 2818.0    |
| 125.00 | 2797.0    |
| 126.10 | 1311.0    |
| 127.00 | 193024.0  |
| 128.00 | 14357.0   |
| 129.00 | 70112.0   |
| 130.00 | 6233.0    |
| 130.90 | 1158.0    |
| 132.20 | 642.0     |
| 133.00 | 879.0     |
| 134.00 | 2215.0    |
| 135.00 | 4885.0    |
| 136.00 | 1865.0    |
| 137.00 | 3188.0    |
| 139.10 | 923.0     |
| 140.00 | 1264.0    |
| 141.00 | 9021.0    |
| 142.00 | 3006.0    |
| 143.00 | 2157.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 144.00 | 629.0    |
| 144.90 | 666.0    |
| 146.00 | 1741.0   |
| 147.00 | 5442.0   |
| 148.00 | 10413.0  |
| 149.00 | 2565.0   |
| 149.90 | 810.0    |
| 151.10 | 1338.0   |
| 152.00 | 645.0    |
| 153.00 | 2702.0   |
| 154.00 | 1959.0   |
| 155.00 | 5390.0   |
| 156.00 | 8125.0   |
| 157.00 | 1140.0   |
| 158.00 | 2084.0   |
| 159.00 | 1441.0   |
| 160.00 | 3381.0   |
| 161.00 | 4736.0   |
| 162.00 | 1342.0   |
| 163.00 | 532.0    |
| 164.00 | 647.0    |
| 165.00 | 4155.0   |
| 166.00 | 3107.0   |
| 167.00 | 20704.0  |
| 168.00 | 12158.0  |
| 169.00 | 1549.0   |
| 169.90 | 783.0    |
| 171.10 | 777.0    |
| 172.00 | 1982.0   |
| 173.00 | 2249.0   |
| 174.00 | 3641.0   |
| 175.00 | 7240.0   |
| 176.00 | 2215.0   |
| 177.00 | 3969.0   |
| 178.10 | 1000.0   |
| 179.00 | 13775.0  |
| 180.00 | 10505.0  |
| 181.00 | 4777.0   |
| 182.00 | 728.0    |
| 183.10 | 362.0    |
| 184.00 | 1399.0   |
| 185.00 | 7403.0   |
| 186.00 | 56664.0  |
| 187.00 | 15923.0  |
| 188.10 | 2200.0   |
| 189.00 | 4459.0   |
| 190.00 | 572.0    |
| 191.00 | 1832.0   |
| 192.00 | 4969.0   |
| 193.00 | 5092.0   |
| 194.10 | 1042.0   |
| 195.10 | 714.0    |
| 196.00 | 13043.0  |
| 197.10 | 2532.0   |
| 198.00 | 410880.0 |
| 199.00 | 30072.0  |
| 200.00 | 2510.0   |
| 201.40 | 1529.0   |
| 203.00 | 2708.0   |
| 204.10 | 14265.0  |
| 205.00 | 22136.0  |
| 206.00 | 104144.0 |
| 207.10 | 16536.0  |
| 208.00 | 3179.0   |
| 209.00 | 2203.0   |
| 210.20 | 1513.0   |
| 211.00 | 4061.0   |
| 214.90 | 1045.0   |
| 216.10 | 1631.0   |
| 217.00 | 25536.0  |
| 218.00 | 3082.0   |
| 219.10 | 355.0    |
| 221.00 | 17064.0  |
| 223.00 | 5664.0   |
| 224.00 | 55944.0  |
| 225.00 | 14603.0  |
| 226.10 | 2335.0   |
| 227.00 | 23408.0  |
| 228.10 | 3038.0   |
| 229.00 | 4794.0   |
| 230.10 | 855.0    |
| 231.00 | 2522.0   |
| 231.90 | 580.0    |
| 233.10 | 386.0    |
| 233.90 | 1890.0   |
| 235.00 | 1722.0   |
| 236.00 | 1092.0   |
| 237.00 | 2335.0   |
| 239.00 | 1187.0   |
| 239.90 | 873.0    |
| 241.10 | 1294.0   |
| 242.00 | 3067.0   |
| 243.10 | 3115.0   |
| 244.10 | 44552.0  |
| 245.10 | 5090.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 246.00 | 8596.0   |
| 247.00 | 1704.0   |
| 247.80 | 424.0    |
| 248.90 | 1808.0   |
| 249.90 | 487.0    |
| 250.90 | 366.0    |
| 252.20 | 638.0    |
| 253.10 | 930.0    |
| 254.20 | 2126.0   |
| 255.00 | 227008.0 |
| 256.00 | 35384.0  |
| 257.00 | 3105.0   |
| 258.00 | 13141.0  |
| 258.90 | 2527.0   |
| 259.90 | 397.0    |
| 260.90 | 627.0    |
| 265.00 | 5441.0   |
| 266.00 | 1067.0   |
| 267.20 | 325.0    |
| 271.00 | 570.0    |
| 272.10 | 935.0    |
| 273.00 | 5721.0   |
| 274.00 | 16872.0  |
| 275.00 | 103288.0 |
| 276.00 | 16141.0  |
| 277.00 | 8752.0   |
| 278.00 | 1944.0   |
| 279.00 | 545.0    |
| 281.10 | 873.0    |
| 281.90 | 460.0    |
| 282.90 | 1051.0   |
| 284.10 | 491.0    |
| 285.00 | 1537.0   |
| 286.00 | 374.0    |
| 290.00 | 347.0    |
| 292.90 | 1874.0   |
| 294.00 | 790.0    |
| 295.20 | 574.0    |
| 296.00 | 25704.0  |
| 297.00 | 4458.0   |
| 298.20 | 377.0    |
| 301.00 | 549.0    |
| 302.00 | 684.0    |
| 303.00 | 3214.0   |
| 303.90 | 910.0    |
| 308.10 | 370.0    |
| 308.90 | 345.0    |
| 314.00 | 1376.0   |
| 315.00 | 3082.0   |
| 316.10 | 2152.0   |
| 321.00 | 918.0    |
| 322.10 | 519.0    |
| 323.00 | 9560.0   |
| 324.00 | 1806.0   |
| 327.10 | 1307.0   |
| 327.90 | 1382.0   |
| 331.90 | 692.0    |
| 333.00 | 832.0    |
| 334.10 | 5533.0   |
| 335.00 | 1773.0   |
| 341.00 | 985.0    |
| 342.10 | 396.0    |
| 346.00 | 1472.0   |
| 347.00 | 510.0    |
| 352.00 | 3143.0   |
| 353.00 | 2383.0   |
| 354.00 | 2943.0   |
| 355.00 | 851.0    |
| 365.00 | 11405.0  |
| 365.90 | 2153.0   |
| 369.90 | 326.0    |
| 371.00 | 753.0    |
| 372.00 | 4969.0   |
| 373.00 | 1293.0   |
| 383.00 | 1176.0   |
| 384.10 | 533.0    |
| 390.00 | 673.0    |
| 392.00 | 329.0    |
| 400.80 | 347.0    |
| 402.00 | 1596.0   |
| 403.00 | 3152.0   |
| 404.00 | 1232.0   |
| 421.00 | 1858.0   |
| 422.10 | 2352.0   |
| 423.00 | 14391.0  |
| 424.10 | 4938.0   |
| 425.00 | 936.0    |
| 441.10 | 37504.0  |
| 442.10 | 252032.0 |
| 443.10 | 48344.0  |
| 444.00 | 4115.0   |
| 445.20 | 343.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 305.0     |
| 37.00  | 1297.0    |
| 38.10  | 3014.0    |
| 39.10  | 16074.0   |
| 40.00  | 2076.0    |
| 41.00  | 795.0     |
| 44.00  | 5762.0    |
| 49.00  | 1122.0    |
| 50.10  | 55264.0   |
| 51.00  | 211712.0  |
| 52.00  | 10708.0   |
| 53.00  | 883.0     |
| 55.00  | 960.0     |
| 56.00  | 6494.0    |
| 57.00  | 13986.0   |
| 58.00  | 715.0     |
| 61.00  | 2911.0    |
| 62.00  | 2875.0    |
| 63.00  | 8667.0    |
| 64.00  | 1123.0    |
| 65.00  | 4666.0    |
| 67.20  | 637.0     |
| 68.10  | 3702.0    |
| 69.00  | 230208.0  |
| 70.00  | 1086.0    |
| 73.00  | 2703.0    |
| 74.00  | 23328.0   |
| 75.00  | 36928.0   |
| 76.10  | 12144.0   |
| 77.10  | 252928.0  |
| 78.00  | 17592.0   |
| 79.00  | 18584.0   |
| 80.00  | 12596.0   |
| 81.00  | 18984.0   |
| 82.00  | 4258.0    |
| 83.10  | 4188.0    |
| 84.00  | 427.0     |
| 85.00  | 3876.0    |
| 86.00  | 5339.0    |
| 87.00  | 2239.0    |
| 88.00  | 1030.0    |
| 88.90  | 656.0     |
| 91.00  | 3854.0    |
| 92.00  | 4119.0    |
| 93.00  | 28096.0   |
| 94.00  | 2001.0    |
| 95.10  | 589.0     |
| 96.00  | 1576.0    |
| 97.20  | 672.0     |
| 98.00  | 22440.0   |
| 99.00  | 18960.0   |
| 100.10 | 1747.0    |
| 101.00 | 11312.0   |
| 101.90 | 463.0     |
| 103.00 | 2629.0    |
| 104.00 | 6158.0    |
| 105.00 | 6392.0    |
| 106.00 | 1874.0    |
| 107.00 | 78336.0   |
| 108.00 | 12242.0   |
| 109.10 | 2583.0    |
| 110.00 | 144896.0  |
| 111.00 | 22840.0   |
| 112.00 | 2599.0    |
| 113.00 | 786.0     |
| 115.10 | 387.0     |
| 116.00 | 4821.0    |
| 117.00 | 61208.0   |
| 118.00 | 5293.0    |
| 119.00 | 662.0     |
| 120.00 | 855.0     |
| 121.10 | 374.0     |
| 122.00 | 5933.0    |
| 123.00 | 7226.0    |
| 124.00 | 3810.0    |
| 125.00 | 3700.0    |
| 127.00 | 277760.0  |
| 128.00 | 21288.0   |
| 129.00 | 108696.0  |
| 130.00 | 8546.0    |
| 131.00 | 1801.0    |
| 132.10 | 692.0     |
| 133.00 | 581.0     |
| 134.00 | 2981.0    |
| 135.00 | 8784.0    |
| 136.00 | 2768.0    |
| 137.00 | 5057.0    |
| 137.90 | 975.0     |
| 139.00 | 688.0     |
| 140.00 | 1365.0    |
| 141.00 | 14524.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 142.00 | 4641.0   |
| 143.00 | 2926.0   |
| 143.90 | 1064.0   |
| 144.90 | 765.0    |
| 146.10 | 2972.0   |
| 147.00 | 7501.0   |
| 148.00 | 16096.0  |
| 149.00 | 3177.0   |
| 150.00 | 1236.0   |
| 151.10 | 1759.0   |
| 151.90 | 978.0    |
| 153.00 | 4392.0   |
| 154.00 | 3476.0   |
| 155.00 | 7073.0   |
| 156.00 | 11711.0  |
| 157.10 | 1673.0   |
| 158.00 | 2495.0   |
| 159.00 | 1796.0   |
| 159.90 | 4089.0   |
| 161.00 | 6235.0   |
| 162.00 | 1850.0   |
| 163.10 | 663.0    |
| 164.10 | 617.0    |
| 165.00 | 5288.0   |
| 166.10 | 5140.0   |
| 167.00 | 28200.0  |
| 168.00 | 15052.0  |
| 169.00 | 2551.0   |
| 169.80 | 527.0    |
| 170.30 | 802.0    |
| 170.80 | 1280.0   |
| 172.00 | 2494.0   |
| 173.10 | 3001.0   |
| 174.00 | 6825.0   |
| 175.00 | 11566.0  |
| 176.00 | 3361.0   |
| 177.00 | 5335.0   |
| 177.90 | 1762.0   |
| 179.00 | 22328.0  |
| 180.00 | 15960.0  |
| 181.00 | 6960.0   |
| 182.00 | 988.0    |
| 183.00 | 1043.0   |
| 184.00 | 1909.0   |
| 185.00 | 10052.0  |
| 186.00 | 84840.0  |
| 187.00 | 23312.0  |
| 188.00 | 2653.0   |
| 189.00 | 5962.0   |
| 190.00 | 1092.0   |
| 191.00 | 2864.0   |
| 191.90 | 6916.0   |
| 193.00 | 7466.0   |
| 194.00 | 1947.0   |
| 195.10 | 1055.0   |
| 196.00 | 19424.0  |
| 197.10 | 3419.0   |
| 198.00 | 631552.0 |
| 199.00 | 42408.0  |
| 200.00 | 3454.0   |
| 201.60 | 3636.0   |
| 203.00 | 4296.0   |
| 204.00 | 21728.0  |
| 205.10 | 36632.0  |
| 206.10 | 161216.0 |
| 207.10 | 24216.0  |
| 208.00 | 6249.0   |
| 209.10 | 1967.0   |
| 210.00 | 2436.0   |
| 211.00 | 6198.0   |
| 211.80 | 1442.0   |
| 213.10 | 635.0    |
| 213.80 | 300.0    |
| 215.00 | 1601.0   |
| 216.00 | 3369.0   |
| 217.00 | 42424.0  |
| 218.00 | 5206.0   |
| 219.00 | 589.0    |
| 219.90 | 715.0    |
| 221.10 | 28296.0  |
| 221.80 | 6871.0   |
| 223.00 | 9462.0   |
| 224.10 | 89944.0  |
| 225.00 | 25096.0  |
| 226.00 | 2698.0   |
| 227.00 | 36584.0  |
| 228.00 | 5194.0   |
| 229.00 | 8006.0   |
| 230.00 | 1141.0   |
| 231.00 | 3034.0   |
| 232.10 | 573.0    |
| 233.00 | 524.0    |
| 234.00 | 2575.0   |
| 235.00 | 2252.0   |
| 236.00 | 2064.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 237.10 | 4142.0   |
| 238.00 | 324.0    |
| 239.00 | 1473.0   |
| 239.90 | 1329.0   |
| 241.00 | 2039.0   |
| 242.00 | 4668.0   |
| 243.00 | 4714.0   |
| 244.10 | 67288.0  |
| 245.00 | 9241.0   |
| 246.00 | 14467.0  |
| 247.00 | 2594.0   |
| 247.90 | 797.0    |
| 249.00 | 2651.0   |
| 249.90 | 694.0    |
| 250.90 | 482.0    |
| 252.10 | 506.0    |
| 253.10 | 1921.0   |
| 254.20 | 2539.0   |
| 255.00 | 376768.0 |
| 256.00 | 56504.0  |
| 257.00 | 4036.0   |
| 258.00 | 20472.0  |
| 259.00 | 3109.0   |
| 260.00 | 691.0    |
| 261.00 | 606.0    |
| 264.00 | 615.0    |
| 265.00 | 8169.0   |
| 266.00 | 1336.0   |
| 268.10 | 343.0    |
| 270.10 | 499.0    |
| 271.00 | 846.0    |
| 272.00 | 1411.0   |
| 273.00 | 11481.0  |
| 274.00 | 28872.0  |
| 275.00 | 172352.0 |
| 276.00 | 22528.0  |
| 277.00 | 14208.0  |
| 278.00 | 3037.0   |
| 279.10 | 529.0    |
| 281.00 | 784.0    |
| 282.10 | 613.0    |
| 283.10 | 1272.0   |
| 284.00 | 1203.0   |
| 285.10 | 2666.0   |
| 286.10 | 529.0    |
| 289.10 | 621.0    |
| 290.10 | 389.0    |
| 290.90 | 328.0    |
| 292.10 | 457.0    |
| 293.00 | 3429.0   |
| 293.90 | 839.0    |
| 295.10 | 1075.0   |
| 296.00 | 44144.0  |
| 297.00 | 6360.0   |
| 298.00 | 763.0    |
| 300.90 | 591.0    |
| 302.10 | 624.0    |
| 303.10 | 6155.0   |
| 304.10 | 1254.0   |
| 304.90 | 341.0    |
| 308.00 | 771.0    |
| 309.00 | 442.0    |
| 310.00 | 576.0    |
| 313.00 | 510.0    |
| 314.10 | 2384.0   |
| 315.00 | 4937.0   |
| 316.00 | 2820.0   |
| 317.00 | 816.0    |
| 321.00 | 1543.0   |
| 322.00 | 718.0    |
| 323.10 | 16936.0  |
| 324.10 | 3048.0   |
| 326.00 | 539.0    |
| 327.00 | 2747.0   |
| 328.00 | 1641.0   |
| 329.10 | 389.0    |
| 331.90 | 845.0    |
| 333.10 | 1260.0   |
| 334.10 | 9829.0   |
| 335.10 | 2510.0   |
| 336.10 | 314.0    |
| 339.10 | 303.0    |
| 341.10 | 1910.0   |
| 342.00 | 662.0    |
| 346.00 | 3193.0   |
| 347.10 | 755.0    |
| 351.10 | 322.0    |
| 352.10 | 4608.0   |
| 353.10 | 3522.0   |
| 354.00 | 5455.0   |
| 355.10 | 1388.0   |
| 359.10 | 392.0    |
| 365.00 | 20848.0  |
| 366.00 | 3097.0   |
| 370.00 | 550.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 371.10 | 1091.0   |
| 372.10 | 8442.0   |
| 373.00 | 2165.0   |
| 383.00 | 2289.0   |
| 384.00 | 552.0    |
| 390.10 | 1076.0   |
| 391.00 | 795.0    |
| 392.20 | 640.0    |
| 401.10 | 515.0    |
| 402.00 | 3795.0   |
| 403.00 | 4537.0   |
| 404.00 | 1897.0   |
| 404.90 | 462.0    |
| 421.10 | 3603.0   |
| 422.00 | 3515.0   |
| 423.10 | 26768.0  |
| 424.10 | 8910.0   |
| 425.10 | 1021.0   |
| 440.10 | 478.0    |
| 441.10 | 71424.0  |
| 442.10 | 472256.0 |
| 443.10 | 93968.0  |
| 444.10 | 8270.0   |
| 445.10 | 424.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 656.0     |
| 38.00  | 1540.0    |
| 39.10  | 12389.0   |
| 40.10  | 2274.0    |
| 41.00  | 420.0     |
| 43.10  | 429.0     |
| 44.00  | 6584.0    |
| 45.00  | 319.0     |
| 49.10  | 987.0     |
| 50.00  | 41568.0   |
| 51.00  | 163904.0  |
| 52.10  | 8572.0    |
| 53.00  | 396.0     |
| 55.00  | 925.0     |
| 56.00  | 4929.0    |
| 57.00  | 11005.0   |
| 58.10  | 347.0     |
| 60.90  | 2309.0    |
| 62.10  | 2532.0    |
| 63.00  | 7765.0    |
| 64.00  | 1041.0    |
| 65.00  | 3059.0    |
| 65.90  | 494.0     |
| 68.10  | 3582.0    |
| 69.00  | 179648.0  |
| 70.00  | 1004.0    |
| 73.00  | 2327.0    |
| 74.00  | 19760.0   |
| 75.00  | 29112.0   |
| 76.10  | 9168.0    |
| 77.10  | 201920.0  |
| 78.10  | 14343.0   |
| 79.00  | 14432.0   |
| 80.00  | 10831.0   |
| 81.00  | 14711.0   |
| 82.00  | 3934.0    |
| 83.00  | 3041.0    |
| 83.90  | 514.0     |
| 85.00  | 2402.0    |
| 86.00  | 4377.0    |
| 87.00  | 2195.0    |
| 88.00  | 867.0     |
| 89.00  | 431.0     |
| 91.00  | 3443.0    |
| 92.00  | 3739.0    |
| 93.00  | 23120.0   |
| 94.00  | 1388.0    |
| 95.00  | 408.0     |
| 96.10  | 1516.0    |
| 98.00  | 17984.0   |
| 99.00  | 16400.0   |
| 100.00 | 1076.0    |
| 101.00 | 8320.0    |
| 101.90 | 415.0     |
| 102.90 | 2565.0    |
| 104.00 | 5544.0    |
| 105.00 | 6164.0    |
| 106.10 | 1539.0    |
| 107.00 | 64272.0   |
| 108.00 | 10052.0   |
| 109.10 | 1698.0    |
| 110.00 | 120136.0  |
| 111.00 | 19368.0   |
| 112.00 | 2241.0    |
| 113.10 | 794.0     |
| 115.00 | 510.0     |
| 116.00 | 3649.0    |
| 117.00 | 50408.0   |
| 118.00 | 3629.0    |
| 119.00 | 714.0     |
| 120.00 | 1206.0    |
| 121.10 | 429.0     |
| 122.00 | 4454.0    |
| 123.00 | 6534.0    |
| 123.90 | 2780.0    |
| 125.00 | 2722.0    |
| 126.20 | 1866.0    |
| 127.00 | 237760.0  |
| 128.00 | 18192.0   |
| 129.00 | 91096.0   |
| 130.00 | 8514.0    |
| 131.00 | 1749.0    |
| 131.90 | 1195.0    |
| 133.00 | 651.0     |
| 134.00 | 2543.0    |
| 135.00 | 8292.0    |
| 136.00 | 3424.0    |
| 137.00 | 3925.0    |
| 137.80 | 838.0     |
| 138.10 | 778.0     |
| 138.80 | 593.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 139.80 | 1087.0   |
| 141.00 | 12209.0  |
| 142.00 | 3889.0   |
| 143.00 | 2730.0   |
| 144.00 | 875.0    |
| 144.90 | 593.0    |
| 146.00 | 2388.0   |
| 147.00 | 6404.0   |
| 148.00 | 13389.0  |
| 149.00 | 3179.0   |
| 150.00 | 1078.0   |
| 151.10 | 1460.0   |
| 151.80 | 1174.0   |
| 153.00 | 3479.0   |
| 154.00 | 2662.0   |
| 155.00 | 5772.0   |
| 156.00 | 10961.0  |
| 157.00 | 1705.0   |
| 158.00 | 2536.0   |
| 159.00 | 1634.0   |
| 160.00 | 4216.0   |
| 161.00 | 6056.0   |
| 162.00 | 1350.0   |
| 163.00 | 814.0    |
| 164.00 | 747.0    |
| 165.00 | 4430.0   |
| 166.00 | 4003.0   |
| 167.00 | 25640.0  |
| 168.00 | 12229.0  |
| 169.00 | 1566.0   |
| 170.00 | 699.0    |
| 170.30 | 729.0    |
| 171.00 | 1269.0   |
| 171.90 | 1880.0   |
| 173.10 | 3165.0   |
| 174.00 | 5031.0   |
| 175.10 | 10267.0  |
| 176.10 | 2404.0   |
| 177.00 | 5346.0   |
| 178.00 | 1648.0   |
| 179.00 | 20024.0  |
| 180.00 | 13495.0  |
| 181.00 | 6269.0   |
| 182.00 | 1350.0   |
| 183.00 | 603.0    |
| 183.90 | 1063.0   |
| 184.20 | 991.0    |
| 185.00 | 9287.0   |
| 186.00 | 73560.0  |
| 187.00 | 21728.0  |
| 188.00 | 2370.0   |
| 189.00 | 4613.0   |
| 190.00 | 1043.0   |
| 191.10 | 2404.0   |
| 192.00 | 6722.0   |
| 193.00 | 6672.0   |
| 193.90 | 1689.0   |
| 195.00 | 1431.0   |
| 196.00 | 17368.0  |
| 197.10 | 3329.0   |
| 198.00 | 585088.0 |
| 199.00 | 36984.0  |
| 200.00 | 2534.0   |
| 201.30 | 2111.0   |
| 202.10 | 1050.0   |
| 203.00 | 4488.0   |
| 204.00 | 18624.0  |
| 205.00 | 33712.0  |
| 206.10 | 149952.0 |
| 207.10 | 20912.0  |
| 208.00 | 5516.0   |
| 209.00 | 2479.0   |
| 209.90 | 1745.0   |
| 211.10 | 5600.0   |
| 213.00 | 327.0    |
| 215.00 | 1567.0   |
| 216.00 | 4051.0   |
| 217.00 | 42360.0  |
| 218.10 | 4938.0   |
| 219.00 | 308.0    |
| 220.30 | 701.0    |
| 221.00 | 25288.0  |
| 221.90 | 4688.0   |
| 223.00 | 9670.0   |
| 224.10 | 83240.0  |
| 225.00 | 21040.0  |
| 226.00 | 2449.0   |
| 227.00 | 36400.0  |
| 228.00 | 4885.0   |
| 229.00 | 7916.0   |
| 230.00 | 1075.0   |
| 231.00 | 3383.0   |
| 232.10 | 575.0    |
| 233.00 | 593.0    |
| 234.10 | 2301.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 235.00 | 2813.0   |
| 236.00 | 2027.0   |
| 237.00 | 3721.0   |
| 237.90 | 471.0    |
| 239.00 | 1398.0   |
| 240.10 | 768.0    |
| 240.90 | 1941.0   |
| 242.00 | 4859.0   |
| 243.10 | 4725.0   |
| 244.10 | 66944.0  |
| 245.10 | 8116.0   |
| 246.00 | 12752.0  |
| 247.00 | 3424.0   |
| 248.00 | 627.0    |
| 249.00 | 2637.0   |
| 250.00 | 427.0    |
| 250.90 | 683.0    |
| 252.00 | 650.0    |
| 253.10 | 1547.0   |
| 254.10 | 2975.0   |
| 255.00 | 361600.0 |
| 256.00 | 51032.0  |
| 257.00 | 4178.0   |
| 258.00 | 20264.0  |
| 259.00 | 2218.0   |
| 260.10 | 791.0    |
| 261.00 | 634.0    |
| 264.00 | 510.0    |
| 265.00 | 8430.0   |
| 266.10 | 1098.0   |
| 269.90 | 406.0    |
| 271.00 | 568.0    |
| 272.10 | 704.0    |
| 273.00 | 11376.0  |
| 274.00 | 28728.0  |
| 275.00 | 173056.0 |
| 276.00 | 23504.0  |
| 277.00 | 13916.0  |
| 278.10 | 2264.0   |
| 279.10 | 564.0    |
| 281.20 | 905.0    |
| 281.90 | 540.0    |
| 283.00 | 1587.0   |
| 284.00 | 1293.0   |
| 285.00 | 3340.0   |
| 286.00 | 385.0    |
| 288.90 | 600.0    |
| 290.10 | 355.0    |
| 290.90 | 484.0    |
| 292.00 | 950.0    |
| 292.90 | 3194.0   |
| 294.00 | 695.0    |
| 295.10 | 835.0    |
| 296.00 | 46680.0  |
| 297.00 | 6263.0   |
| 298.00 | 397.0    |
| 301.10 | 734.0    |
| 302.00 | 953.0    |
| 303.00 | 5305.0   |
| 304.00 | 1232.0   |
| 308.00 | 1061.0   |
| 309.00 | 538.0    |
| 310.10 | 564.0    |
| 313.00 | 577.0    |
| 314.00 | 2274.0   |
| 315.00 | 5701.0   |
| 316.00 | 2986.0   |
| 317.00 | 814.0    |
| 321.10 | 1325.0   |
| 322.10 | 736.0    |
| 323.10 | 16688.0  |
| 324.00 | 3024.0   |
| 326.90 | 3486.0   |
| 328.10 | 1659.0   |
| 329.00 | 329.0    |
| 332.00 | 1305.0   |
| 333.00 | 1743.0   |
| 334.00 | 10850.0  |
| 335.00 | 2826.0   |
| 341.00 | 1968.0   |
| 342.10 | 597.0    |
| 346.00 | 3271.0   |
| 346.90 | 502.0    |
| 351.00 | 360.0    |
| 352.10 | 5112.0   |
| 353.10 | 4179.0   |
| 354.10 | 6637.0   |
| 355.10 | 1106.0   |
| 359.00 | 453.0    |
| 365.00 | 23240.0  |
| 366.00 | 3143.0   |
| 370.10 | 319.0    |
| 371.10 | 1559.0   |
| 372.00 | 10126.0  |
| 373.00 | 2751.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

|        |          |
|--------|----------|
| 383.10 | 2516.0   |
| 384.00 | 1002.0   |
| 390.00 | 1331.0   |
| 391.00 | 869.0    |
| 392.00 | 410.0    |
| 401.00 | 734.0    |
| 402.00 | 4893.0   |
| 403.10 | 6224.0   |
| 404.10 | 2324.0   |
| 421.00 | 5176.0   |
| 422.00 | 4637.0   |
| 423.10 | 33000.0  |
| 424.10 | 7755.0   |
| 425.00 | 858.0    |
| 440.20 | 699.0    |
| 441.10 | 83240.0  |
| 442.10 | 573184.0 |
| 443.10 | 108368.0 |
| 444.10 | 8454.0   |
| 445.00 | 758.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 2180.0    |
| 44.00  | 6383.0    |
| 50.00  | 322.0     |
| 55.00  | 389.0     |
| 63.00  | 437.0     |
| 73.00  | 1355.0    |
| 77.00  | 334.0     |
| 83.80  | 375.0     |
| 115.00 | 416.0     |
| 126.00 | 418.0     |
| 126.20 | 417.0     |
| 132.80 | 323.0     |
| 139.00 | 441.0     |
| 146.90 | 379.0     |
| 154.10 | 399.0     |
| 166.00 | 634.0     |
| 167.10 | 3721.0    |
| 167.90 | 403.0     |
| 191.10 | 345.0     |
| 193.00 | 533.0     |
| 193.90 | 315.0     |
| 207.00 | 3007.0    |
| 208.00 | 834.0     |
| 208.90 | 497.0     |
| 276.10 | 800.0     |
| 281.10 | 1037.0    |
| 282.10 | 365.0     |
| 283.10 | 301.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP56