


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102523\
Data File : BM042432.D
Acq On : 25 Oct 2023 11:30
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Oct 21 02:32:53 2023






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.4	111550	PASS
68	69	0.00	2	1.5	1677	PASS
69	198	0.00	100	39.0	110403	PASS
70	69	0.00	2	0.6	676	PASS
127	198	10	80	42.1	119080	PASS
197	198	0.00	2	0.4	1273	PASS
198	198	100	100	100.0	283008	PASS
199	198	5	9	6.8	19200	PASS
275	198	10	60	25.3	71600	PASS
365	198	1	100	3.0	8430	PASS
441	198	0.01	100	14.8	41813	PASS
442	442	50	100	100.0	269653	PASS
443	442	15	24	19.4	52373	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102523\
Data File : BM042432.D
Acq On : 25 Oct 2023 11:30
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Oct 25 20:30:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration



TIC: BM042432.D\data.ms

(70) Pentachlorophenol (C)

17.015min (-0.006) 16983.07 ng




response 373793

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.30	63.22
264.00	61.70	63.22
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102523\
Data File : BM042432.D
Acq On : 25 Oct 2023 11:30
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Oct 25 20:30:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 21 02:32:53 2023
Response via : Initial Calibration



TIC: BM042432.D\data.ms

(77) Benzidine

19.650min (-0.006) 26820.11 ng

response 1840742

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.07
183.00	11.50	11.43
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/25/2023	BNA_M	<u>BM042432.D</u>
Compound Name	Response	Retention Time
DDT	1106686	20.844
DDD	18483	20.45
DDE	642	19.892
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
19125	1125811	1.70

Instrument :
BNA_M
ClientSampleId :
DFTPP