

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102618\
 Data File : BM017364.D
 Acq On : 27 Oct 2018 00:47
 Operator : MJ/SJ
 Sample : J5686-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 102418-SIDI-F-D-B

Quant Time: Oct 27 03:00:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 17 14:23:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	258792	20.00	ng	-0.02
21) Naphthalene-d8	10.43	136	1108708	20.00	ng	-0.02
39) Acenaphthene-d10	14.29	164	717133	20.00	ng	-0.02
64) Phenanthrene-d10	17.04	188	1784392	20.00	ng	-0.02
76) Chrysene-d12	21.24	240	1995834	20.00	ng	-0.02
87) Perylene-d12	23.45	264	1836647	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	657475	42.97	ng	-0.02
7) Phenol-d6	6.83	99	558551	26.80	ng	-0.02
23) Nitrobenzene-d5	8.81	82	2039248	107.57	ng	-0.02
42) 2,4,6-Tribromophenol	15.79	330	1028618	106.06	ng	-0.02
45) 2-Fluorobiphenyl	12.91	172	5060300	93.89	ng	-0.02
79) Terphenyl-d14	19.69	244	9005656	101.71	ng	-0.01
Target Compounds						
50) Dimethylphthalate	13.76	163	208334	3.765	ng	Qvalue 100
84) Bis(2-ethylhexyl)phthalate	21.17	149	4806	5.374	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102618\
Data File : BM017364.D
Acq On : 27 Oct 2018 00:47
Operator : MJ/SJ
Sample : J5686-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
102418-SIDI-F-D-B

Quant Time: Oct 27 03:00:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 17 14:23:17 2018
Response via : Initial Calibration

