

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102819\
 Data File : BM023427.D
 Acq On : 28 Oct 2019 19:47
 Operator : JU
 Sample : PB124084BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK84

Quant Time: Oct 29 11:49:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 25 07:44:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	202529	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.40	136	818004	20.00	ng/ul	-0.02
36) Acenaphthene-d10	14.27	164	473483	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.02	188	872070	20.00	ng/ul	-0.01
78) Chrysene-d12	21.22	240	762993	20.00	ng/ul	-0.01
86) Perylene-d12	23.41	264	894138	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.15	96	74011	17.36	ng/uL	0.00
5) Phenol-d5	6.81	99	1382341	81.27	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.97	67	824587	82.71	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.16	132	1109185	83.41	ng/ul	-0.01
13) 4-Methylphenol-d8	8.34	113	1076732	77.93	ng/ul	-0.01
19) Nitrobenzene-d5	8.77	128	548286	89.03	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.49	143	599889	87.80	ng/ul	-0.02
26) 2,4-Dichlorophenol-d3	10.03	165	1057971	78.79	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.54	131	1451921	94.04	ng/ul	-0.02
44) Dimethylphthalate-d6	13.69	166	3043614	84.65	ng/ul	-0.01
47) Acenaphthylene-d8	13.96	160	3604100	87.50	ng/ul	-0.02
52) 4-Nitrophenol-d4	14.49	143	435131	69.86	ng/ul	0.00
58) Fluorene-d10	15.27	176	2535219	83.26	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.40	200	328749	60.55	ng/ul	-0.01
71) Anthracene-d10	17.12	188	3688020	89.49	ng/ul	-0.01
79) Pyrene-d10	19.42	212	3675919	87.60	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.27	264	4143824	90.97	ng/ul	-0.01

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102819\
Data File : BM023427.D
Acq On : 28 Oct 2019 19:47
Operator : JU
Sample : PB124084BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK84

Quant Time: Oct 29 11:49:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 25 07:44:12 2019
Response via : Initial Calibration

