

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/30/2023	BNA_M	<u>BM042487.D</u>
Compound Name	Response	Retention Time
DDT	764265	20.786
DDD	12766	20.35
DDE	557	19.833
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13323	777588	1.71