


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103123\
Data File : BM042512.D
Acq On : 31 Oct 2023 15:55
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103023.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Oct 31 02:55:18 2023






AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.8	88403	PASS
68	69	0.00	2	1.7	1333	PASS
69	198	0.00	100	48.9	80312	PASS
70	69	0.00	2	0.6	521	PASS
127	198	10	80	48.6	79904	PASS
197	198	0.00	2	0.5	796	PASS
198	198	100	100	100.0	164323	PASS
199	198	5	9	6.8	11211	PASS
275	198	10	60	25.6	42109	PASS
365	198	1	100	3.7	6094	PASS
441	198	0.01	100	14.2	23320	PASS
442	442	50	100	100.0	149184	PASS
443	442	15	24	19.5	29021	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103123\
Data File : BM042512.D
Acq On : 31 Oct 2023 15:55
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Nov 01 02:55:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 31 02:55:18 2023
Response via : Initial Calibration



TIC: BM042512.D\data.ms

(70) Pentachlorophenol (C)

16.957min (-0.006) 0.00 ng




response 172716

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.60	63.78
264.00	61.80	62.16
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103123\
Data File : BM042512.D
Acq On : 31 Oct 2023 15:55
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Nov 01 02:55:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 31 02:55:18 2023
Response via : Initial Calibration



TIC: BM042512.D\data.ms

(77) Benzidine**19.592min (-0.012) 200036.69 ng****response 738741**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	13.65
183.00	11.20	11.30
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/31/2023	BNA_M	<u>BM042512.D</u>
Compound Name	Response	Retention Time
DDT	571743	20.78
DDD	7608	20.386
DDE	621	19.827
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8229	579972	1.42

Instrument :
BNA_M
ClientSampleId :
DFTPP