


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110218\
 Data File : BM017485.D
 Acq On : 02 Nov 2018 07:50
 Operator : MJ/SJ
 Sample : DFTPP60
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP60

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102418MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat Nov 03 03:58:03 2018



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	165818	PASS
68	69	0.00	2	1.3	2379	PASS
69	198	0.00	100	36.6	179392	PASS
70	69	0.00	2	0.4	804	PASS
127	198	10	80	43.3	212138	PASS
197	198	0.00	2	0.6	3099	PASS
198	198	100	100	100.0	489536	PASS
199	198	5	9	6.7	32584	PASS
275	198	10	60	28.8	141176	PASS
365	198	1	100	3.9	18950	PASS
441	198	0.01	100	12.5	61261	PASS
442	198	50	100	85.3	417365	PASS
443	442	15	24	19.2	79933	PASS

m/z	Abundance
37.10	978.0
38.00	2885.0
39.10	14568.0
40.00	2341.0
41.00	488.0
43.00	492.0
44.00	5159.0
45.10	804.0
49.10	1732.0
50.10	48960.0
51.00	181440.0
52.00	9294.0
52.90	382.0
55.00	1330.0
56.00	5404.0
57.00	12705.0
58.00	655.0
61.00	2802.0
62.00	2670.0
63.10	7963.0
64.00	991.0
65.00	3936.0
65.90	340.0
68.00	3625.0
69.00	196992.0
70.00	939.0
73.10	2957.0
74.00	20192.0
75.00	31120.0
76.00	10642.0
77.10	212416.0
78.10	14569.0
79.00	15713.0
80.00	10773.0
81.00	15795.0
82.00	4389.0
83.00	3713.0
84.10	646.0
85.00	2838.0
86.00	4758.0
87.00	1823.0
88.00	742.0
89.10	505.0
91.00	3822.0
92.00	3992.0
93.00	23984.0
94.00	1654.0
95.00	315.0
96.00	1087.0
97.20	500.0
98.00	21112.0
99.00	16656.0
99.90	1223.0
101.00	8364.0
102.00	643.0
103.00	3344.0
104.00	4906.0
105.00	4747.0
106.00	1893.0
107.00	62072.0
108.00	9987.0
109.10	1852.0
110.00	116400.0
111.00	19064.0
112.00	2354.0
112.90	1055.0
114.80	352.0
116.00	3843.0
117.00	54744.0
118.10	3905.0
119.10	726.0
120.00	941.0
121.10	394.0
122.00	3924.0
123.00	6809.0
124.00	2824.0
124.90	2238.0
126.20	1842.0
127.00	215360.0
128.00	17680.0
129.00	85456.0
130.00	8081.0
131.00	1481.0
131.90	811.0
133.00	841.0
134.00	2629.0
135.10	7084.0
136.00	2811.0
137.00	3899.0
137.80	968.0
139.00	416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

140.00	780.0
141.00	11758.0
142.00	3827.0
143.00	2911.0
144.00	856.0
145.10	495.0
146.00	2363.0
146.90	5547.0
148.00	13024.0
149.00	2648.0
149.80	714.0
151.10	1644.0
151.60	805.0
153.00	3517.0
154.00	2544.0
155.00	6705.0
156.10	8789.0
157.00	1484.0
158.00	2525.0
159.00	1573.0
160.00	3821.0
161.00	5384.0
161.90	1335.0
163.00	506.0
164.00	494.0
164.90	4653.0
166.10	3938.0
167.00	25256.0
168.00	13282.0
169.00	2330.0
170.00	740.0
171.00	518.0
172.00	2191.0
173.00	2067.0
174.00	4425.0
175.00	8475.0
176.00	3668.0
177.00	4982.0
178.00	1612.0
179.00	16760.0
180.00	13791.0
181.00	6444.0
182.00	1565.0
182.90	357.0
183.90	1342.0
185.00	7828.0
186.00	66240.0
187.00	18600.0
188.00	2103.0
189.00	4280.0
189.90	696.0
191.00	2244.0
192.00	5017.0
193.00	7454.0
193.90	1517.0
195.10	1008.0
196.00	14573.0
197.10	3189.0
198.00	465728.0
199.00	33272.0
200.00	2648.0
201.50	1978.0
203.10	3015.0
204.00	16800.0
205.00	27168.0
206.10	121464.0
207.00	16904.0
208.00	4278.0
209.00	1433.0
209.90	1590.0
211.00	3808.0
212.80	422.0
215.00	1580.0
216.10	2016.0
217.00	33200.0
218.00	3725.0
219.10	436.0
221.00	16896.0
221.80	3774.0
223.00	6967.0
224.00	67664.0
225.00	16736.0
226.00	2100.0
227.00	28648.0
228.00	4402.0
229.00	5747.0
229.90	518.0
231.10	2195.0
232.00	341.0
233.10	484.0
234.00	1837.0
235.00	2128.0
236.00	1488.0
237.00	2574.0
237.90	320.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

239.00	1451.0
240.00	1094.0
241.00	1544.0
241.90	2762.0
243.10	3454.0
244.10	51352.0
245.10	6515.0
246.00	10884.0
247.00	2210.0
248.00	461.0
249.00	1993.0
250.00	427.0
250.90	603.0
251.90	526.0
253.00	1243.0
254.20	1906.0
255.00	274624.0
256.00	40224.0
257.00	3580.0
258.00	15195.0
259.10	2790.0
260.00	414.0
260.90	354.0
264.00	354.0
265.00	7443.0
266.00	1395.0
266.90	379.0
269.90	332.0
271.00	943.0
272.10	632.0
273.00	8331.0
274.00	20192.0
275.00	124504.0
276.00	17896.0
277.00	10350.0
277.90	1637.0
279.00	516.0
281.00	1017.0
282.00	390.0
283.00	1578.0
284.10	705.0
285.00	1912.0
286.00	466.0
289.00	497.0
292.10	528.0
293.00	2055.0
294.10	580.0
295.00	930.0
296.00	35000.0
297.00	4803.0
297.90	330.0
301.00	588.0
302.00	728.0
303.00	4366.0
304.10	1121.0
308.10	592.0
309.90	332.0
312.90	324.0
314.00	1664.0
315.00	3356.0
316.00	2154.0
317.10	403.0
320.90	1114.0
321.90	486.0
323.10	11907.0
324.10	1964.0
325.10	350.0
325.90	300.0
326.90	2752.0
328.10	778.0
331.90	698.0
333.10	1162.0
334.00	6413.0
335.00	1419.0
335.90	339.0
339.10	309.0
339.90	351.0
341.10	929.0
342.00	421.0
345.90	1599.0
346.90	641.0
350.90	471.0
352.00	3717.0
353.10	2145.0
354.00	4512.0
355.00	1172.0
359.00	390.0
365.00	15508.0
366.00	1882.0
371.00	873.0
372.10	5910.0
373.10	1462.0
383.20	1466.0
390.10	942.0
391.00	502.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

392.10	486.0
401.00	370.0
402.00	2116.0
403.00	3422.0
404.00	1503.0
421.10	2463.0
422.00	2408.0
423.10	18456.0
424.00	6154.0
425.10	689.0
440.10	376.0
441.10	41856.0
442.10	293824.0
443.10	60544.0
444.00	4829.0
445.00	344.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
37.00	773.0
38.00	2616.0
39.10	14691.0
39.90	1719.0
41.10	1000.0
44.00	5586.0
44.90	543.0
49.10	1415.0
50.10	50664.0
51.00	196672.0
52.00	9452.0
53.20	404.0
55.00	1477.0
56.00	6509.0
57.00	12954.0
58.10	473.0
61.00	2686.0
62.00	3056.0
63.00	8070.0
64.00	882.0
65.10	3906.0
66.10	459.0
66.90	379.0
68.20	4496.0
69.00	208320.0
70.00	810.0
73.00	2946.0
74.00	21600.0
75.00	34544.0
76.10	12033.0
77.10	228032.0
78.10	16032.0
79.00	15956.0
80.00	12764.0
81.00	16123.0
82.00	4000.0
83.00	3808.0
84.00	699.0
85.00	2859.0
86.00	4799.0
87.00	2422.0
88.00	655.0
89.00	501.0
91.00	4802.0
92.00	5152.0
93.00	28568.0
94.00	2301.0
95.00	412.0
96.00	1367.0
97.10	730.0
98.00	20600.0
99.00	16632.0
100.10	1399.0
101.00	9499.0
102.00	412.0
103.00	3719.0
104.00	6617.0
105.00	5587.0
106.00	2260.0
107.00	71688.0
108.00	10943.0
109.20	3199.0
110.00	132992.0
111.00	18984.0
112.00	2592.0
113.00	723.0
114.80	352.0
116.00	4248.0
117.00	65352.0
118.00	4840.0
119.00	615.0
120.00	1285.0
120.90	502.0
122.00	4659.0
123.00	8192.0
124.00	3412.0
125.00	3062.0
127.00	248576.0
128.00	19584.0
129.00	97656.0
130.00	8550.0
131.00	1587.0
132.10	895.0
133.10	685.0
134.00	3217.0
135.00	8352.0
136.00	3077.0
137.10	3626.0
137.80	958.0
138.80	426.0
139.20	439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

140.10	1256.0
141.00	12877.0
142.00	3754.0
143.00	2842.0
144.00	753.0
145.10	872.0
146.00	2517.0
147.00	7353.0
148.00	14240.0
149.00	3229.0
150.00	1008.0
151.20	1480.0
151.80	1185.0
153.00	4658.0
154.00	3004.0
155.00	7682.0
156.00	9951.0
157.10	1936.0
158.00	2929.0
158.90	2080.0
160.00	4252.0
161.00	6647.0
162.00	1862.0
163.00	569.0
164.00	673.0
165.00	5012.0
165.90	3383.0
167.00	27536.0
168.00	13752.0
169.00	2445.0
170.00	1091.0
171.00	1253.0
172.00	2483.0
173.00	3676.0
174.00	6520.0
175.00	9815.0
176.10	2906.0
177.00	5359.0
178.00	1615.0
179.00	20704.0
180.00	15721.0
181.00	6881.0
181.90	1314.0
182.80	457.0
184.00	1729.0
185.00	9186.0
186.00	78392.0
187.00	20856.0
188.10	2923.0
189.00	5869.0
190.00	887.0
191.10	2805.0
192.00	6650.0
193.00	8081.0
194.00	1565.0
195.00	1177.0
196.00	17144.0
197.10	3311.0
198.00	593536.0
199.00	37960.0
199.90	3083.0
201.50	2132.0
203.00	4162.0
204.10	20888.0
205.00	35032.0
206.10	156672.0
207.00	21352.0
208.00	5912.0
209.00	1860.0
210.10	2218.0
211.00	5961.0
213.00	324.0
214.90	2128.0
216.00	2974.0
217.00	41072.0
218.00	5338.0
219.00	529.0
219.60	514.0
221.00	22808.0
221.80	5352.0
223.00	8936.0
224.00	86328.0
225.00	21760.0
226.00	2420.0
227.00	35832.0
227.90	5326.0
229.00	7341.0
230.00	937.0
231.00	2724.0
232.00	503.0
233.00	566.0
234.00	2214.0
235.10	2895.0
236.00	2067.0
237.00	3280.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

237.90	551.0
238.90	1336.0
240.00	923.0
241.00	1535.0
242.10	3835.0
243.00	5060.0
244.10	64440.0
245.10	8726.0
246.00	14073.0
247.00	2503.0
247.90	413.0
249.00	2389.0
250.00	382.0
251.00	690.0
252.00	844.0
252.90	1381.0
254.10	3089.0
255.00	368192.0
256.00	52928.0
257.10	4752.0
258.00	21760.0
259.00	3194.0
259.90	624.0
261.10	595.0
263.00	375.0
264.10	616.0
265.00	7713.0
266.00	1254.0
271.00	861.0
271.90	946.0
273.10	9962.0
274.00	29616.0
275.00	168832.0
276.00	23096.0
277.00	13730.0
278.00	2760.0
278.80	539.0
281.20	732.0
282.10	711.0
283.00	1360.0
284.00	1326.0
285.10	2210.0
286.00	440.0
288.90	539.0
289.90	539.0
290.90	446.0
292.00	571.0
293.00	3630.0
294.00	1187.0
295.10	1075.0
296.00	46968.0
297.00	6398.0
298.00	456.0
301.10	417.0
302.10	869.0
303.00	5766.0
304.10	1217.0
308.00	637.0
308.90	499.0
310.20	632.0
311.00	477.0
313.10	514.0
314.00	2204.0
315.00	4913.0
316.10	2864.0
317.10	675.0
321.00	1445.0
322.10	803.0
323.10	15801.0
324.00	3073.0
325.10	515.0
325.90	463.0
327.00	2855.0
328.00	1329.0
329.10	347.0
332.00	1074.0
333.10	1202.0
334.10	10851.0
335.00	2842.0
336.00	437.0
339.00	322.0
341.00	1777.0
341.90	574.0
346.00	3875.0
347.10	864.0
352.10	4832.0
353.00	4141.0
354.00	5380.0
355.10	1195.0
359.00	338.0
365.00	23496.0
366.10	3177.0
369.90	360.0
371.10	1216.0
372.00	8672.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

373.00	2133.0
383.00	2510.0
384.10	856.0
390.10	1248.0
390.90	839.0
392.10	701.0
401.10	684.0
402.00	3261.0
403.00	5503.0
404.00	1750.0
415.10	302.0
421.10	3386.0
422.00	4852.0
423.10	31464.0
424.10	7602.0
425.00	837.0
441.10	73088.0
442.10	507392.0
443.10	95904.0
444.10	9239.0
445.10	768.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
37.00	585.0
38.00	1262.0
39.00	8941.0
40.00	1554.0
41.00	514.0
44.00	4505.0
45.10	494.0
48.90	844.0
50.00	30504.0
51.00	119344.0
52.00	5337.0
55.00	993.0
55.90	3493.0
57.00	8077.0
58.00	384.0
61.00	1940.0
62.00	1861.0
63.10	5309.0
64.00	597.0
65.00	2401.0
67.10	393.0
68.20	2643.0
69.00	132864.0
70.00	663.0
73.10	2136.0
74.10	13352.0
75.00	19784.0
76.10	7467.0
77.10	144576.0
78.10	8707.0
79.00	10416.0
80.00	7548.0
81.00	10983.0
82.00	2440.0
83.00	2380.0
85.10	2217.0
86.00	2988.0
87.00	1342.0
87.90	617.0
91.00	2278.0
92.00	3125.0
93.00	17424.0
94.00	1224.0
95.10	365.0
96.10	1239.0
98.00	13785.0
99.00	11526.0
99.90	880.0
101.00	6254.0
102.00	627.0
103.00	2066.0
104.00	4337.0
104.90	3648.0
106.10	1354.0
107.00	47528.0
108.00	6523.0
109.10	1906.0
110.00	86416.0
111.00	13205.0
112.00	1697.0
112.90	408.0
114.90	365.0
116.00	3062.0
117.00	37944.0
118.00	2532.0
118.90	645.0
120.00	619.0
120.90	405.0
122.00	3441.0
123.00	5341.0
124.00	2033.0
125.00	2158.0
127.00	172480.0
128.00	11817.0
129.00	68344.0
130.00	5711.0
131.00	1153.0
132.30	362.0
132.90	540.0
134.00	2078.0
135.00	5207.0
136.00	2496.0
137.00	2694.0
138.00	591.0
139.10	405.0
140.10	800.0
141.00	9025.0
142.10	2791.0
142.80	1920.0
143.90	480.0
144.90	407.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

146.00	2250.0
147.10	5133.0
148.00	11085.0
149.00	2132.0
149.90	918.0
151.20	1183.0
151.80	517.0
152.10	530.0
153.00	3034.0
154.00	2411.0
155.10	5109.0
156.00	7881.0
157.00	1538.0
158.00	1810.0
159.00	1661.0
159.90	2942.0
161.00	3785.0
162.00	1467.0
163.10	397.0
164.10	493.0
165.00	4105.0
166.00	2959.0
167.00	17376.0
168.00	9847.0
169.10	1581.0
170.00	747.0
170.90	794.0
172.00	1697.0
173.00	1819.0
174.00	4066.0
175.10	7783.0
176.00	2364.0
177.00	3949.0
178.00	1035.0
179.00	13986.0
180.00	9605.0
181.00	4445.0
182.00	1045.0
183.00	472.0
184.00	1217.0
185.00	7011.0
186.00	54160.0
187.00	16464.0
188.00	1816.0
189.00	3400.0
190.00	863.0
191.00	1968.0
192.00	4983.0
193.00	4887.0
194.00	1198.0
195.10	636.0
196.00	12099.0
197.10	2799.0
198.00	409344.0
199.00	26520.0
200.00	1809.0
201.60	1739.0
203.00	3078.0
204.00	14688.0
205.00	25176.0
206.00	112304.0
207.10	17768.0
208.00	4640.0
209.00	1685.0
210.10	1712.0
211.10	4207.0
215.00	1236.0
216.00	2142.0
217.00	28832.0
218.00	3693.0
218.90	487.0
221.00	17128.0
223.00	6493.0
224.00	60216.0
225.00	14945.0
226.10	1147.0
227.00	28632.0
228.00	3508.0
229.00	5985.0
229.90	696.0
231.00	2281.0
232.00	632.0
233.00	587.0
234.00	1571.0
234.90	1983.0
235.90	1389.0
237.00	2162.0
238.00	649.0
239.00	1252.0
239.90	882.0
241.00	1580.0
242.00	3385.0
243.00	2990.0
244.00	51224.0
245.00	5836.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

246.00	9636.0
247.00	2089.0
249.00	1916.0
249.90	609.0
250.90	512.0
252.00	520.0
253.10	1214.0
254.10	2374.0
255.00	276480.0
256.00	39168.0
257.00	3071.0
258.00	15871.0
259.00	3093.0
259.90	390.0
260.90	518.0
264.00	523.0
265.00	6803.0
266.10	788.0
267.00	446.0
269.90	494.0
271.00	466.0
272.10	1022.0
273.00	8007.0
274.00	21480.0
275.00	130192.0
276.00	16170.0
277.00	10350.0
278.00	1811.0
281.00	1099.0
282.00	374.0
283.00	1538.0
284.00	727.0
285.10	2277.0
289.90	301.0
291.10	311.0
291.80	426.0
292.20	427.0
293.00	2400.0
293.90	473.0
295.10	762.0
296.00	38568.0
297.00	5206.0
301.00	459.0
301.90	1090.0
303.10	4187.0
304.10	1017.0
307.90	588.0
309.00	327.0
309.90	559.0
313.10	307.0
314.10	2383.0
315.00	4631.0
316.10	2324.0
317.10	715.0
321.00	1751.0
321.90	927.0
323.10	12117.0
324.00	2366.0
325.00	304.0
327.00	2556.0
328.10	1387.0
332.10	960.0
333.00	1310.0
334.00	9318.0
335.00	2269.0
341.10	1739.0
342.00	361.0
346.00	2762.0
347.00	451.0
352.00	4051.0
353.00	3026.0
354.10	4503.0
355.00	600.0
356.00	336.0
365.00	17848.0
366.00	2358.0
369.90	416.0
371.10	1231.0
372.10	7441.0
373.00	1372.0
374.20	386.0
383.10	1695.0
384.00	825.0
390.00	764.0
391.20	719.0
392.10	1074.0
401.00	594.0
402.00	2554.0
403.10	4316.0
404.00	1892.0
405.10	363.0
421.00	3944.0
422.00	3278.0
423.10	26416.0
424.00	7309.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

425.00	1062.0
441.10	68840.0
442.10	450880.0
443.10	83352.0
444.00	6968.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60

m/z	Abundance
40.00	1591.0
44.00	5454.0
73.00	1285.0
78.00	383.0
95.80	354.0
96.10	360.0
133.00	384.0
135.10	334.0
191.00	381.0
193.00	302.0
206.90	3239.0
208.00	501.0
209.10	315.0
221.10	400.0
281.20	684.0

Instrument :
BNA_M
ClientSampleId :
DFTPP60