

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110323\
 Data File : BM042577.D
 Acq On : 03 Nov 2023 19:11
 Operator : MA/JU
 Sample : 05220-07 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 L-2(125FT)(0-5)

Quant Time: Nov 03 22:40:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 03 15:52:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	51213	20.000	ng	0.00
21) Naphthalene-d8	10.733	136	191783	20.000	ng	0.00
39) Acenaphthene-d10	14.568	164	119770	20.000	ng	0.00
64) Phenanthrene-d10	17.321	188	267003	20.000	ng	0.00
76) Chrysene-d12	21.503	240	264943	20.000	ng	0.00
86) Perylene-d12	23.927	264	290860	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	70187	22.118	ng	0.01
7) Phenol-d6	7.081	99	90234	20.734	ng	0.00
23) Nitrobenzene-d5	9.116	82	65072	15.817	ng	0.00
42) 2,4,6-Tribromophenol	16.062	330	37069	23.706	ng	0.00
45) 2-Fluorobiphenyl	13.186	172	140649	15.802	ng	0.00
79) Terphenyl-d14	19.933	244	239317	15.316	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	10.786	128	33168	3.347	ng	98
37) 2-Methylnaphthalene	12.610	142	192076	27.525	ng	94
38) 1-Methylnaphthalene	12.610	142	192517	29.310	ng	99
46) 1,1'-Biphenyl	13.404	154	48377	5.284	ng	98
48) 2-Nitroaniline	13.739	65	3786	5.621	ng	# 23
49) Acenaphthylene	14.298	152	192755	17.329	ng	98
52) Acenaphthene	14.633	154	55845	8.272	ng	98
53) 3-Nitroaniline	14.286	138	195	4.437	ng	# 1
54) 2,4-Dinitrophenol	14.910	184	62	7.790	ng	# 1
55) Dibenzofuran	14.968	168	25262	2.302	ng	97
56) 4-Nitrophenol	14.839	139	983	7.489	ng	# 21
57) 2,4-Dinitrotoluene	14.951	165	2908	4.925	ng	# 57
58) Fluorene	15.615	166	172600	19.435	ng	99
62) 4-Nitroaniline	15.621	138	2181	6.778	ng	# 12
65) 4,6-Dinitro-2-methylph...	15.715	198	60	5.865	ng	# 1
71) Phenanthrene	17.362	178	1076502	77.150	ng	100
72) Anthracene	17.456	178	256398	18.552	ng	100
75) Fluoranthene	19.380	202	550973	33.615	ng	98
78) Pyrene	19.745	202	1000056	54.970	ng	98
81) Benzo(a)anthracene	21.486	228	371813	22.057	ng	95
83) Chrysene	21.486	228	371813	21.714	ng	96
87) Indeno(1,2,3-cd)pyrene	26.450	276	123142	7.113	ng	# 92
88) Benzo(b)fluoranthene	23.180	252	332518	20.520	ng	97
89) Benzo(k)fluoranthene	23.180	252	332518	18.415	ng	97
90) Benzo(a)pyrene	23.815	252	289160	18.125	ng	97
91) Dibenzo(a,h)anthracene	26.456	278	42048	6.542	ng	# 78
92) Benzo(g,h,i)perylene	27.244	276	144844	9.493	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110323\
Data File : BM042577.D
Acq On : 03 Nov 2023 19:11
Operator : MA/JU
Sample : 05220-07 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)

Quant Time: Nov 03 22:40:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 03 15:52:20 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.916 min Scan# 806
Delta R.T. 0.006 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:152 Resp: 51213
Ion Ratio Lower Upper
152 100
150 152.2 127.4 191.0
115 59.9 47.9 71.9



#5
2-Fluorophenol
Concen: 22.118 ng
RT: 5.451 min Scan# 385
Delta R.T. 0.011 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:112 Resp: 70187
Ion Ratio Lower Upper
112 100
64 65.7 52.1 78.1
63 40.0 28.3 42.5









#7
Phenol-d6
Concen: 20.734 ng
RT: 7.081 min Scan# 60
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion: 99 Resp: 90234
Ion Ratio Lower Upper
99 100
42 28.0 18.7 28.1
71 38.7 29.1 43.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.733 min Scan# 1283
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion: 136 Resp: 191783
Ion Ratio Lower Upper
136 100
137 11.2 9.0 13.6
54 11.1 8.0 12.0
68 6.3 5.7 8.5





#23

Nitrobenzene-d5

Concen: 15.817 ng

RT: 9.116 min Scan# 1008

Delta R.T. 0.005 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion: 82 Resp: 65072

Ion Ratio Lower Upper

82 100

128 38.8 30.7 46.1

54 58.5 45.3 67.9



#31

Naphthalene

Concen: 3.347 ng

RT: 10.786 min Scan# 1292

Delta R.T. 0.005 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion: 128 Resp: 33168


Ion Ratio Lower Upper

128 100

129 12.3 8.8 13.2

127 13.5 10.6 15.8









#37
2-Methylnaphthalene
Concen: 27.525 ng
RT: 12.610 min Scan# 1602
Delta R.T. 0.223 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:142 Resp: 192076
Ion Ratio Lower Upper
142 100
141 92.3 70.9 106.3
115 45.0 30.2 45.2



#38
1-Methylnaphthalene
Concen: 29.310 ng
RT: 12.610 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:142 Resp: 192517
Ion Ratio Lower Upper
142 100
141 92.3 74.4 111.6
116 4.6 3.2 4.8









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.568 min Scan# 1935
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:164 Resp: 119770
Ion Ratio Lower Upper
164 100
162 100.2 81.6 122.4
160 46.6 36.2 54.4



#42
2,4,6-Tribromophenol
Concen: 23.706 ng
RT: 16.062 min Scan# 2189
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:330 Resp: 37069
Ion Ratio Lower Upper
330 100
332 98.2 78.8 118.2
141 38.8 31.1 46.7









#45
2-Fluorobiphenyl
Concen: 15.802 ng
RT: 13.186 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:172 Resp: 140649
Ion Ratio Lower Upper
172 100
171 34.3 28.2 42.2
170 22.8 19.0 28.4



#46
1,1'-Biphenyl
Concen: 5.284 ng
RT: 13.404 min Scan# 1737
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:154 Resp: 48377
Ion Ratio Lower Upper
154 100
153 39.5 20.3 60.3
76 14.1 0.0 35.0









#48
2-Nitroaniline
Concen: 5.621 ng
RT: 13.739 min Scan# 11
Delta R.T. 0.047 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion: 65 Resp: 3786
Ion Ratio Lower Upper
65 100
92 0.0 49.4 74.2#
138 16.7 66.9 100.3#



#49
Acenaphthylene
Concen: 17.329 ng
RT: 14.298 min Scan# 1889
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:152 Resp: 192755
Ion Ratio Lower Upper
152 100
151 21.1 16.4 24.6
153 14.0 10.4 15.6





#52

Acenaphthene

Concen: 8.272 ng

RT: 14.633 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:154 Resp: 55845

Ion Ratio Lower Upper

154 100

153 116.0 92.0 138.0

152 59.0 44.0 66.0



#53

3-Nitroaniline

Concen: 4.437 ng

RT: 14.286 min Scan# 1887

Delta R.T. -0.236 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:138 Resp: 195


Ion Ratio Lower Upper

138 100

108 80.2 7.9 11.9#

92 0.0 100.7 151.1#









#54
2,4-Dinitrophenol
Concen: 7.790 ng
RT: 14.910 min Scan# 1963
Delta R.T. 0.182 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:184 Resp: 62
Ion Ratio Lower Upper
184 100
63 555.1 69.2 103.8#
154 175.6 53.2 79.8#



#55
Dibenzofuran
Concen: 2.302 ng
RT: 14.968 min Scan# 2003
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:168 Resp: 25262
Ion Ratio Lower Upper
168 100
139 39.8 30.4 45.6
169 12.8 10.6 15.8









#56
4-Nitrophenol
Concen: 7.489 ng
RT: 14.839 min Scan# 1980
Delta R.T. 0.017 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:139 Resp: 983
Ion Ratio Lower Upper
139 100
109 0.0 61.3 101.3#
65 41.8 100.2 140.2#



#57
2,4-Dinitrotoluene
Concen: 4.925 ng
RT: 14.951 min Scan# 2000
Delta R.T. -0.012 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:165 Resp: 2908
Ion Ratio Lower Upper
165 100
63 115.1 66.7 100.1#
89 49.8 80.2 120.4#
182 10.2 2.0 3.0#









#58
Fluorene
Concen: 19.435 ng
RT: 15.615 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:166 Resp: 172600
Ion Ratio Lower Upper
166 100
165 99.7 79.3 118.9
167 14.0 10.5 15.7



#62
4-Nitroaniline
Concen: 6.778 ng
RT: 15.621 min Scan# 2114
Delta R.T. -0.065 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:138 Resp: 2181
Ion Ratio Lower Upper
138 100
92 0.0 37.0 77.0#
108 0.0 67.2 107.2#





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.321 min Scan# 24

Delta R.T. 0.006 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:188 Resp: 267003

Ion Ratio Lower Upper

188 100

94 8.3 8.2 12.4

80 9.5 9.4 14.2



#65

4,6-Dinitro-2-methylphenol

Concen: 5.865 ng

RT: 15.715 min Scan# 2130

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:198 Resp: 60


Ion Ratio Lower Upper

198 100

51 490.1 42.4 82.4#

105 274.3 29.2 69.2#









#71
Phenanthrene
Concen: 77.150 ng
RT: 17.362 min Scan# 2412
Delta R.T. -0.000 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:178 Resp: 1076502
Ion Ratio Lower Upper
178 100
176 19.7 15.8 23.8
179 15.6 12.2 18.4



#72
Anthracene
Concen: 18.552 ng
RT: 17.456 min Scan# 2426
Delta R.T. 0.006 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:178 Resp: 256398
Ion Ratio Lower Upper
178 100
176 18.9 15.2 22.8
179 15.7 12.3 18.5





#75

Fluoranthene

Concen: 33.615 ng

RT: 19.380 min Scan# 21

Delta R.T. 0.005 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:202 Resp: 550973

Ion Ratio Lower Upper

202 100

101 9.0 0.0 30.5

203 17.6 0.0 37.0



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.503 min Scan# 3114

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:240 Resp: 264943


Ion Ratio Lower Upper

240 100

120 8.7 8.3 12.5

236 26.7 20.8 31.2





#78

Pyrene

Concen: 54.970 ng

RT: 19.745 min Scan# 2

Delta R.T. 0.006 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:202 Resp: 1000056

Ion Ratio Lower Upper

202 100

200 21.6 16.9 25.3

203 18.2 13.7 20.5



#79

Terphenyl-d14

Concen: 15.316 ng

RT: 19.933 min Scan# 2847

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:244 Resp: 239317


Ion Ratio Lower Upper

244 100

212 7.5 5.4 8.2

122 8.8 7.2 10.8





#81

Benzo(a)anthracene

Concen: 22.057 ng

RT: 21.486 min Scan# 3111

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:228 Resp: 371813

Ion Ratio Lower Upper

228 100

226 29.9 22.3 33.5

229 23.0 15.8 23.8



#83

Chrysene

Concen: 21.714 ng

RT: 21.486 min Scan# 3111

Delta R.T. -0.053 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:228 Resp: 371813


Ion Ratio Lower Upper

228 100

226 29.9 24.4 36.6

229 23.0 15.7 23.5





#86

Perylene-d12

Concen: 20.000 ng

RT: 23.927 min Scan# 3

Delta R.T. 0.011 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11

Instrument :

BNA_M

ClientSampleId :

L-2(125FT)(0-5)





Tgt Ion:264 Resp: 290860

Ion Ratio Lower Upper

264 100

260 24.0 19.0 28.4

265 23.7 18.7 28.1



#87

Indeno(1,2,3-cd)pyrene

Concen: 7.113 ng

RT: 26.450 min Scan# 3955

Delta R.T. -0.000 min

Lab File: BM042577.D

Acq: 03 Nov 2023 19:11




Tgt Ion:276 Resp: 123142


Ion Ratio Lower Upper

276 100

138 18.6 20.6 31.0#

277 24.9 19.5 29.3









#88
Benzo(b)fluoranthene
Concen: 20.520 ng
RT: 23.180 min Scan# 3399
Delta R.T. 0.005 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:252 Resp: 332518
Ion Ratio Lower Upper
252 100
253 23.4 17.5 26.3
125 10.9 7.8 11.6



#89
Benzo(k)fluoranthene
Concen: 18.415 ng
RT: 23.180 min Scan# 3399
Delta R.T. -0.042 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:252 Resp: 332518
Ion Ratio Lower Upper
252 100
253 23.4 17.6 26.4
125 10.9 8.2 12.4









#90
Benzo(a)pyrene
Concen: 18.125 ng
RT: 23.815 min Scan# 31
Delta R.T. 0.006 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11




Instrument :
BNA_M
ClientSampleId :
L-2(125FT)(0-5)


Tgt Ion:252 Resp: 289160
Ion Ratio Lower Upper
252 100
253 22.8 17.5 26.3
125 10.1 9.4 14.2



#91
Dibenzo(a,h)anthracene
Concen: 6.542 ng
RT: 26.456 min Scan# 3956
Delta R.T. -0.012 min
Lab File: BM042577.D
Acq: 03 Nov 2023 19:11

Tgt Ion:278 Resp: 42048
Ion Ratio Lower Upper
278 100
139 21.9 13.4 20.2#
279 38.9 19.7 29.5#





#92
 Benzo(g,h,i)perylene
 Concen: 9.493 ng
 RT: 27.244 min Scan# 4090
 Delta R.T. -0.000 min
 Lab File: BM042577.D
 Acq: 03 Nov 2023 19:11

Instrument :
 BNA_M
 ClientSampleId :
 L-2(125FT)(0-5)

Tgt Ion:276 Resp: 144844
 Ion Ratio Lower Upper
 276 100
 277 25.4 19.2 28.8
 138 17.1 17.3 25.9#

