

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
 Data File : BM027703.D
 Acq On : 05 Nov 2020 03:10
 Operator : JU/CG
 Sample : SSTD0.806
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.8064

Quant Time: Nov 05 03:58:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 03:44:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	363	0.40	ng/ul	0.00
4) Naphthalene-d8	11.24	136	1240	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.01	164	780	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.72	188	1678	0.40	ng/ul	0.00
19) Chrysene-d12	21.82	240	1514	0.40	ng/ul	0.00
23) Perylene-d12	24.27	264	1390	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.54	96	295	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.80	152	1690	0.85	ng/ul	0.00
17) Fluoranthene-d10	19.71	212	3746	0.85	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.58	88	317	0.628	ng/ul	95
5) Naphthalene	11.29	128	2856	0.828	ng/ul	95
7) 2-Methylnaphthalene	12.87	142	1937	0.833	ng/ul	99
8) 1-Methylnaphthalene	13.09	142	1847	0.836	ng/ul	98
10) Acenaphthylene	14.74	152	2809	0.784	ng/ul	96
11) Acenaphthene	15.07	153	2109	0.790	ng/ul	99
12) Fluorene	16.04	166	2519	0.811	ng/ul	97
14) Pentachlorophenol	17.36	266	457	0.844	ng/ul	96
15) Phenanthrene	17.77	178	3953	0.794	ng/ul	95
16) Anthracene	17.85	178	3864	0.805	ng/ul	98
18) Fluoranthene	19.74	202	4730	0.825	ng/ul	98
20) Pyrene	20.09	202	4717	0.763	ng/ul	97
21) Benzo(a)anthracene	21.80	228	4434	0.792	ng/ul	99
22) Chrysene	21.86	228	4598	0.792	ng/ul	100
24) Benzo(b)fluoranthene	23.52	252	4480	0.781	ng/ul	96
25) Benzo(k)fluoranthene	23.57	252	4494	0.786	ng/ul	95
26) Benzo(a)pyrene	24.15	252	4112	0.811	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	26.80	276	4912	0.835	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.81	278	4121	0.841	ng/ul	92
29) Benzo(g,h,i)perylene	27.58	276	4113	0.816	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
Data File : BM027703.D
Acq On : 05 Nov 2020 03:10
Operator : JU/CG
Sample : SST00.806
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.8064

Quant Time: Nov 05 03:58:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 03:44:30 2020
Response via : Initial Calibration

