

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
 Data File : BM027712.D
 Acq On : 05 Nov 2020 11:12
 Operator : JU/CG
 Sample : PB132684BL
 Misc : BL-2
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK684

Quant Time: Nov 05 12:17:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	378	0.40	ng/ul	0.00
4) Naphthalene-d8	11.24	136	1317	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.01	164	819	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.72	188	1822	0.40	ng/ul	0.00
19) Chrysene-d12	21.82	240	1466	0.40	ng/ul	0.00
23) Perylene-d12	24.27	264	1285	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.54	96	278	0.72	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.80	152	832	0.38	ng/ul	0.00
17) Fluoranthene-d10	19.71	212	1817	0.37	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
Data File : BM027712.D
Acq On : 05 Nov 2020 11:12
Operator : JU/CG
Sample : PB132684BL
Misc : BL-2
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK684

Quant Time: Nov 05 12:17:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

