

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
 Data File : BM027721.D
 Acq On : 05 Nov 2020 16:50
 Operator : JU/CG
 Sample : L4485-09
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleID :
 MDL-WATER-02

Quant Time: Nov 05 17:23:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	290	0.40	ng/ul	0.00
4) Naphthalene-d8	11.24	136	1000	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.01	164	603	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.72	188	1353	0.40	ng/ul	0.00
19) Chrysene-d12	21.82	240	1229	0.40	ng/ul	0.00
23) Perylene-d12	24.27	264	1139	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.54	96	160	0.54	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.80	152	470	0.28	ng/ul	0.00
17) Fluoranthene-d10	19.71	212	1065	0.29	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.59	88	24	0.075	ng/ul#	84
5) Naphthalene	11.29	128	167	0.059	ng/ul#	40
7) 2-Methylnaphthalene	12.87	142	114	0.058	ng/ul	100
8) 1-Methylnaphthalene	13.09	142	107	0.058	ng/ul	99
10) Acenaphthylene	14.74	152	160	0.059	ng/ul#	50
11) Acenaphthene	15.07	153	122	0.060	ng/ul#	92
12) Fluorene	16.04	166	138	0.057	ng/ul#	86
14) Pentachlorophenol	17.37	266	34	0.076	ng/ul#	88
15) Phenanthrene	17.76	178	217	0.055	ng/ul#	60
16) Anthracene	17.85	178	215	0.055	ng/ul#	58
18) Fluoranthene	19.74	202	285	0.062	ng/ul#	55
20) Pyrene	20.09	202	286	0.060	ng/ul#	59
21) Benzo(a)anthracene	21.80	228	301	0.067	ng/ul#	75
22) Chrysene	21.86	228	305	0.066	ng/ul#	81
24) Benzo(b)fluoranthene	23.52	252	308	0.067	ng/ul#	45
25) Benzo(k)fluoranthene	23.57	252	281	0.062	ng/ul#	46
26) Benzo(a)pyrene	24.16	252	276	0.067	ng/ul#	39
27) Indeno(1,2,3-cd)pyrene	26.81	276	309	0.063	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.81	278	259	0.064	ng/ul#	35
29) Benzo(g,h,i)perylene	27.58	276	264	0.064	ng/ul#	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
Data File : BM027721.D
Acq On : 05 Nov 2020 16:50
Operator : JU/CG
Sample : L4485-09
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-WATER-02

Quant Time: Nov 05 17:23:51 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

