

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
 Data File : BM027725.D
 Acq On : 05 Nov 2020 19:15
 Operator : JU/CG
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4078

Quant Time: Nov 06 01:33:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	361	0.40	ng/ul	0.00
4) Naphthalene-d8	11.24	136	1220	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.01	164	704	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.72	188	1518	0.40	ng/ul	0.00
19) Chrysene-d12	21.82	240	1299	0.40	ng/ul	0.00
23) Perylene-d12	24.27	264	1215	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.54	96	169	0.46	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.80	152	809	0.40	ng/ul	0.00
17) Fluoranthene-d10	19.71	212	1709	0.42	ng/ul	0.00
Target Compounds					Ovalue	
2) 1,4-Dioxane	3.58	88	187	0.469	ng/ul	92
5) Naphthalene	11.29	128	1403	0.406	ng/ul	98
7) 2-Methylnaphthalene	12.87	142	956	0.401	ng/ul	99
8) 1-Methylnaphthalene	13.09	142	911	0.405	ng/ul	98
10) Acenaphthylene	14.74	152	1300	0.413	ng/ul	97
11) Acenaphthene	15.07	153	990	0.414	ng/ul	98
12) Fluorene	16.04	166	1155	0.409	ng/ul	96
14) Pentachlorophenol	17.36	266	212	0.424	ng/ul	93
15) Phenanthrene	17.76	178	1843	0.416	ng/ul	98
16) Anthracene	17.85	178	1810	0.416	ng/ul	98
18) Fluoranthene	19.73	202	2127	0.412	ng/ul	96
20) Pyrene	20.09	202	2138	0.422	ng/ul	99
21) Benzo(a)anthracene	21.80	228	2011	0.423	ng/ul	99
22) Chrysene	21.86	228	2046	0.420	ng/ul	99
24) Benzo(b)fluoranthene	23.52	252	2014	0.412	ng/ul	96
25) Benzo(k)fluoranthene	23.57	252	1950	0.401	ng/ul	96
26) Benzo(a)pyrene	24.16	252	1808	0.409	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.79	276	2147	0.413	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.81	278	1809	0.418	ng/ul	98
29) Benzo(g,h,i)perylene	27.58	276	1838	0.420	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
Data File : BM027725.D
Acq On : 05 Nov 2020 19:15
Operator : JU/CG
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4078

Quant Time: Nov 06 01:33:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

