

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
 Data File : BM027717.D
 Acq On : 05 Nov 2020 14:25
 Operator : JU/CG
 Sample : PB132685BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK685

Quant Time: Nov 05 15:01:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.40	152	324	0.40	ng/ul	0.00
4) Naphthalene-d8	11.24	136	1072	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.01	164	675	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.72	188	1499	0.40	ng/ul	0.00
19) Chrysene-d12	21.82	240	1309	0.40	ng/ul	0.00
23) Perylene-d12	24.27	264	1253	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.54	96	249	0.75	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.80	152	683	0.38	ng/ul	0.00
17) Fluoranthene-d10	19.71	212	1544	0.38	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110520\
Data File : BM027717.D
Acq On : 05 Nov 2020 14:25
Operator : JU/CG
Sample : PB132685BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK685

Quant Time: Nov 05 15:01:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

