


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM110520\
 Data File : BM027699.D
 Acq On : 05 Nov 2020 00:10
 Operator : JU/CG
 Sample : DFTPP84
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP841

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 05 04:58:48 2020



AutoFind: Scans 2573, 2574, 2575; Background Corrected with Scan 2567

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	68.0	21773	PASS
68	69	0.00	2	1.5	293	PASS
69	198	0.00	100	61.6	19736	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.7	17503	PASS
197	198	0.00	2	1.1	361	PASS
198	198	100	100	100.0	32021	PASS
199	198	5	9	7.4	2360	PASS
275	198	10	60	28.1	8991	PASS
365	198	1	100	5.1	1623	PASS
441	198	0.01	100	13.1	4202	PASS
442	198	50	100	87.2	27930	PASS
443	442	15	24	18.7	5210	PASS

m/z	Abundance
207.00	179.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

m/z	Abundance
36.90	150.0
38.00	358.0
39.10	1370.0
49.00	208.0
50.00	4927.0
51.10	17064.0
52.00	681.0
56.00	620.0
57.00	1327.0
60.90	341.0
62.00	300.0
63.00	815.0
65.00	255.0
67.90	368.0
69.00	15664.0
72.90	244.0
74.00	1819.0
75.00	2471.0
76.00	821.0
77.10	15770.0
78.00	1020.0
79.00	1400.0
79.90	882.0
80.90	1277.0
82.00	391.0
82.80	225.0
85.00	213.0
86.00	357.0
91.00	345.0
92.00	257.0
93.00	2726.0
93.90	210.0
98.00	2380.0
99.00	1639.0
100.00	150.0
100.90	628.0
102.90	158.0
104.00	474.0
104.90	439.0
107.00	5333.0
108.00	875.0
108.90	153.0
110.00	6455.0
111.00	1365.0
112.00	241.0
115.90	431.0
117.00	6327.0
117.90	400.0
120.00	161.0
122.00	425.0
122.90	456.0
123.90	308.0
125.00	172.0
127.00	12781.0
128.00	1199.0
129.00	6060.0
129.90	439.0
135.00	592.0
136.00	206.0
137.00	231.0
140.90	800.0
142.00	177.0
143.00	167.0
146.00	196.0
147.00	387.0
147.90	1186.0
149.00	242.0
152.80	301.0
153.90	219.0
155.00	335.0
156.00	664.0
160.10	256.0
160.90	393.0
162.00	161.0
165.00	372.0
166.00	298.0
166.90	1730.0
168.00	1083.0
169.00	173.0
172.00	169.0
174.00	466.0
175.10	697.0
177.00	342.0
178.90	1220.0
180.00	791.0
181.00	317.0
185.00	583.0
186.00	4202.0
187.00	1137.0
188.90	339.0
190.90	224.0
192.00	395.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

193.10	513.0
195.90	497.0
197.00	323.0
198.00	21880.0
199.00	1624.0
200.10	158.0
203.00	221.0
204.00	1112.0
205.00	1818.0
206.00	6565.0
207.00	1005.0
207.90	269.0
211.00	184.0
215.90	278.0
216.90	2286.0
217.90	289.0
221.00	778.0
222.00	281.0
222.90	467.0
224.00	4046.0
225.10	878.0
227.00	2197.0
228.00	324.0
228.90	524.0
234.00	196.0
235.00	204.0
237.00	361.0
241.00	166.0
243.00	157.0
244.00	2676.0
245.10	284.0
246.00	761.0
254.20	274.0
255.00	13462.0
256.00	2297.0
258.00	1097.0
258.90	197.0
265.00	452.0
272.90	441.0
274.10	959.0
275.00	5515.0
275.90	521.0
277.00	578.0
296.00	2419.0
297.00	327.0
302.90	154.0
314.80	156.0
323.00	696.0
326.90	165.0
334.00	471.0
352.00	222.0
364.90	903.0
371.90	198.0
403.00	158.0
422.00	179.0
423.00	668.0
423.90	319.0
441.00	1994.0
442.00	13288.0
443.00	2385.0
443.90	319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

m/z	Abundance
38.00	407.0
39.10	2410.0
40.00	212.0
49.00	267.0
50.00	8086.0
51.10	26640.0
52.00	1433.0
55.10	201.0
56.00	793.0
56.90	1569.0
60.90	440.0
62.00	423.0
63.00	1172.0
65.00	413.0
68.10	532.0
69.00	23376.0
72.90	284.0
74.00	2936.0
75.00	4536.0
76.10	1417.0
77.10	24488.0
78.10	1672.0
79.00	1871.0
80.00	1688.0
80.90	2255.0
82.00	287.0
82.90	382.0
85.00	267.0
85.90	754.0
87.00	350.0
91.00	521.0
92.00	541.0
93.00	4070.0
94.00	247.0
98.00	3665.0
99.00	2494.0
99.90	218.0
100.90	1008.0
102.90	346.0
104.00	693.0
105.00	588.0
106.00	249.0
107.00	8384.0
108.00	1162.0
109.10	164.0
110.00	10873.0
111.00	2053.0
112.00	298.0
115.90	507.0
117.00	10947.0
117.90	564.0
122.00	663.0
123.00	837.0
123.90	445.0
124.90	260.0
127.00	21000.0
127.90	1796.0
129.00	9994.0
129.90	958.0
131.00	226.0
133.90	253.0
134.90	811.0
135.90	344.0
137.00	456.0
140.90	1357.0
142.00	432.0
143.00	329.0
146.00	230.0
147.00	742.0
148.00	2107.0
148.90	422.0
153.00	240.0
153.90	323.0
154.90	657.0
156.00	1098.0
156.90	307.0
158.10	187.0
158.90	329.0
159.90	372.0
160.90	512.0
161.90	187.0
164.90	677.0
166.00	570.0
167.00	2993.0
168.00	1802.0
169.10	161.0
169.90	166.0
171.90	272.0
172.90	395.0
174.00	522.0
175.00	863.0
176.10	280.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

177.00	460.0
178.00	159.0
179.00	2340.0
180.00	1415.0
181.00	512.0
185.00	917.0
186.00	6522.0
187.00	1821.0
187.80	176.0
188.90	415.0
190.90	246.0
192.00	800.0
193.00	933.0
193.90	173.0
196.00	1121.0
197.00	351.0
198.00	37736.0
199.00	3171.0
199.90	250.0
202.90	511.0
204.00	2111.0
205.00	3381.0
206.00	11563.0
207.00	1264.0
208.00	714.0
209.00	163.0
209.90	277.0
211.00	311.0
214.90	219.0
216.10	348.0
216.90	4367.0
217.90	621.0
221.00	948.0
221.90	486.0
223.00	769.0
224.00	7221.0
225.00	1892.0
227.00	4421.0
228.00	721.0
229.00	736.0
230.90	265.0
233.90	252.0
235.00	236.0
236.80	302.0
239.90	275.0
240.90	239.0
242.00	413.0
243.00	480.0
244.00	4907.0
245.10	771.0
245.90	1410.0
246.90	374.0
249.00	161.0
253.00	233.0
253.80	289.0
254.20	255.0
255.00	24872.0
256.00	4029.0
257.00	304.0
258.00	1986.0
259.00	387.0
264.90	890.0
273.00	758.0
274.00	2244.0
275.00	10257.0
276.00	1291.0
277.00	1153.0
277.80	240.0
285.00	233.0
292.80	240.0
296.00	4256.0
297.00	597.0
302.90	365.0
314.90	459.0
323.00	939.0
324.00	157.0
326.90	183.0
334.10	589.0
335.00	205.0
341.00	151.0
345.90	271.0
352.10	264.0
353.00	188.0
354.00	261.0
364.90	1826.0
365.90	267.0
370.90	150.0
371.90	359.0
402.10	186.0
402.90	231.0
420.90	184.0
422.00	249.0
423.00	1255.0
424.00	361.0
441.00	4401.0
442.00	30520.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

443.006102.0

444.00495.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

m/z	Abundance
38.10	349.0
39.00	1526.0
48.90	200.0
50.00	6920.0
51.10	21616.0
52.00	927.0
56.00	821.0
57.00	1251.0
60.90	297.0
61.90	244.0
63.10	1021.0
64.90	244.0
68.20	347.0
69.00	20168.0
73.00	215.0
74.00	2283.0
75.00	3597.0
76.00	969.0
77.10	19744.0
78.00	1314.0
79.00	1728.0
80.00	1011.0
81.00	2393.0
82.00	443.0
82.90	254.0
85.00	229.0
85.90	551.0
87.00	244.0
91.00	399.0
92.00	475.0
93.00	4143.0
93.90	221.0
98.00	2783.0
99.00	1798.0
101.00	869.0
103.00	253.0
104.00	682.0
105.00	611.0
107.00	7020.0
108.00	900.0
109.00	194.0
110.00	9624.0
111.00	1527.0
112.00	339.0
116.10	501.0
117.00	9231.0
118.00	566.0
121.90	505.0
122.90	816.0
123.90	290.0
125.00	263.0
127.00	18728.0
128.00	1454.0
129.00	9154.0
129.90	931.0
131.80	156.0
133.90	295.0
134.90	737.0
135.90	384.0
137.00	358.0
141.00	1106.0
142.00	386.0
143.00	314.0
145.90	175.0
147.00	748.0
148.00	1953.0
149.00	382.0
149.90	189.0
152.90	385.0
154.00	389.0
155.00	660.0
156.00	1079.0
156.90	241.0
157.90	299.0
159.00	192.0
160.10	351.0
161.00	530.0
162.00	239.0
165.00	514.0
165.80	347.0
167.00	3278.0
168.00	1369.0
169.00	333.0
172.00	179.0
173.00	357.0
173.90	615.0
175.10	1094.0
176.10	193.0
177.00	515.0
177.90	216.0
179.00	2108.0
180.00	1181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841

181.00	429.0
185.00	831.0
186.00	6627.0
187.00	1888.0
188.00	151.0
189.00	291.0
191.00	418.0
192.00	536.0
193.00	966.0
196.00	1229.0
197.10	409.0
197.90	36448.0
198.90	2287.0
199.90	353.0
203.00	383.0
204.00	2018.0
205.00	3146.0
206.00	11820.0
207.00	1720.0
208.00	801.0
209.90	229.0
210.90	435.0
215.90	429.0
217.00	4021.0
217.90	376.0
221.00	1050.0
222.00	441.0
223.00	822.0
224.00	6643.0
225.00	1730.0
226.00	269.0
227.00	4111.0
228.00	581.0
228.90	696.0
230.80	173.0
234.00	205.0
235.00	170.0
236.00	232.0
237.10	164.0
241.90	343.0
243.00	433.0
244.00	5048.0
245.00	607.0
245.90	1340.0
246.90	516.0
248.80	172.0
254.20	367.0
255.00	27792.0
256.00	4152.0
256.90	240.0
258.00	1955.0
259.00	299.0
265.00	922.0
266.00	162.0
272.90	754.0
274.00	2475.0
275.00	11202.0
275.90	1471.0
277.00	1353.0
277.90	255.0
282.90	265.0
284.90	151.0
293.00	215.0
296.00	4489.0
296.90	438.0
302.90	414.0
315.00	276.0
315.90	192.0
323.00	1169.0
327.00	257.0
332.90	165.0
333.90	676.0
334.90	236.0
346.00	212.0
352.00	520.0
353.10	237.0
354.00	254.0
365.00	2141.0
365.90	358.0
371.90	533.0
401.90	416.0
403.00	356.0
404.00	206.0
421.00	437.0
421.90	219.0
423.00	1914.0
424.00	427.0
441.00	6212.0
442.00	39984.0
443.00	7143.0
444.00	641.0

Instrument :
BNA_M
ClientSampleId :
DFTPP841