


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110623\
Data File : BM042598.D
Acq On : 06 Nov 2023 20:31
Operator : MA/JU
Sample : DFTPP179
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP179

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Nov 07 02:23:03 2023



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.2	86976	PASS
68	69	0.00	2	1.6	1193	PASS
69	198	0.00	100	46.0	73899	PASS
70	69	0.00	2	0.7	486	PASS
127	198	10	80	44.8	71867	PASS
197	198	0.00	2	0.4	635	PASS
198	198	100	100	100.0	160512	PASS
199	198	5	9	6.6	10548	PASS
275	198	10	60	27.2	43683	PASS
365	198	1	100	4.0	6462	PASS
441	198	0.01	100	15.3	24573	PASS
442	198	50	100	99.4	159581	PASS
443	442	15	24	19.6	31256	PASS

m/z	Abundance
39.00	194.0
40.00	481.0
41.00	235.0
44.00	1296.0
73.00	498.0
77.00	188.0
84.10	152.0
132.90	201.0
192.90	175.0
207.10	1255.0
208.00	162.0
209.00	186.0
281.00	222.0

Instrument :

BNA_M

ClientSampleId :

DFTPP179

m/z	Abundance
37.00	376.0
38.10	1381.0
39.10	7793.0
40.00	976.0
41.00	492.0
43.00	249.0
44.00	1327.0
45.00	256.0
49.00	805.0
50.10	26448.0
51.10	94424.0
52.10	5083.0
53.00	257.0
55.00	625.0
56.00	2438.0
57.00	5367.0
57.90	253.0
61.00	926.0
62.00	1298.0
63.10	3227.0
64.00	527.0
65.00	1492.0
67.10	169.0
68.00	1424.0
69.00	80112.0
70.00	460.0
73.00	1197.0
74.10	9337.0
75.00	13048.0
76.10	5122.0
77.10	91848.0
78.10	6904.0
79.00	6072.0
80.00	4337.0
81.00	6712.0
82.00	1512.0
83.10	1338.0
83.90	266.0
85.00	959.0
86.00	2028.0
87.00	796.0
88.00	314.0
91.00	1465.0
92.00	1664.0
93.00	11622.0
94.00	715.0
95.00	150.0
96.00	524.0
97.20	211.0
98.00	8457.0
99.00	6157.0
100.00	677.0
101.00	3313.0
102.00	166.0
103.00	798.0
104.00	2102.0
105.00	1876.0
106.00	655.0
107.00	25064.0
108.10	3802.0
108.90	492.0
110.00	38192.0
111.00	6345.0
112.00	996.0
112.90	272.0
116.00	1450.0
117.00	29176.0
118.00	2139.0
119.00	312.0
120.00	308.0
122.00	1458.0
123.00	2386.0
124.00	1114.0
125.00	978.0
126.10	273.0
127.10	71760.0
128.00	5881.0
129.00	30656.0
130.00	2329.0
130.90	491.0
131.90	256.0
132.90	328.0
133.90	825.0
135.00	2669.0
136.00	1091.0
137.00	1318.0
140.00	369.0
141.00	3815.0
142.00	1389.0
143.00	1004.0
144.00	375.0
144.90	208.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

146.00	855.0
147.00	1971.0
148.00	6394.0
149.00	1456.0
150.00	284.0
151.00	511.0
151.90	273.0
153.00	1413.0
154.00	960.0
155.00	2435.0
156.00	3313.0
157.10	554.0
157.90	806.0
159.00	740.0
160.00	1486.0
161.00	1915.0
162.00	661.0
163.00	211.0
163.90	176.0
165.00	1499.0
166.00	1347.0
167.00	11002.0
168.00	6649.0
169.00	802.0
170.00	327.0
171.10	426.0
172.00	767.0
173.00	1100.0
174.00	1790.0
175.10	3492.0
176.00	713.0
177.00	1344.0
178.00	424.0
179.00	6489.0
180.00	4612.0
181.00	2113.0
182.00	328.0
183.00	150.0
184.00	577.0
185.00	3285.0
186.10	22824.0
187.10	6623.0
188.00	645.0
189.00	1609.0
190.00	333.0
191.10	891.0
192.00	2133.0
193.00	2578.0
194.00	538.0
195.10	321.0
196.10	3756.0
196.80	652.0
198.00	145600.0
199.00	9710.0
200.00	696.0
201.40	534.0
203.00	1441.0
204.10	6304.0
205.00	9966.0
206.10	42016.0
207.10	7074.0
208.00	1844.0
209.00	659.0
210.00	615.0
211.00	1616.0
215.00	392.0
216.00	1090.0
217.00	12951.0
218.00	1836.0
220.10	204.0
221.10	4434.0
222.00	1800.0
223.00	2557.0
224.10	23928.0
225.10	6087.0
226.10	687.0
227.00	10949.0
228.00	1703.0
229.00	2056.0
230.00	260.0
231.00	817.0
233.00	159.0
234.00	735.0
235.00	817.0
236.00	599.0
237.00	892.0
239.00	337.0
240.00	465.0
241.00	699.0
242.00	1468.0
243.10	1351.0
244.10	17656.0
245.10	2437.0
246.00	4122.0
247.00	856.0
248.00	233.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

249.00	517.0
251.90	201.0
252.90	581.0
254.00	642.0
255.10	88872.0
256.10	13492.0
257.00	1143.0
258.00	5842.0
259.00	879.0
260.00	187.0
264.00	192.0
265.00	2609.0
265.90	300.0
271.00	400.0
272.10	244.0
273.00	3150.0
274.00	6776.0
275.10	36176.0
276.10	5546.0
277.00	3588.0
277.90	535.0
281.00	319.0
282.10	205.0
283.00	501.0
284.00	307.0
285.10	655.0
292.00	163.0
293.00	737.0
294.00	212.0
295.00	269.0
296.00	13427.0
297.00	1809.0
301.00	233.0
303.10	1240.0
304.00	449.0
308.10	171.0
313.00	164.0
314.00	619.0
315.00	1560.0
316.00	728.0
321.10	399.0
322.10	220.0
323.10	3097.0
324.10	367.0
327.00	737.0
328.00	330.0
332.00	307.0
333.00	300.0
334.00	1891.0
335.00	528.0
341.00	365.0
345.90	591.0
346.80	182.0
352.10	1174.0
353.10	712.0
354.10	1059.0
355.10	183.0
365.10	5249.0
366.00	606.0
371.00	233.0
372.10	1531.0
373.10	434.0
383.10	390.0
390.10	155.0
402.10	608.0
403.10	660.0
404.00	363.0
421.10	685.0
422.10	804.0
423.10	5196.0
424.10	1307.0
425.10	190.0
441.10	16672.0
442.10	106712.0
443.10	21904.0
444.10	1824.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

m/z	Abundance
37.00	559.0
38.10	1586.0
39.10	7592.0
40.00	777.0
41.10	431.0
43.10	209.0
44.00	1146.0
45.00	376.0
49.00	726.0
50.10	29272.0
51.10	108376.0
52.10	5102.0
53.00	247.0
55.00	709.0
56.00	3130.0
57.00	6107.0
58.00	249.0
61.00	1076.0
62.00	1383.0
63.00	3523.0
64.00	563.0
65.10	1645.0
67.00	174.0
68.00	1442.0
69.00	90896.0
70.10	630.0
73.10	1486.0
74.10	10515.0
75.00	15737.0
76.10	5299.0
77.10	106352.0
78.10	8223.0
79.00	7021.0
80.00	5294.0
81.00	7535.0
82.00	1986.0
83.00	1479.0
84.00	345.0
85.00	1186.0
86.10	2201.0
87.00	1113.0
88.00	301.0
89.10	217.0
91.00	1606.0
92.00	1752.0
93.00	13667.0
94.00	758.0
94.90	306.0
96.00	529.0
97.00	227.0
98.00	11023.0
99.00	7191.0
100.00	554.0
101.10	3808.0
102.00	271.0
103.00	1273.0
104.00	2577.0
105.00	2336.0
106.00	872.0
107.00	31920.0
108.10	4565.0
109.10	732.0
110.00	47496.0
111.00	7809.0
112.00	868.0
112.90	256.0
116.00	1638.0
117.00	35976.0
118.00	2601.0
119.00	358.0
120.00	342.0
122.00	1774.0
123.00	2657.0
124.00	1539.0
125.00	1297.0
126.20	459.0
127.10	90928.0
128.00	7150.0
129.00	38224.0
130.00	3432.0
131.00	791.0
132.00	230.0
133.00	407.0
134.00	1396.0
135.00	3286.0
136.00	1340.0
137.00	1738.0
138.00	422.0
138.90	267.0
139.90	528.0
141.00	5444.0
142.00	1765.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

143.10	1335.0
144.00	314.0
144.90	293.0
146.00	1236.0
147.00	3070.0
148.00	8620.0
149.00	1747.0
150.00	413.0
151.10	610.0
151.90	338.0
153.00	1660.0
154.00	1476.0
155.00	3081.0
156.00	4091.0
157.00	628.0
158.00	810.0
159.00	756.0
160.00	1680.0
161.00	2917.0
162.00	767.0
163.00	212.0
164.00	259.0
165.00	1993.0
166.00	1674.0
167.00	14824.0
168.00	7782.0
169.10	1067.0
170.10	468.0
171.00	440.0
172.00	1066.0
173.00	1472.0
174.00	2223.0
175.10	4631.0
176.00	1223.0
177.00	1948.0
177.90	734.0
179.00	9247.0
180.00	6485.0
181.00	3003.0
182.00	431.0
182.90	206.0
183.90	735.0
185.00	4324.0
186.10	29432.0
187.10	8820.0
188.00	1186.0
189.00	2179.0
189.90	388.0
191.00	1163.0
192.00	3414.0
193.00	3382.0
194.00	738.0
195.00	460.0
196.10	5136.0
196.80	1078.0
197.10	1069.0
198.00	206464.0
199.00	13839.0
200.00	1177.0
201.30	600.0
203.00	2007.0
204.10	8884.0
205.10	14527.0
206.10	57912.0
207.10	8724.0
208.00	2467.0
209.00	981.0
210.00	852.0
211.10	2178.0
211.80	336.0
213.10	235.0
215.00	694.0
216.10	1594.0
217.00	17784.0
218.00	2489.0
219.10	179.0
219.90	271.0
221.10	7377.0
222.00	2173.0
223.00	3609.0
224.10	33224.0
225.10	9263.0
226.00	1006.0
227.00	16408.0
228.00	2266.0
229.00	3234.0
230.00	341.0
231.00	1116.0
232.10	190.0
233.10	216.0
234.00	1066.0
235.10	1109.0
236.00	925.0
237.10	1448.0
238.00	279.0
239.00	549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

239.90	517.0
241.00	784.0
242.00	1689.0
243.10	2231.0
244.10	27016.0
245.10	4043.0
246.00	6238.0
247.00	1236.0
248.10	235.0
249.00	1043.0
250.10	269.0
251.10	293.0
251.90	189.0
253.00	786.0
254.10	734.0
255.10	131968.0
256.10	18992.0
257.10	1524.0
258.00	8549.0
259.00	1406.0
260.10	238.0
261.10	183.0
264.00	219.0
265.00	3601.0
266.10	450.0
271.00	234.0
272.10	299.0
273.10	4089.0
274.00	10783.0
275.10	55176.0
276.00	7365.0
277.00	5871.0
278.00	1124.0
279.10	235.0
281.10	441.0
282.00	284.0
283.00	789.0
284.00	562.0
285.10	972.0
286.00	254.0
288.80	243.0
290.00	184.0
291.10	203.0
293.00	1293.0
294.00	269.0
295.00	355.0
296.10	18712.0
297.00	2997.0
298.10	223.0
301.10	152.0
302.00	282.0
303.10	1916.0
304.10	579.0
308.00	219.0
310.10	180.0
313.00	241.0
314.00	841.0
315.00	2055.0
316.00	891.0
321.10	453.0
322.00	190.0
323.10	5541.0
324.10	1135.0
327.00	1210.0
328.10	504.0
329.10	188.0
332.00	355.0
333.10	615.0
334.10	3680.0
335.10	1138.0
341.10	625.0
342.00	237.0
346.00	1248.0
347.00	241.0
352.10	1469.0
353.10	904.0
354.10	1538.0
355.20	322.0
359.10	150.0
365.10	8260.0
366.00	1194.0
370.00	210.0
371.10	564.0
372.10	2780.0
373.10	745.0
383.00	752.0
384.00	242.0
390.10	511.0
391.10	322.0
401.00	197.0
402.00	1028.0
403.10	1449.0
404.10	469.0
421.10	1311.0
422.10	1417.0
423.10	8567.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

424.10	2033.0
425.00	278.0
441.10	29528.0
442.10	194752.0
443.10	38512.0
444.10	3220.0
445.10	241.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

m/z	Abundance
37.10	280.0
38.00	851.0
39.10	4576.0
40.00	594.0
41.10	220.0
43.20	194.0
44.00	1571.0
45.00	200.0
49.10	639.0
50.10	16464.0
51.10	58128.0
52.10	2953.0
54.90	446.0
56.00	1835.0
57.00	3403.0
57.90	185.0
61.00	710.0
62.00	810.0
63.00	2339.0
63.90	291.0
65.00	1042.0
68.10	714.0
69.00	50688.0
70.00	369.0
73.00	973.0
74.00	6350.0
75.00	9113.0
76.10	3261.0
77.10	61992.0
78.10	4257.0
79.00	3566.0
80.10	2958.0
81.00	4142.0
82.00	1276.0
83.10	829.0
84.00	161.0
85.00	706.0
86.00	1299.0
87.00	468.0
91.00	991.0
92.00	1016.0
93.00	7801.0
94.00	462.0
95.00	192.0
96.00	408.0
98.00	6455.0
99.00	4263.0
99.90	263.0
101.00	2096.0
103.10	741.0
104.00	1369.0
105.00	1672.0
106.00	390.0
107.00	17744.0
108.00	2839.0
109.00	497.0
110.00	27784.0
111.00	4484.0
112.00	621.0
113.00	241.0
116.10	953.0
117.00	20784.0
118.00	1383.0
118.90	284.0
119.90	215.0
122.00	1314.0
123.00	1887.0
124.00	881.0
125.00	675.0
126.20	224.0
127.00	52912.0
128.00	3784.0
129.00	23104.0
130.00	1804.0
131.00	418.0
133.00	223.0
134.00	589.0
135.00	2046.0
136.00	609.0
137.00	1050.0
138.00	307.0
140.10	206.0
141.00	3186.0
142.00	1032.0
143.00	851.0
144.00	166.0
146.00	560.0
147.10	2073.0
148.00	4888.0
149.10	874.0
150.00	198.0
151.10	562.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

151.90	189.0
153.00	1177.0
154.00	769.0
155.00	1884.0
156.00	2382.0
157.20	461.0
158.00	612.0
159.00	544.0
159.90	1102.0
161.00	1763.0
162.00	504.0
163.10	221.0
163.80	171.0
165.00	1249.0
166.00	1102.0
167.00	9896.0
168.00	5083.0
169.00	624.0
170.10	274.0
171.00	266.0
172.10	697.0
173.00	871.0
174.00	1502.0
175.00	2911.0
176.00	465.0
177.00	1409.0
177.90	512.0
179.00	6157.0
180.00	3552.0
181.00	1654.0
182.00	347.0
183.90	452.0
185.10	2574.0
186.10	19632.0
187.10	5488.0
188.00	657.0
189.00	1326.0
190.00	301.0
191.00	838.0
192.00	2163.0
193.00	2532.0
194.00	489.0
195.00	229.0
196.00	3397.0
197.10	837.0
198.00	129472.0
199.00	8096.0
200.00	884.0
201.60	452.0
203.00	1197.0
204.10	6122.0
205.10	9556.0
206.10	38592.0
207.10	6660.0
208.00	1629.0
209.00	795.0
210.10	656.0
211.10	1741.0
211.80	273.0
215.00	522.0
216.00	954.0
217.00	11303.0
218.00	1788.0
220.10	200.0
221.00	4868.0
221.90	1345.0
223.00	2569.0
224.10	22864.0
225.10	5957.0
226.00	598.0
227.00	10854.0
228.00	1592.0
229.00	2332.0
230.00	320.0
231.00	859.0
233.00	206.0
234.00	719.0
235.00	661.0
236.00	514.0
237.10	901.0
239.00	410.0
240.00	343.0
241.00	532.0
242.00	1304.0
243.10	1372.0
244.10	17784.0
245.10	2420.0
246.00	4191.0
247.00	999.0
248.00	252.0
249.00	639.0
251.00	179.0
252.10	222.0
253.00	529.0
254.00	716.0
255.10	92312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179

256.00	13818.0
257.00	1246.0
258.00	6017.0
259.00	1059.0
261.00	176.0
265.00	2983.0
266.00	350.0
267.00	174.0
270.90	180.0
272.00	294.0
273.10	2867.0
274.00	7840.0
275.10	39696.0
276.10	5345.0
277.00	3897.0
278.10	805.0
279.10	179.0
281.00	305.0
282.00	192.0
283.00	505.0
283.90	292.0
285.00	622.0
286.10	180.0
289.00	189.0
289.90	169.0
292.00	194.0
293.00	953.0
293.90	299.0
295.00	268.0
296.00	14248.0
297.00	1794.0
298.00	173.0
301.00	307.0
302.00	188.0
303.10	1530.0
304.00	404.0
307.90	255.0
309.90	260.0
314.10	607.0
315.00	1742.0
316.00	662.0
321.00	307.0
322.10	187.0
323.10	4012.0
324.00	869.0
327.00	780.0
328.00	528.0
332.10	219.0
333.10	459.0
334.10	2776.0
335.10	648.0
341.00	521.0
342.10	179.0
346.10	821.0
347.10	280.0
352.10	1375.0
353.10	897.0
354.10	1206.0
355.10	268.0
365.10	5878.0
366.10	958.0
371.00	311.0
372.10	2070.0
373.00	658.0
383.10	617.0
383.90	159.0
390.00	234.0
391.10	170.0
402.10	672.0
403.10	1056.0
403.90	528.0
421.10	1281.0
422.10	973.0
423.10	7219.0
424.10	1664.0
425.00	242.0
441.10	27520.0
442.10	177280.0
443.10	33352.0
444.10	2987.0
445.00	218.0

Instrument :
BNA_M
ClientSampleId :
DFTPP179