

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
 Data File : BM042639.D
 Acq On : 08 Nov 2023 07:57
 Operator : MA/JU
 Sample : 04948-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AG4

Quant Time: Nov 08 08:27:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 02:23:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.859	152	1081	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.671	136	2751	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.510	164	1582	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.262	188	3558	0.400	ng/ul	0.00
17) Chrysene-d12	21.450	240	2413	0.400	ng/ul	0.00
23) Perylene-d12	23.826	264	3201	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.198	96	8179	5.726	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.266	152	1235	0.322	ng/ul	0.00
18) Fluoranthene-d10	19.289	212	2895	0.347	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
Data File : BM042639.D
Acq On : 08 Nov 2023 07:57
Operator : MA/JU
Sample : 04948-02
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AG4

Quant Time: Nov 08 08:27:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 02:23:03 2023
Response via : Initial Calibration

