

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
 Data File : BM042662.D
 Acq On : 08 Nov 2023 22:47
 Operator : MA/JU
 Sample : PB156830BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK830

Quant Time: Nov 08 23:20:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 02:23:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.855	152	1131	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.666	136	2485	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.510	164	1394	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.262	188	3112	0.400	ng/ul	0.00
17) Chrysene-d12	21.453	240	2090	0.400	ng/ul	0.00
23) Perylene-d12	23.820	264	3460	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.197	96	7144	4.780	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.271	152	878	0.253	ng/ul	0.00
18) Fluoranthene-d10	19.289	212	1923	0.266	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
Data File : BM042662.D
Acq On : 08 Nov 2023 22:47
Operator : MA/JU
Sample : PB156830BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK830

Quant Time: Nov 08 23:20:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 02:23:03 2023
Response via : Initial Calibration

