

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
 Data File : BM042637.D
 Acq On : 08 Nov 2023 06:45
 Operator : MA/JU
 Sample : 05026-18
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH341

Quant Time: Nov 08 07:17:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 02:23:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.859	152	1423	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.671	136	3561	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.510	164	2086	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.262	188	4851	0.400	ng/ul	0.00
17) Chrysene-d12	21.447	240	3138	0.400	ng/ul	0.00
23) Perylene-d12	23.820	264	4102	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.202	96	7138	3.796	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.260	152	1952	0.393	ng/ul	-0.01
18) Fluoranthene-d10	19.289	212	5137	0.473	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	5304	2.373	ng/ul#	56


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110723\
Data File : BM042637.D
Acq On : 08 Nov 2023 06:45
Operator : MA/JU
Sample : 05026-18
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH341

Quant Time: Nov 08 07:17:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 02:23:03 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 2.373 ng/ul
 RT: 3.240 min Scan# 61
 Delta R.T. -0.006 min
 Lab File: BM042637.D
 Acq: 08 Nov 2023 06:45

Instrument :
 BNA_M
 ClientSampleId :
 BH341

Tgt Ion: 88 Resp: 5304
 Ion Ratio Lower Upper
 88 100
 43 48.3 77.5 116.3#
 58 74.3 40.9 61.3#

