



Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM110923\  
 Data File : BM042667.D  
 Acq On : 09 Nov 2023 11:17  
 Operator : MA/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM103023.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Nov 09 13:21:00 2023






AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2482

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 58.9         | 46387      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 730        | PASS                |
| 69             | 198             | 0.00            | 100             | 48.0         | 37827      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 224        | PASS                |
| 127            | 198             | 10              | 80              | 46.7         | 36779      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 298        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 78779      | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 5329       | PASS                |
| 275            | 198             | 10              | 60              | 28.2         | 22245      | PASS                |
| 365            | 198             | 1               | 100             | 4.7          | 3726       | PASS                |
| 441            | 198             | 0.01            | 100             | 16.6         | 13093      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 79571      | PASS                |
| 443            | 442             | 15              | 24              | 19.1         | 15185      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM110923\  
Data File : BM042667.D  
Acq On : 09 Nov 2023 11:17  
Operator : MA/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP

Quant Time: Nov 09 19:27:55 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM103023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 09 13:21:00 2023  
Response via : Initial Calibration



TIC: BM042667.D\data.ms

**(70) Pentachlorophenol (C)**

**16.886min (-0.000) 0.00 ng**




**response 110876**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.60  | 63.63  |
| 264.00 | 61.80  | 62.28  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM110923\  
 Data File : BM042667.D  
 Acq On : 09 Nov 2023 11:17  
 Operator : MA/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Quant Time: Nov 09 19:27:55 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\8270-BM103023.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 09 13:21:00 2023  
 Response via : Initial Calibration



TIC: BM042667.D\data.ms

**(77) Benzidine****19.533min (-0.000) 32283.70 ng****response 526756**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.00  | 14.10  |
| 183.00 | 11.20  | 11.73  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 11/9/2023     | BNA_M            | <u>BM042667.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 415310           | 20.715             |
| DDD           | 5528             | 20.321             |
| DDE           | 127              | 19.774             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 5655          | 420965           | 1.34               |
|               |                  |                    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP