



Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM111018\
Data File : BM017590.D
Acq On : 09 Nov 2018 14:00
Operator : MJ/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 01:56:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 15:08:36 2018
Response via : Initial Calibration



TIC: BM017590.D

(77) Benzidine



19.286min (-0.018) 282119.82ng

response 1588093

Ion	Exp%	Act%
184.00	100	100
185.00	14.80	13.82
183.00	12.10	11.51
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM111018\
Data File : BM017590.D
Acq On : 09 Nov 2018 14:00
Operator : MJ/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 01:56:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM101718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 15:08:36 2018
Response via : Initial Calibration



TIC: BM017590.D

(70) Pentachlorophenol (C)

16.669min (-0.029) 0.00ng

response 269885

Ion	Exp%	Act%
265.70	100	100
268.00	63.40	63.83
264.00	63.00	61.78
0.00	0.00	0.00