


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111323\
Data File : BM042703.D
Acq On : 13 Nov 2023 12:03
Operator : MA/JU
Sample : DFTPP181
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP181

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM111323.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Nov 14 00:02:28 2023



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.2	38021	PASS
68	69	0.00	2	1.7	538	PASS
69	198	0.00	100	46.1	31771	PASS
70	69	0.00	2	0.5	159	PASS
127	198	10	80	44.7	30832	PASS
197	198	0.00	2	0.4	269	PASS
198	198	100	100	100.0	68901	PASS
199	198	5	9	6.6	4559	PASS
275	198	10	60	28.0	19326	PASS
365	198	1	100	4.8	3285	PASS
441	198	0.01	100	13.7	9432	PASS
442	198	50	100	87.5	60317	PASS
443	442	15	24	18.4	11081	PASS

m/z	Abundance
36.00	265.0
36.90	521.0
40.10	258.0
41.20	187.0
42.10	164.0
43.00	226.0
44.00	590.0
47.00	1267.0
48.00	658.0
49.00	6259.0
50.00	252.0
51.00	1838.0
73.00	306.0
83.00	151.0
84.00	3627.0
85.00	187.0
86.00	2131.0
87.90	376.0
207.10	625.0
208.00	178.0
281.00	230.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
36.00	288.0
37.00	817.0
38.10	889.0
39.10	4073.0
40.00	471.0
41.00	569.0
42.00	165.0
43.00	342.0
44.00	951.0
45.00	179.0
47.00	1600.0
48.10	899.0
49.00	6783.0
50.10	15359.0
51.10	52304.0
52.10	3039.0
53.10	183.0
55.00	318.0
56.00	1283.0
57.00	2869.0
58.10	164.0
61.00	439.0
61.90	570.0
63.10	1577.0
64.10	263.0
65.10	781.0
68.00	688.0
69.00	41344.0
70.00	221.0
73.10	750.0
74.10	5106.0
75.10	7296.0
76.10	2712.0
77.10	49576.0
78.10	3419.0
79.00	3346.0
80.00	2744.0
81.00	3192.0
82.00	906.0
83.00	785.0
84.00	4046.0
85.00	544.0
86.00	3642.0
87.00	617.0
88.10	562.0
91.00	775.0
92.00	885.0
93.00	6525.0
94.00	473.0
96.00	279.0
98.00	4851.0
99.00	3302.0
100.00	319.0
101.00	1749.0
102.90	392.0
104.00	1102.0
105.00	1343.0
106.10	398.0
107.00	14018.0
108.00	2256.0
109.10	396.0
110.00	20360.0
111.00	3355.0
112.00	338.0
113.00	204.0
116.00	731.0
117.00	16392.0
118.00	1423.0
119.00	231.0
120.00	191.0
122.00	1009.0
123.10	1305.0
124.00	760.0
125.00	591.0
126.20	176.0
127.10	37304.0
128.00	3016.0
129.00	16010.0
130.00	1484.0
131.10	287.0
134.10	439.0
135.00	1462.0
136.00	595.0
137.10	986.0
138.00	219.0
140.00	191.0
141.00	2149.0
142.00	636.0
143.00	496.0
144.10	191.0
146.00	691.0
147.00	1357.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

148.00	4005.0
149.10	714.0
151.00	377.0
151.80	179.0
153.00	623.0
154.00	542.0
155.00	1397.0
156.10	1890.0
157.00	415.0
158.10	464.0
159.00	268.0
160.00	797.0
161.10	942.0
162.10	336.0
165.00	999.0
166.00	749.0
167.00	6724.0
168.00	3888.0
169.00	462.0
170.90	245.0
172.00	450.0
173.00	553.0
174.00	1083.0
175.10	1986.0
176.00	457.0
177.00	893.0
178.00	276.0
179.00	3998.0
180.00	2618.0
181.10	1116.0
182.10	181.0
183.90	337.0
185.10	1801.0
186.10	13094.0
187.00	4146.0
188.00	475.0
189.00	913.0
190.00	163.0
191.00	468.0
192.00	1435.0
193.00	1682.0
194.00	308.0
196.00	1884.0
198.00	79160.0
199.00	5351.0
200.00	525.0
201.40	322.0
203.00	750.0
204.00	3676.0
205.10	6103.0
206.10	23616.0
207.10	3934.0
208.00	1197.0
209.00	423.0
210.10	422.0
211.00	1182.0
211.70	211.0
212.10	169.0
215.00	298.0
216.00	603.0
217.00	7366.0
218.00	887.0
221.00	2660.0
222.00	1089.0
223.00	1727.0
224.10	13967.0
225.10	3373.0
226.00	471.0
227.00	6551.0
228.00	1019.0
229.00	1388.0
230.00	199.0
231.00	558.0
234.00	376.0
235.00	509.0
235.90	367.0
237.00	602.0
239.00	307.0
240.00	335.0
241.00	367.0
242.00	666.0
243.10	628.0
244.10	10128.0
245.10	1317.0
246.00	2202.0
247.00	593.0
249.00	315.0
253.00	315.0
254.00	442.0
255.10	51568.0
256.10	7083.0
257.00	553.0
258.00	3689.0
259.10	712.0
265.00	1461.0
266.00	220.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

272.00	172.0
273.00	1554.0
274.00	4029.0
275.10	20392.0
276.10	2952.0
277.00	2127.0
278.00	289.0
281.10	174.0
283.10	331.0
285.00	439.0
293.00	409.0
293.90	178.0
296.00	7537.0
297.00	847.0
303.00	723.0
304.00	243.0
314.10	321.0
315.00	806.0
316.00	475.0
321.00	295.0
323.10	2170.0
324.00	416.0
327.00	393.0
328.10	261.0
332.00	150.0
333.10	157.0
334.10	1362.0
335.00	262.0
340.90	161.0
346.00	565.0
352.10	553.0
353.00	468.0
354.00	626.0
365.10	3189.0
366.00	388.0
371.00	173.0
372.10	1047.0
373.20	287.0
383.00	246.0
402.10	394.0
403.20	443.0
404.00	269.0
421.10	521.0
422.10	428.0
423.10	2845.0
424.00	713.0
441.10	8293.0
442.10	52632.0
443.10	9751.0
444.10	961.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
36.00	414.0
36.90	685.0
38.00	887.0
39.10	3678.0
40.00	383.0
41.00	429.0
42.00	212.0
44.00	938.0
47.00	1996.0
48.00	752.0
49.10	6446.0
50.10	12205.0
51.10	46000.0
52.00	1776.0
55.00	279.0
56.00	1109.0
57.00	2273.0
61.00	490.0
62.00	638.0
63.00	1617.0
64.00	234.0
65.10	701.0
68.00	619.0
69.00	37392.0
70.00	257.0
73.10	477.0
74.10	4759.0
75.10	6607.0
76.10	2287.0
77.10	45760.0
78.10	2998.0
79.00	2765.0
80.10	2102.0
81.10	3505.0
82.00	743.0
83.00	887.0
84.00	3725.0
85.00	508.0
86.00	3443.0
87.00	476.0
88.00	524.0
91.00	513.0
92.00	701.0
93.00	5691.0
94.00	377.0
96.00	197.0
98.00	4480.0
99.00	3170.0
100.10	298.0
101.00	1462.0
103.00	450.0
104.00	1033.0
105.10	1048.0
106.00	357.0
107.00	13266.0
108.00	2135.0
109.10	304.0
110.00	18344.0
111.00	3161.0
112.00	322.0
113.10	190.0
116.00	628.0
117.00	16416.0
118.00	1117.0
119.00	165.0
119.90	174.0
122.00	928.0
123.00	1482.0
124.00	705.0
125.00	659.0
126.00	182.0
127.00	37272.0
128.00	2865.0
129.00	16010.0
130.00	1292.0
131.00	353.0
134.00	373.0
135.00	1351.0
136.10	605.0
137.00	744.0
137.90	261.0
140.00	205.0
141.00	2095.0
142.00	729.0
143.00	458.0
144.00	157.0
146.00	537.0
147.00	1178.0
148.00	3888.0
149.00	714.0
151.10	417.0
152.90	874.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

153.90	531.0
155.00	1242.0
156.00	1663.0
157.00	396.0
157.90	315.0
159.00	442.0
159.90	600.0
161.00	1014.0
162.00	315.0
165.00	1132.0
166.00	734.0
167.10	6995.0
168.00	3390.0
169.00	417.0
170.90	220.0
171.90	653.0
173.00	687.0
174.00	1106.0
175.00	1797.0
176.00	569.0
177.00	788.0
177.90	348.0
179.00	4422.0
180.00	2614.0
181.00	1305.0
182.00	163.0
184.00	447.0
185.00	1742.0
186.10	13497.0
187.10	3916.0
188.10	329.0
189.00	947.0
189.90	151.0
191.00	593.0
192.00	1402.0
193.10	1696.0
194.00	394.0
195.00	162.0
196.00	2105.0
197.10	467.0
198.00	84272.0
199.00	5550.0
200.00	481.0
201.50	506.0
203.00	660.0
204.10	4123.0
205.00	6338.0
206.10	24000.0
207.10	4250.0
208.00	1067.0
209.00	372.0
211.00	1004.0
211.80	236.0
215.00	376.0
216.00	731.0
217.00	7687.0
218.10	1068.0
221.00	2811.0
222.00	804.0
223.00	1788.0
224.10	15654.0
225.10	3627.0
226.10	429.0
227.00	7315.0
228.00	919.0
229.00	1352.0
231.10	544.0
234.00	470.0
235.00	435.0
236.00	446.0
237.00	516.0
238.90	355.0
240.00	201.0
241.00	331.0
242.00	947.0
243.10	988.0
244.10	12211.0
245.10	1420.0
246.10	2669.0
247.00	609.0
249.00	542.0
253.00	233.0
254.00	352.0
255.00	57272.0
256.10	8489.0
257.10	663.0
258.00	4127.0
259.00	717.0
265.10	1966.0
266.00	281.0
271.10	186.0
272.00	213.0
273.00	1837.0
274.10	4906.0
275.10	24120.0
276.10	3265.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

277.00	2664.0
278.10	407.0
281.10	184.0
283.00	257.0
285.00	536.0
293.00	588.0
295.00	165.0
296.10	9289.0
297.00	1392.0
301.00	155.0
303.10	1088.0
304.10	298.0
310.20	150.0
314.00	391.0
315.00	1039.0
316.10	497.0
320.90	214.0
323.10	2219.0
324.00	359.0
327.00	513.0
328.00	262.0
332.00	232.0
332.90	284.0
334.10	1620.0
335.10	448.0
341.00	211.0
346.10	575.0
352.10	914.0
353.10	614.0
354.10	783.0
355.00	229.0
365.10	4023.0
366.00	559.0
371.10	268.0
372.10	1348.0
373.00	381.0
382.90	381.0
390.10	179.0
402.00	501.0
403.00	650.0
404.10	237.0
421.10	665.0
422.10	553.0
423.10	3934.0
424.10	981.0
441.10	12112.0
442.10	75592.0
443.10	14174.0
444.10	1233.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
35.90	219.0
37.10	620.0
38.00	286.0
39.10	1372.0
40.00	465.0
41.00	324.0
43.00	172.0
44.00	756.0
47.00	1646.0
48.00	694.0
49.00	6369.0
50.10	5708.0
51.10	21272.0
52.10	1065.0
55.10	195.0
56.10	713.0
57.10	1382.0
61.00	223.0
62.00	395.0
63.10	655.0
65.00	313.0
68.00	307.0
69.00	16576.0
73.00	444.0
74.00	2033.0
75.00	3118.0
76.00	1176.0
77.10	20144.0
78.10	1312.0
79.10	1234.0
80.00	1062.0
81.00	1268.0
82.00	452.0
82.90	539.0
84.00	3791.0
85.00	362.0
86.00	2499.0
86.90	228.0
88.00	376.0
91.00	344.0
92.00	286.0
93.00	2952.0
98.00	2155.0
99.00	1710.0
101.10	833.0
102.90	280.0
103.90	471.0
105.10	550.0
106.10	187.0
107.10	5710.0
108.00	956.0
109.00	216.0
110.00	8427.0
111.10	1589.0
112.00	206.0
116.10	348.0
117.00	8023.0
118.00	475.0
119.90	190.0
122.10	451.0
123.10	557.0
124.10	343.0
124.90	266.0
127.10	17920.0
128.00	1418.0
129.00	7808.0
130.00	759.0
131.10	173.0
134.00	177.0
135.00	764.0
135.90	215.0
137.00	397.0
141.10	1296.0
142.00	362.0
142.90	160.0
144.00	169.0
146.00	186.0
147.00	620.0
148.00	1556.0
149.00	455.0
153.00	344.0
154.10	278.0
155.00	582.0
156.10	976.0
158.00	324.0
158.90	162.0
160.00	491.0
161.00	589.0
162.00	151.0
164.90	452.0
166.10	358.0
167.00	3373.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

168.00	1756.0
168.90	239.0
171.10	168.0
173.00	169.0
174.00	578.0
175.00	1073.0
176.20	247.0
177.00	632.0
178.00	292.0
179.00	1779.0
180.00	1256.0
181.00	496.0
185.10	834.0
186.10	6491.0
187.00	1805.0
188.10	222.0
188.90	490.0
189.90	165.0
191.00	274.0
192.00	625.0
193.10	914.0
194.00	210.0
196.00	1591.0
197.10	341.0
198.00	43272.0
199.00	2777.0
199.90	266.0
201.50	164.0
203.00	533.0
204.00	2045.0
205.00	3828.0
206.10	12809.0
207.10	2295.0
208.00	765.0
209.00	210.0
210.10	279.0
211.10	571.0
211.70	155.0
215.00	222.0
216.00	253.0
217.00	4307.0
218.10	545.0
221.00	1462.0
222.00	605.0
223.00	970.0
224.10	7818.0
225.00	2155.0
226.00	184.0
227.10	4361.0
228.00	542.0
229.10	925.0
231.10	198.0
233.90	285.0
234.90	155.0
237.00	467.0
241.00	227.0
242.00	610.0
243.00	438.0
244.10	6447.0
245.10	833.0
246.00	1366.0
247.00	328.0
248.80	272.0
254.10	242.0
255.10	32888.0
256.10	4847.0
256.90	416.0
258.00	2471.0
259.00	416.0
265.00	1033.0
265.90	250.0
273.00	1211.0
274.00	2931.0
275.10	13466.0
276.00	2026.0
277.00	1721.0
278.00	176.0
281.00	252.0
283.00	269.0
284.10	154.0
285.10	262.0
292.90	242.0
296.00	4842.0
297.00	602.0
303.10	683.0
314.00	270.0
315.00	670.0
316.00	270.0
323.00	1309.0
324.10	413.0
327.00	370.0
328.10	156.0
333.00	164.0
334.00	979.0
335.10	185.0
341.00	154.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

346.00	376.0
352.10	401.0
353.10	403.0
354.10	512.0
355.00	159.0
365.00	2643.0
366.10	396.0
372.00	772.0
373.10	217.0
383.00	246.0
402.10	400.0
403.00	469.0
421.00	478.0
422.10	426.0
423.10	2801.0
424.10	520.0
441.10	7891.0
442.10	52728.0
443.10	9317.0
444.10	847.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181