

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111324\
 Data File : BM048648.D
 Acq On : 13 Nov 2024 23:58
 Operator : RC/JU
 Sample : PB164898BS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS898

Quant Time: Nov 14 02:37:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 11:38:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	4661	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.474	136	15823	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.317	164	9095	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.074	188	18663	0.400	ng/ul	-0.01
17) Chrysene-d12	21.257	240	14750	0.400	ng/ul	0.00
23) Perylene-d12	23.475	264	15495	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.126	96	4626	0.823	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.086	152	8335	0.407	ng/ul	-0.02
18) Fluoranthene-d10	19.100	212	18028	0.435	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.155	88	12363	1.913	ng/ul#	61
5) Naphthalene	10.529	128	15606	0.394	ng/ul	99
7) 2-Methylnaphthalene	12.163	142	9350	0.375	ng/ul	95
8) 1-Methylnaphthalene	12.361	142	11547	0.452	ng/ul	98
10) Acenaphthylene	14.044	152	14278	0.413	ng/ul	99
11) Acenaphthene	14.382	153	10953	0.397	ng/ul	100
12) Fluorene	15.390	166	11817	0.394	ng/ul	98
14) Pentachlorophenol	16.744	266	3651	0.533	ng/ul	99
15) Phenanthrene	17.116	178	17601	0.359	ng/ul	99
16) Anthracene	17.230	178	16222	0.356	ng/ul	100
19) Fluoranthene	19.132	202	22444	0.411	ng/ul	98
20) Pyrene	19.495	202	24097	0.407	ng/ul	99
21) Benzo(a)anthracene	21.245	228	17992	0.373	ng/ul	99
22) Chrysene	21.292	228	25310	0.437	ng/ul	99
24) Benzo(b)fluoranthene	22.805	252	20746	0.430	ng/ul	91
25) Benzo(k)fluoranthene	22.852	252	24902	0.440	ng/ul#	91
26) Benzo(a)pyrene	23.384	252	19811	0.420	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	25.728	276	25989	0.395	ng/ul#	97
28) Dibenzo(a,h)anthracene	25.742	278	19789	0.386	ng/ul	95
29) Benzo(g,h,i)perylene	26.409	276	22005	0.398	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111324\
Data File : BM048648.D
Acq On : 13 Nov 2024 23:58
Operator : RC/JU
Sample : PB164898BS
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS898

Quant Time: Nov 14 02:37:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 11:38:55 2024
Response via : Initial Calibration

