

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111324\
 Data File : BM048701.D
 Acq On : 15 Nov 2024 08:59
 Operator : RC/JU
 Sample : P4761-17
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A0AN2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/15/2024
 Supervised By :mohammad ahmed 11/15/2024

Quant Time: Nov 15 10:12:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 11:38:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	4275	0.400	ng/ul	-0.08
4) Naphthalene-d8	10.436	136	13517	0.400	ng/ul	-0.06
9) Acenaphthene-d10	14.299	164	7643	0.400	ng/ul	-0.03
13) Phenanthrene-d10	17.044	188	16441m	0.400	ng/ul	-0.04
17) Chrysene-d12	21.245	240	13126m	0.400	ng/ul	-0.02
23) Perylene-d12	23.452	264	11381m	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.118	96	23858	4.626	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.053	152	4502	0.258	ng/ul	-0.05
18) Fluoranthene-d10	19.081	212	13808	0.375	ng/ul	-0.03
Target Compounds						
5) Naphthalene	10.486	128	3883	0.115	ng/ul#	90
7) 2-Methylnaphthalene	12.130	142	448	0.021	ng/ul	90
8) 1-Methylnaphthalene	12.339	142	5010	0.230	ng/ul	99
11) Acenaphthene	14.364	153	28479	1.227	ng/ul	98
12) Fluorene	15.354	166	15873	0.629	ng/ul	98
15) Phenanthrene	17.087	178	1947m	0.045	ng/ul	
16) Anthracene	17.180	178	2121m	0.053	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111324\
Data File : BM048701.D
Acq On : 15 Nov 2024 08:59
Operator : RC/JU
Sample : P4761-17
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A0AN2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/15/2024
Supervised By :mohammad ahmed 11/15/2024

Quant Time: Nov 15 10:12:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 11:38:55 2024
Response via : Initial Calibration

