



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111324\
Data File : BM048626.D
Acq On : 13 Nov 2024 08:55
Operator : RC/JU
Sample : DFTPP064
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP064

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM110624.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Nov 13 11:38:55 2024



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.7	145624	PASS
68	69	0.00	2	1.6	2419	PASS
69	198	0.00	100	31.4	153971	PASS
70	69	0.00	2	0.5	811	PASS
127	198	10	80	43.9	215147	PASS
197	198	0.00	2	0.3	1527	PASS
198	198	100	100	100.0	490197	PASS
199	198	5	9	6.6	32408	PASS
275	198	10	60	25.7	126192	PASS
365	198	1	100	3.7	18020	PASS
441	198	0.01	100	13.1	64331	PASS
442	198	50	100	83.2	407808	PASS
443	442	15	24	19.1	77872	PASS

m/z	Abundance
39.00	184.0
40.00	2235.0
41.00	189.0
44.00	3643.0
45.00	258.0
49.00	389.0
51.00	235.0
73.00	525.0
76.90	250.0
83.90	334.0
85.90	211.0
90.90	181.0
96.20	207.0
133.00	298.0
147.00	267.0
191.10	328.0
193.00	261.0
207.10	2751.0
208.10	613.0
209.00	308.0
221.00	174.0
281.10	586.0
281.90	165.0
283.10	181.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

m/z	Abundance
36.00	245.0
37.10	682.0
38.10	2128.0
39.10	12504.0
40.00	2073.0
41.10	410.0
44.00	3360.0
45.00	580.0
46.90	186.0
49.00	1256.0
50.10	40296.0
51.10	169280.0
52.10	9373.0
53.00	444.0
55.00	919.0
56.10	4761.0
57.00	11406.0
58.00	454.0
59.00	176.0
60.00	182.0
61.10	1785.0
62.10	2392.0
63.10	6507.0
64.00	899.0
65.10	3539.0
66.00	317.0
67.10	211.0
68.10	2948.0
69.00	174656.0
70.00	776.0
73.10	1425.0
74.10	15634.0
75.10	27344.0
76.10	8827.0
77.10	206272.0
78.10	14299.0
79.10	11958.0
80.00	9939.0
81.00	13902.0
82.00	3742.0
83.00	3226.0
84.00	649.0
85.00	2332.0
86.10	3791.0
87.10	1834.0
88.10	769.0
89.10	303.0
90.10	171.0
91.10	3190.0
92.10	3707.0
93.00	21592.0
94.10	1561.0
95.00	290.0
96.10	1102.0
97.00	395.0
98.10	16968.0
99.10	14736.0
100.10	1175.0
101.10	8564.0
102.00	437.0
103.10	2161.0
104.00	4735.0
105.10	4916.0
106.10	1766.0
107.10	64360.0
108.00	10085.0
109.10	1460.0
110.10	120360.0
111.10	17840.0
112.10	2308.0
113.10	737.0
114.80	361.0
116.00	3217.0
117.00	49976.0
118.10	3639.0
119.00	558.0
120.00	857.0
121.00	412.0
122.10	3558.0
123.10	5784.0
124.10	3107.0
125.10	3126.0
126.20	866.0
127.10	237568.0
128.10	18216.0
129.10	87632.0
130.10	7507.0
131.10	1534.0
132.20	677.0
133.00	525.0
134.10	2331.0
135.10	6409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

136.00	2798.0
137.10	4096.0
138.00	739.0
139.00	594.0
140.10	1029.0
141.10	11239.0
142.00	3608.0
143.00	2887.0
144.00	672.0
145.00	610.0
146.00	1974.0
147.10	5798.0
148.00	12874.0
149.00	3020.0
150.00	888.0
151.10	1582.0
152.10	721.0
153.00	3956.0
154.00	2798.0
155.10	6698.0
156.10	9362.0
157.10	1974.0
158.00	1698.0
159.00	1652.0
160.00	3665.0
161.10	5071.0
161.90	1461.0
163.10	466.0
164.00	736.0
165.00	3819.0
166.10	3707.0
167.10	23760.0
168.10	15185.0
169.10	1895.0
170.10	876.0
171.10	1026.0
172.00	1897.0
173.10	2674.0
174.10	4600.0
175.10	8743.0
176.10	2201.0
177.10	3883.0
178.00	1610.0
179.00	18352.0
180.10	13254.0
181.10	5868.0
182.10	1047.0
183.10	446.0
184.00	1758.0
185.10	8456.0
186.10	66848.0
187.10	19488.0
188.10	2372.0
189.00	3895.0
190.00	583.0
191.00	2010.0
192.10	5662.0
193.10	6725.0
194.00	1380.0
195.10	641.0
196.10	12569.0
196.90	2748.0
198.00	510400.0
199.00	34800.0
200.10	2626.0
201.50	2143.0
203.00	3801.0
204.10	17616.0
205.10	30232.0
206.10	130368.0
207.10	20408.0
208.10	4842.0
209.10	1212.0
210.00	2040.0
211.10	4319.0
212.00	801.0
213.00	308.0
213.80	175.0
215.00	1540.0
216.10	2673.0
217.00	35160.0
218.00	4025.0
219.00	510.0
220.30	641.0
221.10	18096.0
222.00	4505.0
223.10	7271.0
224.10	71536.0
225.10	18880.0
226.10	1740.0
227.10	29088.0
228.00	4249.0
229.10	6103.0
230.00	984.0
231.10	2420.0
232.10	486.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

233.10	601.0
234.00	1890.0
235.10	2417.0
236.10	1589.0
237.10	2520.0
238.10	455.0
239.00	1275.0
240.10	916.0
241.10	1642.0
242.10	3935.0
243.10	3759.0
244.10	54032.0
245.10	7657.0
246.10	10618.0
247.00	2161.0
248.10	583.0
249.10	1761.0
249.90	408.0
251.10	513.0
252.10	555.0
253.00	1180.0
254.10	1669.0
255.10	283264.0
256.10	42264.0
257.10	3299.0
258.10	16054.0
259.10	2359.0
260.00	517.0
260.90	400.0
264.10	439.0
265.10	6717.0
266.10	1036.0
267.10	159.0
270.10	279.0
271.10	1155.0
272.10	891.0
273.10	8031.0
274.10	22184.0
275.10	121408.0
276.10	16440.0
277.10	9762.0
278.10	1758.0
279.20	411.0
281.10	716.0
282.00	502.0
283.10	1348.0
284.00	731.0
285.10	1687.0
286.10	390.0
289.00	411.0
290.00	391.0
291.00	277.0
291.90	370.0
293.10	2470.0
294.00	811.0
295.10	868.0
296.10	36016.0
297.10	5294.0
298.00	329.0
301.10	455.0
302.10	441.0
303.10	4374.0
304.00	957.0
308.10	505.0
309.00	297.0
310.10	414.0
312.90	262.0
314.10	1563.0
315.00	3749.0
316.10	2402.0
317.10	559.0
321.10	956.0
322.10	605.0
323.10	12042.0
324.10	2236.0
325.10	267.0
326.10	225.0
327.00	2024.0
328.00	1042.0
329.00	181.0
332.20	906.0
333.10	1005.0
334.10	6787.0
335.10	1886.0
336.00	294.0
339.00	217.0
340.00	172.0
341.10	1514.0
342.10	471.0
346.10	2272.0
347.00	682.0
350.00	168.0
351.00	213.0
352.10	3293.0
353.20	2173.0
354.10	3364.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

355.10	782.0
358.90	227.0
359.20	215.0
365.10	16264.0
366.10	2494.0
366.90	182.0
370.10	329.0
371.10	889.0
372.10	6545.0
373.10	1199.0
377.00	195.0
383.10	1474.0
384.10	368.0
385.00	153.0
390.10	747.0
391.00	429.0
392.20	426.0
401.00	288.0
402.10	2274.0
403.10	3615.0
404.10	835.0
405.20	299.0
415.00	222.0
421.10	2777.0
422.20	2839.0
423.10	19520.0
424.10	5193.0
425.10	562.0
440.00	167.0
441.10	53584.0
442.10	337856.0
443.10	66224.0
444.10	6174.0
445.10	595.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

m/z	Abundance
36.00	173.0
37.00	609.0
38.10	2009.0
39.10	12321.0
40.00	2309.0
41.00	548.0
43.00	160.0
44.10	3396.0
45.00	455.0
49.10	1437.0
50.10	41360.0
51.10	169344.0
52.10	8366.0
53.10	271.0
55.00	596.0
56.00	4910.0
57.10	11727.0
58.00	452.0
61.00	2092.0
62.10	2481.0
63.10	6352.0
64.00	996.0
65.10	3593.0
66.10	220.0
67.00	180.0
68.00	2789.0
69.00	178880.0
70.00	1033.0
73.10	1383.0
74.10	16512.0
75.10	27984.0
76.10	10230.0
77.10	209088.0
78.10	14461.0
79.10	12070.0
80.00	9793.0
81.10	14172.0
82.10	4012.0
83.00	3035.0
84.00	674.0
85.10	2244.0
86.10	3824.0
87.10	1991.0
88.10	853.0
89.00	399.0
91.10	3182.0
92.10	3836.0
93.10	23504.0
94.00	1828.0
95.10	335.0
96.10	1237.0
97.00	419.0
98.10	16346.0
99.10	15214.0
100.00	1421.0
101.00	8199.0
102.00	434.0
103.10	2702.0
104.00	4950.0
105.10	5133.0
106.10	1590.0
107.10	66872.0
108.10	9915.0
109.10	1880.0
110.10	123352.0
111.10	19008.0
112.00	2084.0
113.10	766.0
115.10	440.0
116.00	3447.0
117.10	51584.0
118.00	3702.0
119.20	594.0
120.10	916.0
121.20	240.0
122.10	4405.0
123.00	6444.0
124.00	2809.0
125.00	3152.0
126.20	892.0
127.10	251456.0
128.10	18272.0
129.10	92744.0
130.00	8012.0
131.10	1382.0
132.10	895.0
133.00	615.0
134.00	2848.0
135.10	7906.0
136.10	2844.0
137.10	4016.0
138.00	985.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

139.00	411.0
140.00	1127.0
141.10	11457.0
142.10	3677.0
143.00	2507.0
144.10	545.0
145.10	641.0
146.10	2305.0
147.10	6249.0
148.10	14338.0
149.10	2978.0
150.10	1070.0
151.20	1479.0
151.90	964.0
153.10	3906.0
154.10	2972.0
155.10	6983.0
156.10	10431.0
157.10	2120.0
158.00	2660.0
159.00	2028.0
160.00	4080.0
161.10	5714.0
162.00	1750.0
163.10	474.0
164.00	706.0
165.10	4273.0
166.10	3560.0
167.10	24872.0
168.10	14280.0
169.10	2219.0
169.90	709.0
171.00	1105.0
172.00	2286.0
173.10	2833.0
174.10	5603.0
175.10	9726.0
176.10	2818.0
177.00	4884.0
178.00	1475.0
179.10	19192.0
180.10	14841.0
181.10	6237.0
182.10	962.0
183.10	443.0
184.10	1520.0
185.00	9109.0
186.10	71800.0
187.10	21592.0
188.10	2401.0
189.10	4235.0
190.00	791.0
191.10	2604.0
192.10	6458.0
193.10	7236.0
194.10	1548.0
195.00	833.0
196.10	14581.0
197.10	2595.0
198.00	574720.0
199.00	37696.0
200.00	2838.0
201.50	2349.0
203.10	4525.0
204.10	20568.0
205.10	32408.0
206.10	145344.0
207.10	22448.0
208.10	5218.0
209.10	1769.0
210.10	2067.0
211.10	5707.0
211.90	1035.0
213.00	509.0
214.10	150.0
215.00	1544.0
216.10	3010.0
217.10	39344.0
218.10	5297.0
219.10	473.0
221.10	21176.0
221.90	5134.0
223.10	9038.0
224.10	82416.0
225.10	21000.0
226.10	2470.0
227.10	34400.0
228.10	4230.0
229.00	7501.0
230.00	821.0
231.10	2899.0
232.10	568.0
233.10	554.0
234.00	2013.0
235.10	2252.0
236.00	1719.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

237.10	2734.0
238.00	457.0
239.10	1269.0
240.10	1106.0
241.00	1950.0
242.10	4500.0
243.10	4569.0
244.10	65160.0
245.10	9004.0
246.10	12123.0
247.10	2392.0
248.00	518.0
249.00	2626.0
250.00	527.0
250.90	479.0
252.00	562.0
253.10	1371.0
254.10	1936.0
255.10	339264.0
256.10	49056.0
257.10	4139.0
258.10	19184.0
259.10	2968.0
260.00	394.0
261.00	791.0
263.00	153.0
264.00	529.0
265.00	7755.0
266.00	1285.0
267.10	185.0
267.90	240.0
270.10	308.0
271.10	1183.0
272.10	1047.0
273.10	9719.0
274.10	25944.0
275.10	150464.0
276.10	19072.0
277.10	13188.0
278.00	2331.0
279.00	469.0
281.00	554.0
282.00	372.0
283.10	1697.0
284.10	789.0
285.20	2038.0
286.10	420.0
289.00	432.0
290.00	232.0
291.00	365.0
292.00	668.0
293.10	3039.0
294.10	710.0
295.10	725.0
296.10	44720.0
297.10	6307.0
297.90	417.0
301.10	436.0
302.10	741.0
303.10	4913.0
304.10	1290.0
305.00	311.0
308.10	683.0
309.00	379.0
310.10	638.0
312.00	176.0
313.10	369.0
314.10	1994.0
315.00	5346.0
316.10	2475.0
317.00	485.0
319.90	157.0
321.10	1602.0
322.10	640.0
323.10	15112.0
324.10	2662.0
325.20	291.0
326.00	355.0
327.00	2626.0
328.10	1414.0
329.10	217.0
332.10	1218.0
333.10	1545.0
334.10	9389.0
335.10	2463.0
336.00	269.0
340.00	201.0
341.10	1683.0
342.00	494.0
342.90	163.0
346.10	3398.0
347.10	495.0
350.00	255.0
351.10	286.0
352.10	4256.0
353.10	3197.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

354.10	4858.0
355.10	873.0
356.00	161.0
359.10	407.0
364.20	183.0
365.10	21600.0
366.10	3021.0
366.90	176.0
370.20	317.0
371.00	1100.0
372.10	7541.0
373.10	1928.0
374.10	243.0
376.90	226.0
383.10	2123.0
384.10	829.0
385.10	212.0
390.10	1020.0
391.10	724.0
392.00	405.0
401.00	495.0
402.10	3596.0
403.10	4527.0
404.10	1897.0
405.20	303.0
410.00	184.0
415.20	170.0
421.10	3942.0
422.10	3788.0
423.10	26848.0
424.10	6223.0
425.10	749.0
437.90	201.0
439.10	252.0
440.10	156.0
441.10	75704.0
442.10	481536.0
443.10	93408.0
444.10	8400.0
445.20	472.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

m/z	Abundance
37.00	365.0
38.10	1088.0
39.10	6974.0
40.10	2158.0
41.10	384.0
42.00	160.0
43.10	204.0
44.10	4104.0
45.00	282.0
49.00	924.0
50.10	24352.0
51.10	98952.0
52.10	5031.0
53.00	234.0
55.10	477.0
56.00	2510.0
57.10	6922.0
58.00	273.0
59.00	247.0
61.00	1177.0
62.00	1470.0
63.10	3567.0
64.10	522.0
65.10	1963.0
66.90	154.0
68.10	1520.0
69.00	108376.0
70.10	623.0
73.00	1132.0
74.10	10104.0
75.10	16960.0
76.10	5714.0
77.10	127080.0
78.10	7963.0
79.10	7760.0
80.10	5980.0
81.00	7889.0
82.00	2310.0
83.00	2139.0
84.00	408.0
85.00	1602.0
86.00	2528.0
87.10	1313.0
87.90	281.0
89.00	271.0
91.00	1996.0
92.00	2084.0
93.00	14663.0
94.00	1112.0
94.90	334.0
96.10	979.0
97.00	228.0
98.10	9847.0
99.00	9365.0
100.00	665.0
101.00	4947.0
101.90	175.0
103.00	1548.0
104.00	2941.0
105.00	2988.0
106.10	945.0
107.10	40560.0
108.10	6260.0
109.10	992.0
110.00	76752.0
111.00	10767.0
112.10	1228.0
113.00	667.0
115.00	289.0
116.00	2589.0
117.10	31128.0
118.00	2310.0
119.10	467.0
120.00	583.0
121.20	232.0
122.10	2535.0
123.10	4312.0
124.00	2025.0
125.00	1778.0
126.10	666.0
127.10	156416.0
128.10	11832.0
129.10	57960.0
130.10	4623.0
131.00	1052.0
132.10	698.0
133.00	374.0
134.00	1979.0
135.10	4194.0
136.10	2004.0
137.10	2672.0
138.00	549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

139.00	296.0
140.00	746.0
141.00	7486.0
142.10	2271.0
143.10	1928.0
144.10	672.0
145.00	391.0
146.00	1163.0
147.00	4365.0
148.00	9153.0
149.00	2027.0
150.10	687.0
151.20	921.0
152.00	542.0
153.00	2297.0
154.10	1716.0
155.10	4724.0
156.10	6418.0
157.00	1190.0
158.10	1443.0
159.00	1088.0
160.00	2154.0
161.00	3986.0
162.10	1084.0
163.10	245.0
164.00	449.0
165.10	2689.0
166.00	2378.0
167.10	15874.0
168.10	9754.0
169.00	1387.0
170.00	585.0
171.10	764.0
172.00	1465.0
173.10	1998.0
174.00	3314.0
175.10	6480.0
176.10	1794.0
177.00	3298.0
178.00	925.0
179.00	12820.0
180.10	9014.0
181.10	4732.0
182.00	533.0
183.10	493.0
184.10	888.0
185.10	5992.0
186.10	48600.0
187.10	13854.0
188.10	1391.0
189.00	2935.0
190.10	429.0
191.10	1569.0
192.00	4447.0
193.10	5162.0
194.00	1093.0
195.10	610.0
196.10	9592.0
196.80	1834.0
198.00	385472.0
199.00	24728.0
200.00	1743.0
201.50	1338.0
203.00	2741.0
204.10	13359.0
205.10	21120.0
206.10	97888.0
207.10	15560.0
208.10	3797.0
209.10	1360.0
210.20	1655.0
211.10	3961.0
213.00	282.0
214.00	169.0
215.10	1050.0
216.00	1873.0
217.00	26560.0
218.10	3835.0
219.00	428.0
221.10	14124.0
221.90	3911.0
223.10	6536.0
224.10	55344.0
225.10	14721.0
226.10	1570.0
227.10	22304.0
228.00	3395.0
229.10	4806.0
230.00	574.0
231.10	1979.0
232.10	320.0
233.10	450.0
234.00	1566.0
235.00	1558.0
236.00	1356.0
237.00	1907.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

238.00	242.0
239.00	911.0
240.00	733.0
241.10	1498.0
242.10	3095.0
243.10	2930.0
244.10	45728.0
245.10	5639.0
246.10	8261.0
247.10	1553.0
248.10	429.0
249.00	1757.0
250.00	394.0
251.00	340.0
252.10	416.0
253.20	967.0
254.10	1406.0
255.10	238208.0
256.10	33920.0
257.10	2761.0
258.10	13724.0
259.10	2553.0
260.10	336.0
261.00	361.0
264.10	240.0
265.10	5803.0
266.00	687.0
267.10	184.0
270.10	328.0
271.00	668.0
272.10	587.0
273.10	6773.0
274.10	18800.0
275.10	106704.0
276.10	13653.0
277.10	8704.0
278.10	1654.0
279.00	385.0
281.20	578.0
282.00	510.0
283.10	1034.0
284.10	842.0
285.10	1543.0
286.10	236.0
289.10	311.0
289.90	189.0
290.10	189.0
291.00	268.0
292.10	404.0
293.10	1886.0
294.10	567.0
295.10	568.0
296.10	32240.0
297.10	4427.0
298.00	263.0
301.00	325.0
302.00	447.0
303.10	3399.0
304.10	1014.0
308.10	348.0
309.10	422.0
310.10	491.0
312.00	164.0
313.00	288.0
314.10	1450.0
315.00	3466.0
316.10	1996.0
317.10	392.0
321.10	1020.0
322.00	595.0
323.10	11526.0
324.10	1870.0
325.10	253.0
325.90	157.0
327.10	2052.0
328.10	1011.0
329.10	213.0
332.10	973.0
333.10	1335.0
334.10	6779.0
335.10	1700.0
336.00	210.0
339.00	188.0
341.20	1396.0
342.20	280.0
346.10	2594.0
347.20	348.0
351.00	161.0
352.10	3047.0
353.10	2439.0
354.10	3247.0
355.10	753.0
359.00	297.0
365.10	16196.0
366.00	2356.0
367.00	152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064

370.00	275.0
371.00	946.0
372.20	5639.0
373.10	1576.0
374.00	238.0
376.90	155.0
383.10	1834.0
384.00	393.0
384.90	175.0
390.10	749.0
391.10	462.0
392.10	436.0
401.00	472.0
402.10	2981.0
403.10	3760.0
404.10	1303.0
405.10	274.0
421.10	3622.0
422.10	3203.0
423.10	22280.0
424.10	5074.0
425.10	534.0
441.10	63704.0
442.10	404032.0
443.10	73984.0
444.10	7104.0
445.00	351.0

Instrument :
BNA_M
ClientSampleId :
DFTPP064