

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033127.D
 Acq On : 17 Nov 2021 19:47
 Operator : CG/JU
 Sample : PB140791BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK791

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
 Supervised By :mohammad ahmed 11/26/2021

Quant Time: Nov 18 00:35:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.972	152	72009	20.000	ng/u1	0.00
20) Naphthalene-d8	10.772	136	278463	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.589	164	180465	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.324	188	375698	20.000	ng/u1	0.00
79) Chrysene-d12	21.477	240	371689	20.000	ng/u1	0.00
88) Perylene-d12	23.830	264	371915	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.425	96	11757m	6.294	ng/uL	0.00
4) Pyridine-d5	3.843	84	160292	31.154	ng/u1	0.00
7) Phenol-d5	7.125	99	192880	31.666	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.301	67	124143	32.133	ng/u1	0.00
11) 2-Chlorophenol-d4	7.501	132	149277	32.357	ng/u1	0.00
15) 4-Methylphenol-d8	8.666	113	146605	31.014	ng/u1	0.00
21) Nitrobenzene-d5	9.131	128	67597	33.811	ng/u1	0.00
24) 2-Nitrophenol-d4	9.854	143	67542	33.733	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.384	165	140878	30.724	ng/u1	0.00
31) 4-Chloroaniline-d4	10.907	131	197774	32.434	ng/u1	0.00
46) Dimethylphthalate-d6	14.001	166	457806	34.586	ng/u1	0.00
49) Acenaphthylene-d8	14.289	160	543404	31.866	ng/u1	0.00
54) 4-Nitrophenol-d4	14.772	143	63145	29.286	ng/u1	0.00
60) Fluorene-d10	15.583	176	388647	32.730	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	48376	27.399	ng/u1	0.00
73) Anthracene-d10	17.424	188	639323	35.302	ng/u1	0.00
81) Pyrene-d10	19.706	212	758739	34.543	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.677	264	713740	35.758	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
Data File : BM033127.D
Acq On : 17 Nov 2021 19:47
Operator : CG/JU
Sample : PB140791BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK791

Quant Time: Nov 18 00:35:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 17 14:14:11 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021

