

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033144.D
 Acq On : 18 Nov 2021 06:34
 Operator : CG/JU
 Sample : PB140814BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK814

Quant Time: Nov 18 07:10:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.972	152	98022	20.000	ng/u1	0.00
20) Naphthalene-d8	10.766	136	391901	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.589	164	257111	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.324	188	524810	20.000	ng/u1	0.00
79) Chrysene-d12	21.477	240	497144	20.000	ng/u1	0.00
88) Perylene-d12	23.824	264	488132	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.419	96	16444	6.467	ng/uL	0.00
4) Pyridine-d5	3.837	84	218605	31.212	ng/u1	-0.01
7) Phenol-d5	7.119	99	263684	31.802	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.296	67	172091	32.723	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.502	132	202404	32.230	ng/u1	0.00
15) 4-Methylphenol-d8	8.660	113	208715	32.436	ng/u1	-0.01
21) Nitrobenzene-d5	9.131	128	97880	34.787	ng/u1	0.00
24) 2-Nitrophenol-d4	9.848	143	101687	36.085	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.378	165	202307	31.350	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.901	131	282055	32.866	ng/u1	0.00
46) Dimethylphthalate-d6	13.995	166	649361	34.433	ng/u1	-0.01
49) Acenaphthylene-d8	14.283	160	787584	32.417	ng/u1	0.00
54) 4-Nitrophenol-d4	14.766	143	90400	29.428	ng/u1	0.00
60) Fluorene-d10	15.577	176	560119	33.109	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.683	200	71693	29.068	ng/u1	-0.01
73) Anthracene-d10	17.419	188	855945	33.835	ng/u1	-0.01
81) Pyrene-d10	19.701	212	969638	33.005	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.671	264	879019	33.554	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
Data File : BM033144.D
Acq On : 18 Nov 2021 06:34
Operator : CG/JU
Sample : PB140814BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK814

Quant Time: Nov 18 07:10:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 17 14:14:11 2021
Response via : Initial Calibration

