

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033130.D
 Acq On : 17 Nov 2021 21:36
 Operator : CG/JU
 Sample : PB140812BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS812

Quant Time: Nov 18 00:36:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
 Supervised By :mohammad ahmed 11/26/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.972	152	93834	20.000	ng/u1	0.00
20) Naphthalene-d8	10.772	136	389989	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.589	164	269208	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.330	188	580622	20.000	ng/u1	0.00
79) Chrysene-d12	21.483	240	568750	20.000	ng/u1	0.00
88) Perylene-d12	23.830	264	561689	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.425	96	11309	4.646	ng/uL	0.00
4) Pyridine-d5	3.843	84	161513	24.090	ng/u1	0.00
7) Phenol-d5	7.125	99	201346	25.368	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.301	67	127316	25.290	ng/u1	0.00
11) 2-Chlorophenol-d4	7.507	132	153798	25.583	ng/u1	0.00
15) 4-Methylphenol-d8	8.672	113	160571	26.068	ng/u1	0.00
21) Nitrobenzene-d5	9.136	128	73077	26.099	ng/u1	0.00
24) 2-Nitrophenol-d4	9.854	143	76112	27.142	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.389	165	166977	26.002	ng/u1	0.00
31) 4-Chloroaniline-d4	10.907	131	196929	23.060	ng/u1	0.00
46) Dimethylphthalate-d6	14.001	166	513845	26.023	ng/u1	0.00
49) Acenaphthylene-d8	14.289	160	641875	25.232	ng/u1	0.00
54) 4-Nitrophenol-d4	14.771	143	83355	25.915	ng/u1	0.00
60) Fluorene-d10	15.577	176	465358	26.271	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	71669	26.265	ng/u1	0.00
73) Anthracene-d10	17.424	188	726099	25.943	ng/u1	0.00
81) Pyrene-d10	19.706	212	859193	25.564	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.677	264	796202	26.412	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.460	88	24324m	9.843	ng/uL	
5) Pyridine	3.860	79	165423	24.179	ng/u1	99
6) Benzaldehyde	7.119	77	126220	27.903	ng/u1	99
8) Phenol	7.154	94	204716	25.900	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.395	93	161959	25.795	ng/u1	98
12) 2-Chlorophenol	7.537	128	161509	26.029	ng/u1	99
13) 2-Methylphenol	8.407	108	156368	25.825	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.501	45	263163	27.138	ng/u1	98
16) Acetophenone	8.795	105	257568	26.492	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.784	70	141736	27.057	ng/u1	98
18) 4-Methylphenol	8.731	108	167056	26.319	ng/u1	99
19) Hexachloroethane	9.054	117	74159	26.235	ng/u1	96
22) Nitrobenzene	9.178	77	206392	26.413	ng/u1	98
23) Isophorone	9.707	82	378095	26.473	ng/u1	98
25) 2-Nitrophenol	9.883	139	83843	28.208	ng/u1	96
26) 2,4-Dimethylphenol	9.936	107	207013	26.507	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.178	93	219046	25.772	ng/u1	97
29) 2,4-Dichlorophenol	10.413	162	164278	26.382	ng/u1	96
30) Naphthalene	10.825	128	531584	25.633	ng/u1	99
32) 4-Chloroaniline	10.930	127	197301	22.870	ng/u1	97
33) Hexachlorobutadiene	11.101	225	124061	25.577	ng/u1	99
34) Caprolactam	11.719	113	45603m	25.473	ng/u1	
35) 4-Chloro-3-methylphenol	12.036	107	183887	27.119	ng/u1	98
36) 2-Methylnaphthalene	12.425	142	381308	26.537	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033130.D
 Acq On : 17 Nov 2021 21:36
 Operator : CG/JU
 Sample : PB140812BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS812

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
 Supervised By :mohammad ahmed 11/26/2021

Quant Time: Nov 18 00:36:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.642	142	384204	26.193	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.783	216	222999	24.898	ng/ul	99
40) Hexachlorocyclopentadiene	12.766	237	145439	23.077	ng/ul	96
41) 2,4,6-Trichlorophenol	13.024	196	139521	26.450	ng/ul	100
42) 2,4,5-Trichlorophenol	13.095	196	149101	26.361	ng/ul	97
43) 1,1'-Biphenyl	13.430	154	527456	25.147	ng/ul	98
44) 2-Chloronaphthalene	13.472	162	405391	25.062	ng/ul	99
45) 2-Nitroaniline	13.671	65	122426	28.715	ng/ul	95
47) Dimethylphthalate	14.048	163	506845	26.336	ng/ul	99
48) 2,6-Dinitrotoluene	14.166	165	93863	28.505	ng/ul	98
50) Acenaphthylene	14.319	152	660429	25.584	ng/ul	100
51) 3-Nitroaniline	14.495	138	83900	25.288	ng/ul	92
52) Acenaphthene	14.654	153	435500	25.861	ng/ul	99
53) 2,4-Dinitrophenol	14.695	184	43368	25.804	ng/ul	93
55) 4-Nitrophenol	14.789	109	87539	26.647	ng/ul	96
56) Dibenzofuran	14.989	168	645528	26.155	ng/ul	100
57) 2,4-Dinitrotoluene	14.948	165	138062	30.487	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.207	232	126845	27.474	ng/ul	95
59) Diethylphthalate	15.407	149	534591	27.733	ng/ul	99
61) Fluorene	15.636	166	520141	26.514	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.630	204	272844	26.647	ng/ul	98
63) 4-Nitroaniline	15.648	138	94100	28.844	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.707	198	72152	26.382	ng/ul	100
67) N-Nitrosodiphenylamine	15.842	169	453589	26.561	ng/ul	99
68) 4-Bromophenyl-phenylether	16.518	248	172347	26.692	ng/ul	96
69) Hexachlorobenzene	16.630	284	190569	25.763	ng/ul	97
70) Atrazine	16.789	200	170107	25.779	ng/ul	99
71) Pentachlorophenol	16.971	266	109650	25.645	ng/ul	99
72) Phenanthrene	17.371	178	850473	26.443	ng/ul	98
74) Anthracene	17.459	178	858072	26.604	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.395	216	228040	23.911	ng/uL	97
76) Pentachlorobenzene	14.907	250	226867	24.265	ng/uL	99
77) Carbazole	17.724	167	765352	26.711	ng/ul	99
78) Di-n-butylphthalate	18.289	149	901921	28.787	ng/ul	100
80) Fluoranthene	19.371	202	1018722	25.872	ng/ul	99
82) Pyrene	19.736	202	1040501	26.026	ng/ul	99
83) Butylbenzylphthalate	20.624	149	383042	28.043	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.395	252	307656	24.011	ng/ul	99
85) Benzo(a)anthracene	21.465	228	984603	26.766	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.389	149	550331	28.425	ng/ul	99
87) Chrysene	21.518	228	957115	26.643	ng/ul	99
89) Di-n-octyl phthalate	22.300	149	929500	27.159	ng/ul	100
90) Benzo(b)fluoranthene	23.118	252	1026976	27.148	ng/ul	98
91) Benzo(k)fluoranthene	23.165	252	912577	26.334	ng/ul	99
93) Benzo(a)pyrene	23.730	252	953024	26.657	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.241	276	1050728	26.392	ng/ul	97
95) Dibenzo(a,h)anthracene	26.253	278	902094	26.485	ng/ul	98
96) Benzo(g,h,i)perylene	26.977	276	914759	26.473	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033130.D
 Acq On : 17 Nov 2021 21:36
 Operator : CG/JU
 Sample : PB140812BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :


SLCS812


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
 Supervised By :mohammad ahmed 11/26/2021


Quant Time: Nov 18 00:36:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 9.843 ng/uL m
RT: 3.460 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion: 88 Resp: 24324
Ion Ratio Lower Upper
88 100
43 41.0 34.9 52.3
58 88.9 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#5
Pyridine
Concen: 24.179 ng/uL
RT: 3.860 min Scan# 89
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36



Tgt Ion: 79 Resp: 165423
Ion Ratio Lower Upper
79 100
52 87.1 69.1 103.7
51 43.5 35.0 52.6









#6
Benzaldehyde
Concen: 27.903 ng/ul
RT: 7.119 min Scan# 643
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Instrument :
BNA_M
ClientSampleId :
SLCS812

Tgt Ion: 77 Resp: 126220
Ion Ratio Lower Upper
77 100
105 89.7 70.7 106.1
106 83.2 66.8 100.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#8
Phenol
Concen: 25.900 ng/ul
RT: 7.154 min Scan# 649
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion: 94 Resp: 204716
Ion Ratio Lower Upper
94 100
65 35.1 27.1 40.7
66 49.1 39.4 59.2





#10
Bis(2-Chloroethyl)ether
Concen: 25.795 ng/ul
RT: 7.395 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion: 93 Resp: 161959
Ion Ratio Lower Upper
93 100
63 82.4 67.4 101.2
95 32.1 24.8 37.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#12
2-Chlorophenol
Concen: 26.029 ng/ul
RT: 7.537 min Scan# 714
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36



Tgt Ion:128 Resp: 161509
Ion Ratio Lower Upper
128 100
64 52.6 42.6 63.8
130 31.0 24.4 36.6









#13
2-Methylphenol
Concen: 25.825 ng/ul
RT: 8.407 min Scan# 862
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Instrument :
BNA_M
ClientSampleId :
SLCS812

Tgt Ion:108 Resp: 156368
Ion Ratio Lower Upper
108 100
107 87.9 69.6 104.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.138 ng/ul
RT: 8.501 min Scan# 878
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion: 45 Resp: 263163
Ion Ratio Lower Upper
45 100
77 12.8 10.9 16.3
79 10.2 7.6 11.4





#16
Acetophenone
Concen: 26.492 ng/ul
RT: 8.795 min Scan# 912
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion: 105 Resp: 257568
Ion Ratio Lower Upper
105 100
77 83.3 68.0 102.0
51 34.3 27.8 41.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#17
N-Nitroso-di-n-propylamine
Concen: 27.057 ng/ul
RT: 8.784 min Scan# 926
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion: 70 Resp: 141736
Ion Ratio Lower Upper
70 100
42 68.4 52.8 79.2
101 9.1 7.1 10.7
130 19.3 15.0 22.4





#18
4-Methylphenol
Concen: 26.319 ng/ul
RT: 8.731 min Scan# 917
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:108 Resp: 167050
Ion Ratio Lower Upper
108 100
107 113.0 89.4 134.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#19
Hexachloroethane
Concen: 26.235 ng/ul
RT: 9.054 min Scan# 972
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:117 Resp: 74159
Ion Ratio Lower Upper
117 100
201 104.5 87.0 130.4
199 64.5 53.8 80.6





#22
Nitrobenzene
Concen: 26.413 ng/ul
RT: 9.178 min Scan# 993
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion: 77 Resp: 206392
Ion Ratio Lower Upper
77 100
123 37.8 31.2 46.8
65 14.5 11.8 17.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#23
Isophorone
Concen: 26.473 ng/ul
RT: 9.707 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Tgt Ion: 82 Resp: 378095
Ion Ratio Lower Upper
82 100
95 8.6 6.6 9.8
138 16.8 12.6 19.0









#25
2-Nitrophenol
Concen: 28.208 ng/ul
RT: 9.883 min Scan# 1113
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Instrument :
BNA_M
ClientSampleId :
SLCS812

Tgt Ion:139 Resp: 8384
Ion Ratio Lower Upper
139 100
65 64.8 55.9 83.9
109 41.8 33.8 50.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#26
2,4-Dimethylphenol
Concen: 26.507 ng/ul
RT: 9.936 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Tgt Ion:107 Resp: 207013
Ion Ratio Lower Upper
107 100
121 43.8 36.2 54.2
122 76.2 63.6 95.4





#27

Bis(2-Chloroethoxy)methane

Concen: 25.772 ng/ul

RT: 10.178 min Scan# 1163

Delta R.T. -0.006 min

Lab File: BM033130.D

Acq: 17 Nov 2021 21:36

Instrument :

BNA_M

ClientSampleId :

SLCS812

Tgt Ion: 93 Resp: 219040

Ion Ratio Lower Upper

93 100

95 32.0 24.3 36.5





123 10.3 9.0 13.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021

Supervised By :mohammad ahmed 11/26/2021



#29

2,4-Dichlorophenol

Concen: 26.382 ng/ul

RT: 10.413 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BM033130.D

Acq: 17 Nov 2021 21:36




Tgt Ion:162 Resp: 164278


Ion Ratio Lower Upper

162 100

164 66.0 49.8 74.8


98 37.6 31.2 46.8





#30
Naphthalene
Concen: 25.633 ng/ul
RT: 10.825 min Scan# 1273
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:128 Resp: 531584
Ion Ratio Lower Upper
128 100
129 10.4 9.1 13.7
127 12.9 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#32
4-Chloroaniline
Concen: 22.870 ng/ul
RT: 10.930 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:127 Resp: 197301
Ion Ratio Lower Upper
127 100
129 32.5 24.9 37.3





#33
Hexachlorobutadiene
Concen: 25.577 ng/ul
RT: 11.101 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:225 Resp: 12406
Ion Ratio Lower Upper
225 100
223 63.2 51.1 76.7
227 65.2 51.7 77.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#34
Caprolactam
Concen: 25.473 ng/ul m
RT: 11.719 min Scan# 1425
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:113 Resp: 45603
Ion Ratio Lower Upper
113 100
55 203.9 157.3 235.9
56 146.4 118.2 177.4





#35
4-Chloro-3-methylphenol
Concen: 27.119 ng/ul
RT: 12.036 min Scan# 1479
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:107 Resp: 18388
Ion Ratio Lower Upper
107 100
144 24.8 19.3 28.9
142 75.2 61.4 92.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#36
2-Methylnaphthalene
Concen: 26.537 ng/ul
RT: 12.425 min Scan# 1545
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36



Tgt Ion:142 Resp: 381308
Ion Ratio Lower Upper
142 100
141 89.8 71.0 106.6





#37

1-Methylnaphthalene

Concen: 26.193 ng/ul

RT: 12.642 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM033130.D


Acq: 17 Nov 2021 21:36

Instrument :

BNA_M

ClientSampleId :

SLCS812






Tgt Ion:	142	Resp:	384204
Ion Ratio	Lower	Upper	
142	100		
141	91.8	73.2	109.8
116	4.5	3.6	5.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.898 ng/ul




RT: 12.783 min Scan# 1606


Delta R.T. -0.006 min

Lab File: BM033130.D

Acq: 17 Nov 2021 21:36


Tgt Ion:	216	Resp:	222999
Ion Ratio	Lower	Upper	
216	100		
214	79.2	62.4	93.6
179	20.8	17.0	25.6
108	18.6	14.6	21.8





#40
Hexachlorocyclopentadiene
Concen: 23.077 ng/ul
RT: 12.766 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:237 Resp: 145439
Ion Ratio Lower Upper
237 100
235 60.4 51.2 76.8
272 12.7 10.4 15.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#41
2,4,6-Trichlorophenol
Concen: 26.450 ng/ul
RT: 13.024 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:196 Resp: 139521
Ion Ratio Lower Upper
196 100
198 93.0 74.3 111.5
200 28.7 23.3 34.9





#42
2,4,5-Trichlorophenol
Concen: 26.361 ng/ul
RT: 13.095 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:196 Resp: 14910
Ion Ratio Lower Upper
196 100
198 97.1 75.3 112.9
200 32.1 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#43
1,1'-Biphenyl
Concen: 25.147 ng/ul
RT: 13.430 min Scan# 1716
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:154 Resp: 527456
Ion Ratio Lower Upper
154 100
153 39.8 33.0 49.4
76 12.0 10.1 15.1





#44
2-Chloronaphthalene
Concen: 25.062 ng/ul
RT: 13.472 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:162 Resp: 40539
Ion Ratio Lower Upper
162 100
164 32.6 26.0 39.0
127 40.2 31.0 46.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#45
2-Nitroaniline
Concen: 28.715 ng/ul
RT: 13.671 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion: 65 Resp: 122426
Ion Ratio Lower Upper
65 100
92 63.4 53.2 79.8
138 88.3 74.8 112.2





#47
Dimethylphthalate
Concen: 26.336 ng/ul
RT: 14.048 min Scan# 1821
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:163 Resp: 506849
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.2 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#48
2,6-Dinitrotoluene
Concen: 28.505 ng/ul
RT: 14.166 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:165 Resp: 93863
Ion Ratio Lower Upper
165 100
89 72.3 57.1 85.7
121 25.5 21.8 32.8





#50
Acenaphthylene
Concen: 25.584 ng/ul
RT: 14.319 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:152 Resp: 660429
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 12.6 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#51
3-Nitroaniline
Concen: 25.288 ng/ul
RT: 14.495 min Scan# 1897
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:138 Resp: 83900
Ion Ratio Lower Upper
138 100
108 13.1 12.6 18.8
92 127.0 109.4 164.2





#52
Acenaphthene
Concen: 25.861 ng/ul
RT: 14.654 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:153 Resp: 435500
Ion Ratio Lower Upper
153 100
152 48.0 38.9 58.3
154 88.4 71.7 107.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#53
2,4-Dinitrophenol
Concen: 25.804 ng/ul
RT: 14.695 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:184 Resp: 43368
Ion Ratio Lower Upper
184 100
63 78.1 55.2 82.8
154 69.9 54.2 81.4





#55
4-Nitrophenol
Concen: 26.647 ng/ul
RT: 14.789 min Scan# 1946
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:109 Resp: 87539
Ion Ratio Lower Upper
109 100
139 100.0 85.4 128.0
65 123.7 100.9 151.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#56
Dibenzofuran
Concen: 26.155 ng/ul
RT: 14.989 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:168 Resp: 645528
Ion Ratio Lower Upper
168 100
139 39.4 31.7 47.5





#57
2,4-Dinitrotoluene
Concen: 30.487 ng/ul
RT: 14.948 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:165 Resp: 138062
Ion Ratio Lower Upper
165 100
63 52.3 41.5 62.3
182 3.1 2.0 3.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#58
2,3,4,6-Tetrachlorophenol
Concen: 27.474 ng/ul
RT: 15.207 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:232 Resp: 126845
Ion Ratio Lower Upper
232 100
131 48.2 34.8 52.2
130 2.6 2.1 3.1
166 32.4 24.5 36.7





#59
Diethylphthalate
Concen: 27.733 ng/ul
RT: 15.407 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:149 Resp: 53459
Ion Ratio Lower Upper
149 100
177 22.2 18.0 27.0
150 12.2 9.0 13.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#61
Fluorene
Concen: 26.514 ng/ul
RT: 15.636 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:166 Resp: 520141
Ion Ratio Lower Upper
166 100
165 97.4 76.8 115.2
167 12.8 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 26.647 ng/ul
RT: 15.630 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:204 Resp: 272844
Ion Ratio Lower Upper
204 100
206 31.8 25.7 38.5
141 62.3 48.6 72.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#63
4-Nitroaniline
Concen: 28.844 ng/ul
RT: 15.648 min Scan# 2093
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:138 Resp: 94100
Ion Ratio Lower Upper
138 100
92 59.0 47.3 70.9
108 102.0 83.0 124.6





#66
4,6-Dinitro-2-methylphenol
Concen: 26.382 ng/ul
RT: 15.707 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:198 Resp: 72152
Ion Ratio Lower Upper
198 100
182 4.8 3.8 5.8
77 27.5 22.0 33.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#67
N-Nitrosodiphenylamine
Concen: 26.561 ng/ul
RT: 15.842 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:169 Resp: 453589
Ion Ratio Lower Upper
169 100
168 66.4 53.7 80.5
167 35.9 28.8 43.2





#68
4-Bromophenyl-phenylether
Concen: 26.692 ng/ul
RT: 16.518 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:248 Resp: 17234
Ion Ratio Lower Upper
248 100
250 97.2 75.5 113.3
141 78.4 59.0 88.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#69
Hexachlorobenzene
Concen: 25.763 ng/ul
RT: 16.630 min Scan# 2260
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:284 Resp: 190569
Ion Ratio Lower Upper
284 100
142 33.7 28.4 42.6
249 32.2 24.8 37.2





#70
Atrazine
Concen: 25.779 ng/ul
RT: 16.789 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:200 Resp: 17010
Ion Ratio Lower Upper
200 100
173 25.1 20.8 31.2
215 47.5 38.6 57.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#71
Pentachlorophenol
Concen: 25.645 ng/ul
RT: 16.971 min Scan# 2318
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:266 Resp: 109650
Ion Ratio Lower Upper
266 100
264 64.7 50.6 75.8
268 63.2 51.1 76.7





#72
Phenanthrene
Concen: 26.443 ng/ul
RT: 17.371 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:178 Resp: 85047
Ion Ratio Lower Upper
178 100
179 15.4 11.9 17.9
176 19.4 14.6 21.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#74
Anthracene
Concen: 26.604 ng/ul
RT: 17.459 min Scan# 2401
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:178 Resp: 858072
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 18.2 15.3 22.9





#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.911 ng/uL
RT: 13.395 min Scan# 1710
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:216 Resp: 228040
Ion Ratio Lower Upper
216 100
214 77.3 64.4 96.6
218 46.3 38.1 57.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#76
Pentachlorobenzene
Concen: 24.265 ng/uL
RT: 14.907 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:250 Resp: 226867
Ion Ratio Lower Upper
250 100
248 63.4 50.2 75.2
252 62.3 50.2 75.2





#77
 Carbazole
 Concen: 26.711 ng/ul
 RT: 17.724 min Scan# 2446
 Delta R.T. -0.006 min
 Lab File: BM033130.D
 Acq: 17 Nov 2021 21:36




Instrument :
 BNA_M
 ClientSampleId :
 SLCS812




Tgt Ion:167 Resp: 765352
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.4 26.2
 139 13.0 10.2 15.4

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
 Supervised By :mohammad ahmed 11/26/2021




#78
 Di-n-butylphthalate
 Concen: 28.787 ng/ul
 RT: 18.289 min Scan# 2542
 Delta R.T. -0.006 min
 Lab File: BM033130.D
 Acq: 17 Nov 2021 21:36




Tgt Ion:149 Resp: 901921
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.2 10.8
 104 5.6 4.4 6.6





#80
Fluoranthene
Concen: 25.872 ng/ul
RT: 19.371 min Scan# 2726
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:202 Resp: 101872
Ion Ratio Lower Upper
202 100
101 9.4 7.0 10.6
100 6.9 5.4 8.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#82
Pyrene
Concen: 26.026 ng/ul
RT: 19.736 min Scan# 2788
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:202 Resp: 1040501
Ion Ratio Lower Upper
202 100
101 10.2 8.3 12.5
100 9.0 7.0 10.4





#83
Butylbenzylphthalate
Concen: 28.043 ng/ul
RT: 20.624 min Scan# 2939
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812




Tgt Ion:149 Resp: 383042
Ion Ratio Lower Upper
149 100
91 73.8 62.2 93.4
206 22.7 17.0 25.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021




#84
3,3'-Dichlorobenzidine
Concen: 24.011 ng/ul
RT: 21.395 min Scan# 3070
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Tgt Ion:252 Resp: 307656
Ion Ratio Lower Upper
252 100
254 64.3 52.2 78.4
126 10.5 8.1 12.1





#85
Benzo(a)anthracene
Concen: 26.766 ng/ul
RT: 21.465 min Scan# 3082
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:228 Resp: 984600
Ion Ratio Lower Upper
228 100
229 19.5 15.8 23.8
226 26.3 21.6 32.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#86
Bis(2-ethylhexyl)phthalate
Concen: 28.425 ng/ul
RT: 21.389 min Scan# 3069
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Tgt Ion:149 Resp: 550331
Ion Ratio Lower Upper
149 100
167 27.0 22.1 33.1
279 4.1 3.4 5.2









#87
Chrysene
Concen: 26.643 ng/ul
RT: 21.518 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Instrument :
BNA_M
ClientSampleId :
SLCS812

Tgt Ion:228 Resp: 957115
Ion Ratio Lower Upper
228 100
226 29.4 23.0 34.6
229 19.4 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#89
Di-n-octyl phthalate
Concen: 27.159 ng/ul
RT: 22.300 min Scan# 3224
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:149 Resp: 929500





#90
Benzo(b)fluoranthene
Concen: 27.148 ng/ul
RT: 23.118 min Scan# 3363
Delta R.T. -0.000 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:252 Resp: 1026970
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.8
125 7.7 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#91
Benzo(k)fluoranthene
Concen: 26.334 ng/ul
RT: 23.165 min Scan# 3371
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36


Tgt Ion:252 Resp: 912577
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 8.2 5.8 8.8





#93
Benzo(a)pyrene
Concen: 26.657 ng/ul
RT: 23.730 min Scan# 3467
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36




Instrument :
BNA_M
ClientSampleId :
SLCS812



Tgt Ion:252 Resp: 953024
Ion Ratio Lower Upper
252 100
253 21.1 16.9 25.3
125 8.3 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#94
Indeno(1,2,3-cd)pyrene
Concen: 26.392 ng/ul
RT: 26.241 min Scan# 3894
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Tgt Ion:276 Resp: 1050728
Ion Ratio Lower Upper
276 100
138 19.4 13.6 20.4
277 25.9 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 26.485 ng/ul
RT: 26.253 min Scan# 3896
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Instrument : BNA_M
ClientSampleId : SLCS812

Tgt Ion:278 Resp: 902094
Ion Ratio Lower Upper
278 100
139 13.0 9.8 14.6
279 23.8 18.6 27.8

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/18/2021
Supervised By :mohammad ahmed 11/26/2021



#96
Benzo(g,h,i)perylene
Concen: 26.473 ng/ul
RT: 26.977 min Scan# 4019
Delta R.T. -0.006 min
Lab File: BM033130.D
Acq: 17 Nov 2021 21:36

Tgt Ion:276 Resp: 914759
Ion Ratio Lower Upper
276 100
138 17.3 13.1 19.7
277 24.5 18.5 27.7

