

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033150.D
 Acq On : 18 Nov 2021 10:10
 Operator : CG/JU
 Sample : M4677-08MSD
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0AB5MSD

Quant Time: Nov 18 10:40:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.972	152	218493	20.000	ng/u1	0.00
20) Naphthalene-d8	10.766	136	906454	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.589	164	575622	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.324	188	1114566	20.000	ng/u1	0.00
79) Chrysene-d12	21.477	240	954903	20.000	ng/u1	0.00
88) Perylene-d12	23.824	264	942362	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.425	96	7101	1.253	ng/uL	0.00
4) Pyridine-d5	3.843	84	85534	5.479	ng/u1	0.00
7) Phenol-d5	7.125	99	74202	4.015	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.301	67	229556	19.583	ng/u1	0.00
11) 2-Chlorophenol-d4	7.501	132	214539	15.326	ng/u1	0.00
15) 4-Methylphenol-d8	8.666	113	146526	10.216	ng/u1	0.00
21) Nitrobenzene-d5	9.131	128	143173	21.999	ng/u1	0.00
24) 2-Nitrophenol-d4	9.848	143	146142	22.422	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.383	165	290823	19.485	ng/u1	0.00
31) 4-Chloroaniline-d4	10.901	131	333267	16.790	ng/u1	0.00
46) Dimethylphthalate-d6	14.001	166	1015786	24.059	ng/u1	0.00
49) Acenaphthylene-d8	14.283	160	1226025	22.540	ng/u1	0.00
54) 4-Nitrophenol-d4	14.771	143	32259	4.691	ng/u1	0.00
60) Fluorene-d10	15.577	176	872815	23.045	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.689	200	150385	28.710	ng/u1	0.00
73) Anthracene-d10	17.424	188	1322103	24.608	ng/u1	0.00
81) Pyrene-d10	19.700	212	1457044	25.821	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.677	264	1264458	25.002	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.454	88	15329	2.664	ng/uL	90
5) Pyridine	3.860	79	98578	6.188	ng/u1	98
6) Benzaldehyde	7.113	77	242401	23.013	ng/u1	96
8) Phenol	7.148	94	87999	4.781	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.390	93	304696	20.841	ng/u1	99
12) 2-Chlorophenol	7.531	128	230935	15.984	ng/u1	98
13) 2-Methylphenol	8.401	108	177331	12.578	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.495	45	473814	20.984	ng/u1	99
16) Acetophenone	8.795	105	477375	21.086	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.778	70	268878	22.043	ng/u1	97
18) 4-Methylphenol	8.731	108	164327	11.119	ng/u1	99
19) Hexachloroethane	9.048	117	104015	15.803	ng/u1	96
22) Nitrobenzene	9.172	77	385368	21.218	ng/u1	100
23) Isophorone	9.701	82	736289	22.180	ng/u1	99
25) 2-Nitrophenol	9.883	139	157775	22.837	ng/u1	88
26) 2,4-Dimethylphenol	9.936	107	314397	17.320	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.178	93	430831	21.809	ng/u1	98
29) 2,4-Dichlorophenol	10.407	162	289864	20.027	ng/u1	95
30) Naphthalene	10.819	128	925578	19.202	ng/u1	99
32) 4-Chloroaniline	10.925	127	355116	17.709	ng/u1	99
33) Hexachlorobutadiene	11.095	225	185132	16.421	ng/u1	98
34) Caprolactam	11.725	113	10896	2.619	ng/u1	94
35) 4-Chloro-3-methylphenol	12.030	107	306468	19.445	ng/u1	98
36) 2-Methylnaphthalene	12.419	142	690844	20.685	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111721\
 Data File : BM033150.D
 Acq On : 18 Nov 2021 10:10
 Operator : CG/JU
 Sample : M4677-08MSD
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0AB5MSD

Quant Time: Nov 18 10:40:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 17 14:14:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.642	142	711388	20.866	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.783	216	400063	20.890	ng/ul	99
40) Hexachlorocyclopentadiene	12.760	237	216672	16.079	ng/ul	97
41) 2,4,6-Trichlorophenol	13.019	196	273875	24.282	ng/ul	97
42) 2,4,5-Trichlorophenol	13.089	196	293296	24.251	ng/ul	99
43) 1,1'-Biphenyl	13.424	154	997232	22.235	ng/ul	98
44) 2-Chloronaphthalene	13.471	162	750700	21.705	ng/ul	100
45) 2-Nitroaniline	13.671	65	261420	28.676	ng/ul	99
47) Dimethylphthalate	14.048	163	1062693	25.824	ng/ul	99
48) 2,6-Dinitrotoluene	14.166	165	204639	29.065	ng/ul	93
50) Acenaphthylene	14.313	152	1265016	22.918	ng/ul	99
51) 3-Nitroaniline	14.495	138	184989	26.077	ng/ul#	89
52) Acenaphthene	14.654	153	831518	23.093	ng/ul	98
53) 2,4-Dinitrophenol	14.689	184	102183	28.434	ng/ul	92
55) 4-Nitrophenol	14.783	109	32576	4.638	ng/ul	95
56) Dibenzofuran	14.983	168	1232671	23.358	ng/ul	98
57) 2,4-Dinitrotoluene	14.948	165	285449	29.479	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.207	232	252648	25.593	ng/ul	99
59) Diethylphthalate	15.401	149	1034953	25.110	ng/ul	98
61) Fluorene	15.630	166	994027	23.697	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.624	204	518947	23.703	ng/ul	99
63) 4-Nitroaniline	15.648	138	190774	27.349	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.707	198	153851	29.305	ng/ul#	98
67) N-Nitrosodiphenylamine	15.836	169	862923	26.324	ng/ul	99
68) 4-Bromophenyl-phenylether	16.518	248	327606	26.432	ng/ul	98
69) Hexachlorobenzene	16.624	284	364285	25.655	ng/ul	99
70) Atrazine	16.789	200	282901	22.334	ng/ul	100
71) Pentachlorophenol	16.971	266	220091	26.815	ng/ul	99
72) Phenanthrene	17.365	178	1561092	25.286	ng/ul	99
74) Anthracene	17.459	178	1563603	25.255	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.389	216	413912	22.609	ng/ul	99
76) Pentachlorobenzene	14.901	250	430129	23.966	ng/ul	98
77) Carbazole	17.724	167	1379856	25.087	ng/ul	99
78) Di-n-butylphthalate	18.283	149	1720779	28.611	ng/ul	99
80) Fluoranthene	19.371	202	1781801	26.953	ng/ul	99
82) Pyrene	19.730	202	1776167	26.461	ng/ul	97
83) Butylbenzylphthalate	20.618	149	721453	31.459	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.394	252	144708	6.727	ng/ul	95
85) Benzo(a)anthracene	21.459	228	1609872	26.066	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.383	149	1038742	31.955	ng/ul	99
87) Chrysene	21.512	228	1572208	26.067	ng/ul	99
89) Di-n-octyl phthalate	22.294	149	1705786	29.707	ng/ul	100
90) Benzo(b)fluoranthene	23.112	252	1623683	25.583	ng/ul	99
91) Benzo(k)fluoranthene	23.159	252	1523351	26.202	ng/ul	99
93) Benzo(a)pyrene	23.724	252	1532533	25.550	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.235	276	1730333	25.906	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.247	278	1483041	25.953	ng/ul	98
96) Benzo(g,h,i)perylene	26.971	276	1495596	25.798	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 2.664 ng/uL
RT: 3.454 min Scan# 20
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion: 88 Resp: 15329
Ion Ratio Lower Upper
88 100
43 42.4 34.9 52.3
58 74.7 70.0 105.0



#5
Pyridine
Concen: 6.188 ng/uL
RT: 3.860 min Scan# 89
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion: 79 Resp: 98578
Ion Ratio Lower Upper
79 100
52 88.2 69.1 103.7
51 43.1 35.0 52.6









#6
Benzaldehyde
Concen: 23.013 ng/ul
RT: 7.113 min Scan# 642
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion: 77 Resp: 242401
Ion Ratio Lower Upper
77 100
105 90.4 70.7 106.1
106 88.8 66.8 100.2



#8
Phenol
Concen: 4.781 ng/ul
RT: 7.148 min Scan# 648
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion: 94 Resp: 87999
Ion Ratio Lower Upper
94 100
65 31.9 27.1 40.7
66 45.4 39.4 59.2





#10

Bis(2-Chloroethyl)ether

Concen: 20.841 ng/ul

RT: 7.390 min Scan# 689

Delta R.T. -0.012 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :

H0AB5MSD





Tgt Ion: 93 Resp: 304696

Ion Ratio Lower Upper

93 100

63 83.5 67.4 101.2

95 32.0 24.8 37.2



#12

2-Chlorophenol

Concen: 15.984 ng/ul

RT: 7.531 min Scan# 713

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10




Tgt Ion: 128 Resp: 230935


Ion Ratio Lower Upper

128 100

64 53.9 42.6 63.8

130 32.5 24.4 36.6





#13

2-Methylphenol

Concen: 12.578 ng/ul

RT: 8.401 min Scan# 861

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :





H0AB5MSD

Tgt Ion:108 Resp: 177331

Ion Ratio Lower Upper

108 100

107 88.1 69.6 104.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.984 ng/ul

RT: 8.495 min Scan# 877

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10




Tgt Ion: 45 Resp: 473814


Ion Ratio Lower Upper

45 100

77 13.5 10.9 16.3

79 10.3 7.6 11.4









#16
Acetophenone
Concen: 21.086 ng/ul
RT: 8.795 min Scan# 912
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:105 Resp: 477375
Ion Ratio Lower Upper
105 100
77 83.8 68.0 102.0
51 32.5 27.8 41.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.043 ng/ul
RT: 8.778 min Scan# 925
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion: 70 Resp: 268878
Ion Ratio Lower Upper
70 100
42 63.0 52.8 79.2
101 9.6 7.1 10.7
130 19.3 15.0 22.4





#18

4-Methylphenol

Concen: 11.119 ng/ul

RT: 8.731 min Scan# 917

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :





H0AB5MSD

Tgt Ion:108 Resp: 164327

Ion Ratio Lower Upper

108 100

107 113.3 89.4 134.0



#19

Hexachloroethane

Concen: 15.803 ng/ul

RT: 9.048 min Scan# 971

Delta R.T. -0.012 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10




Tgt Ion:117 Resp: 104015


Ion Ratio Lower Upper

117 100

201 103.8 87.0 130.4

199 65.2 53.8 80.6





#22

Nitrobenzene

Concen: 21.218 ng/ul

RT: 9.172 min Scan# 992

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :

H0AB5MSD





Tgt Ion: 77 Resp: 385368

Ion Ratio Lower Upper

77 100

123 39.0 31.2 46.8

65 14.9 11.8 17.6



#23

Isophorone

Concen: 22.180 ng/ul

RT: 9.701 min Scan# 1082

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10




Tgt Ion: 82 Resp: 736289


Ion Ratio Lower Upper

82 100

95 7.9 6.6 9.8

138 16.1 12.6 19.0









#25
2-Nitrophenol
Concen: 22.837 ng/ul
RT: 9.883 min Scan# 1113
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:139 Resp: 157775
Ion Ratio Lower Upper
139 100
65 59.2 55.9 83.9
109 36.2 33.8 50.6



#26
2,4-Dimethylphenol
Concen: 17.320 ng/ul
RT: 9.936 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:107 Resp: 314397
Ion Ratio Lower Upper
107 100
121 46.9 36.2 54.2
122 77.9 63.6 95.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.809 ng/ul
 RT: 10.178 min Scan# 1163
 Delta R.T. -0.006 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10




Instrument :
 BNA_M
 ClientSampleId :
 H0AB5MSD


Tgt Ion: 93 Resp: 430831
 Ion Ratio Lower Upper
 93 100
 95 32.0 24.3 36.5
 123 11.1 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 20.027 ng/ul
 RT: 10.407 min Scan# 1202
 Delta R.T. -0.012 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10

Tgt Ion:162 Resp: 289864
 Ion Ratio Lower Upper
 162 100
 164 67.6 49.8 74.8
 98 37.5 31.2 46.8









#30
Naphthalene
Concen: 19.202 ng/ul
RT: 10.819 min Scan# 1272
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:128 Resp: 925578
Ion Ratio Lower Upper
128 100
129 10.7 9.1 13.7
127 13.1 10.2 15.2



#32
4-Chloroaniline
Concen: 17.709 ng/ul
RT: 10.925 min Scan# 1290
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:127 Resp: 355116
Ion Ratio Lower Upper
127 100
129 31.4 24.9 37.3









#33
Hexachlorobutadiene
Concen: 16.421 ng/ul
RT: 11.095 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:225 Resp: 185132
Ion Ratio Lower Upper
225 100
223 61.9 51.1 76.7
227 63.1 51.7 77.5



#34
Caprolactam
Concen: 2.619 ng/ul
RT: 11.725 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:113 Resp: 10896
Ion Ratio Lower Upper
113 100
55 193.3 157.3 235.9
56 132.5 118.2 177.4









#35
4-Chloro-3-methylphenol
Concen: 19.445 ng/ul
RT: 12.030 min Scan# 1478
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:107 Resp: 306468
Ion Ratio Lower Upper
107 100
144 23.4 19.3 28.9
142 75.0 61.4 92.2



#36
2-Methylnaphthalene
Concen: 20.685 ng/ul
RT: 12.419 min Scan# 1544
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:142 Resp: 690844
Ion Ratio Lower Upper
142 100
141 89.6 71.0 106.6









#37
1-Methylnaphthalene
Concen: 20.866 ng/ul
RT: 12.642 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:142 Resp: 711388
Ion Ratio Lower Upper
142 100
141 92.0 73.2 109.8
116 4.1 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.890 ng/ul
RT: 12.783 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:216 Resp: 400063
Ion Ratio Lower Upper
216 100
214 78.4 62.4 93.6
179 21.1 17.0 25.6
108 17.9 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 16.079 ng/ul
RT: 12.760 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:237 Resp: 216672
Ion Ratio Lower Upper
237 100
235 61.8 51.2 76.8
272 12.5 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 24.282 ng/ul
RT: 13.019 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:196 Resp: 273875
Ion Ratio Lower Upper
196 100
198 96.1 74.3 111.5
200 31.0 23.3 34.9









#42
2,4,5-Trichlorophenol
Concen: 24.251 ng/ul
RT: 13.089 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:196 Resp: 293296
Ion Ratio Lower Upper
196 100
198 95.7 75.3 112.9
200 30.5 24.6 36.8



#43
1,1'-Biphenyl
Concen: 22.235 ng/ul
RT: 13.424 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:154 Resp: 997232
Ion Ratio Lower Upper
154 100
153 39.9 33.0 49.4
76 12.6 10.1 15.1









#44
2-Chloronaphthalene
Concen: 21.705 ng/ul
RT: 13.471 min Scan# 1722
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:162 Resp: 750700
Ion Ratio Lower Upper
162 100
164 32.7 26.0 39.0
127 39.0 31.0 46.6



#45
2-Nitroaniline
Concen: 28.676 ng/ul
RT: 13.671 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion: 65 Resp: 261420
Ion Ratio Lower Upper
65 100
92 64.7 53.2 79.8
138 93.0 74.8 112.2









#47
Dimethylphthalate
Concen: 25.824 ng/ul
RT: 14.048 min Scan# 1820
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:163 Resp: 1062693
Ion Ratio Lower Upper
163 100
194 4.2 3.1 4.7
164 9.9 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 29.065 ng/ul
RT: 14.166 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:165 Resp: 204639
Ion Ratio Lower Upper
165 100
89 66.2 57.1 85.7
121 22.2 21.8 32.8









#50
Acenaphthylene
Concen: 22.918 ng/ul
RT: 14.313 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:152 Resp: 1265016
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.2 9.8 14.8



#51
3-Nitroaniline
Concen: 26.077 ng/ul
RT: 14.495 min Scan# 1897
Delta R.T. -0.000 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:138 Resp: 184989
Ion Ratio Lower Upper
138 100
108 11.6 12.6 18.8#
92 123.9 109.4 164.2









#52
Acenaphthene
Concen: 23.093 ng/ul
RT: 14.654 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:153 Resp: 831518
Ion Ratio Lower Upper
153 100
152 47.0 38.9 58.3
154 88.5 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 28.434 ng/ul
RT: 14.689 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:184 Resp: 102183
Ion Ratio Lower Upper
184 100
63 80.3 55.2 82.8
154 69.5 54.2 81.4









#55
4-Nitrophenol
Concen: 4.638 ng/ul
RT: 14.783 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:	109	Resp:	32576
Ion	Ratio	Lower	Upper
109	100		
139	106.4	85.4	128.0
65	135.9	100.9	151.3



#56
Dibenzofuran
Concen: 23.358 ng/ul
RT: 14.983 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:	168	Resp:	1232671
Ion	Ratio	Lower	Upper
168	100		
139	38.4	31.7	47.5





#57

2,4-Dinitrotoluene

Concen: 29.479 ng/ul

RT: 14.948 min Scan# 1973

Delta R.T. -0.000 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :

H0AB5MSD





Tgt Ion:165 Resp: 285449

Ion Ratio Lower Upper

165 100

63 49.1 41.5 62.3

182 2.8 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 25.593 ng/ul

RT: 15.207 min Scan# 2018

Delta R.T. -0.006 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Tgt Ion:232 Resp: 252648




Ion Ratio Lower Upper


232 100

131 44.1 34.8 52.2

130 2.3 2.1 3.1

166 31.1 24.5 36.7









#59
Diethylphthalate
Concen: 25.110 ng/ul
RT: 15.401 min Scan# 2051
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:149 Resp: 1034953
Ion Ratio Lower Upper
149 100
177 22.3 18.0 27.0
150 12.9 9.0 13.6



#61
Fluorene
Concen: 23.697 ng/ul
RT: 15.630 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:166 Resp: 994027
Ion Ratio Lower Upper
166 100
165 96.5 76.8 115.2
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 23.703 ng/ul
RT: 15.624 min Scan# 2089
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:204 Resp: 518947
Ion Ratio Lower Upper
204 100
206 32.2 25.7 38.5
141 59.8 48.6 72.8



#63
4-Nitroaniline
Concen: 27.349 ng/ul
RT: 15.648 min Scan# 2093
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:138 Resp: 190774
Ion Ratio Lower Upper
138 100
92 56.7 47.3 70.9
108 96.8 83.0 124.6









#66
4,6-Dinitro-2-methylphenol
Concen: 29.305 ng/ul
RT: 15.707 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:198 Resp: 153851
Ion Ratio Lower Upper
198 100
182 3.7 3.8 5.8#
77 26.6 22.0 33.0



#67
N-Nitrosodiphenylamine
Concen: 26.324 ng/ul
RT: 15.836 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:169 Resp: 862923
Ion Ratio Lower Upper
169 100
168 66.1 53.7 80.5
167 36.2 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 26.432 ng/ul
RT: 16.518 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:248 Resp: 327606
Ion Ratio Lower Upper
248 100
250 96.9 75.5 113.3
141 72.5 59.0 88.4



#69
Hexachlorobenzene
Concen: 25.655 ng/ul
RT: 16.624 min Scan# 2259
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:284 Resp: 364285
Ion Ratio Lower Upper
284 100
142 35.4 28.4 42.6
249 31.8 24.8 37.2









#70
Atrazine
Concen: 22.334 ng/ul
RT: 16.789 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:200 Resp: 282901
Ion Ratio Lower Upper
200 100
173 26.2 20.8 31.2
215 48.1 38.6 57.8



#71
Pentachlorophenol
Concen: 26.815 ng/ul
RT: 16.971 min Scan# 2318
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:266 Resp: 220091
Ion Ratio Lower Upper
266 100
264 63.1 50.6 75.8
268 62.5 51.1 76.7









#72
Phenanthrene
Concen: 25.286 ng/u1
RT: 17.365 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:178 Resp: 1561092
Ion Ratio Lower Upper
178 100
179 15.1 11.9 17.9
176 18.7 14.6 21.8



#74
Anthracene
Concen: 25.255 ng/u1
RT: 17.459 min Scan# 2401
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:178 Resp: 1563603
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 18.6 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.609 ng/uL
RT: 13.389 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:216 Resp: 413912
Ion Ratio Lower Upper
216 100
214 79.5 64.4 96.6
218 47.2 38.1 57.1



#76
Pentachlorobenzene
Concen: 23.966 ng/uL
RT: 14.901 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:250 Resp: 430129
Ion Ratio Lower Upper
250 100
248 64.1 50.2 75.2
252 63.9 50.2 75.2









#77
 Carbazole
 Concen: 25.087 ng/ul
 RT: 17.724 min Scan# 2446
 Delta R.T. -0.006 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10




Instrument :
 BNA_M
 ClientSampleId :
 H0AB5MSD


Tgt Ion:167 Resp: 1379856
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.2
 139 12.7 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 28.611 ng/ul
 RT: 18.283 min Scan# 2541
 Delta R.T. -0.012 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10

Tgt Ion:149 Resp: 1720779
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.2 10.8
 104 5.6 4.4 6.6









#80
Fluoranthene
Concen: 26.953 ng/ul
RT: 19.371 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:202 Resp: 1781801
Ion Ratio Lower Upper
202 100
101 9.2 7.0 10.6
100 7.2 5.4 8.2



#82
Pyrene
Concen: 26.461 ng/ul
RT: 19.730 min Scan# 2787
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:202 Resp: 1776167
Ion Ratio Lower Upper
202 100
101 11.5 8.3 12.5
100 9.7 7.0 10.4









#83
Butylbenzylphthalate
Concen: 31.459 ng/ul
RT: 20.618 min Scan# 2938
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:149 Resp: 721453
Ion Ratio Lower Upper
149 100
91 76.4 62.2 93.4
206 22.1 17.0 25.6



#84
3,3'-Dichlorobenzidine
Concen: 6.727 ng/ul
RT: 21.394 min Scan# 3070
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:252 Resp: 144708
Ion Ratio Lower Upper
252 100
254 61.2 52.2 78.4
126 10.8 8.1 12.1









#85
Benzo(a)anthracene
Concen: 26.066 ng/ul
RT: 21.459 min Scan# 3081
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:228 Resp: 1609872
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.8
226 26.8 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 31.955 ng/ul
RT: 21.383 min Scan# 3068
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:149 Resp: 1038742
Ion Ratio Lower Upper
149 100
167 27.2 22.1 33.1
279 4.1 3.4 5.2





#87

Chrysene

Concen: 26.067 ng/ul

RT: 21.512 min Scan# 3090

Delta R.T. -0.012 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Instrument :

BNA_M

ClientSampleId :

H0AB5MSD





Tgt Ion:228 Resp: 1572208

Ion Ratio Lower Upper

228 100

226 29.5 23.0 34.6

229 19.8 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 29.707 ng/ul




RT: 22.294 min Scan# 3223


Delta R.T. -0.012 min

Lab File: BM033150.D

Acq: 18 Nov 2021 10:10

Tgt Ion:149 Resp: 1705786









#90
Benzo(b)fluoranthene
Concen: 25.583 ng/ul
RT: 23.112 min Scan# 3362
Delta R.T. -0.006 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:252 Resp: 1623683
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.8
125 8.5 6.8 10.2



#91
Benzo(k)fluoranthene
Concen: 26.202 ng/ul
RT: 23.159 min Scan# 3370
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:252 Resp: 1523351
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 8.7 5.8 8.8









#93
Benzo(a)pyrene
Concen: 25.550 ng/ul
RT: 23.724 min Scan# 3466
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10




Instrument :
BNA_M
ClientSampleId :
H0AB5MSD


Tgt Ion:252 Resp: 1532533
Ion Ratio Lower Upper
252 100
253 22.3 16.9 25.3
125 9.5 6.5 9.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 25.906 ng/ul
RT: 26.235 min Scan# 3893
Delta R.T. -0.012 min
Lab File: BM033150.D
Acq: 18 Nov 2021 10:10

Tgt Ion:276 Resp: 1730333
Ion Ratio Lower Upper
276 100
138 21.6 13.6 20.4#
277 24.7 21.0 31.6









#95
 Dibenzo(a,h)anthracene
 Concen: 25.953 ng/ul
 RT: 26.247 min Scan# 3895
 Delta R.T. -0.012 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10

Instrument :
 BNA_M
 ClientSampleId :
 H0AB5MSD

Tgt Ion:278 Resp: 1483041

Ion	Ratio	Lower	Upper
278	100		
139	13.4	9.8	14.6
279	23.7	18.6	27.8



#96
 Benzo(g,h,i)perylene
 Concen: 25.798 ng/ul
 RT: 26.971 min Scan# 4018
 Delta R.T. -0.012 min
 Lab File: BM033150.D
 Acq: 18 Nov 2021 10:10

Tgt Ion:276 Resp: 1495596

Ion	Ratio	Lower	Upper
276	100		
138	17.8	13.1	19.7
277	23.9	18.5	27.7

