


Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM111723\  
Data File : BM042848.D  
Acq On : 17 Nov 2023 17:38  
Operator : MA/JU  
Sample : DFTPP182  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-BM111323.MA.M  
Title : SVOA CALIBRATION  
Last Update : Sat Nov 18 22:06:09 2023



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2478

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 77.9         | 97027      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 1288       | PASS                |
| 69             | 198             | 0.00            | 100             | 61.0         | 75909      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 354        | PASS                |
| 127            | 198             | 10              | 80              | 50.6         | 63003      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.6          | 752        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 124499     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 8377       | PASS                |
| 275            | 198             | 10              | 60              | 27.0         | 33605      | PASS                |
| 365            | 198             | 1               | 100             | 4.6          | 5745       | PASS                |
| 441            | 198             | 0.01            | 100             | 12.1         | 15123      | PASS                |
| 442            | 198             | 50              | 100             | 76.7         | 95507      | PASS                |
| 443            | 442             | 15              | 24              | 19.0         | 18140      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 241.0     |
| 44.00  | 430.0     |
| 48.90  | 180.0     |
| 73.00  | 426.0     |
| 147.00 | 194.0     |
| 207.00 | 821.0     |
| 209.00 | 197.0     |
| 281.00 | 154.0     |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP182

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 186.0     |
| 37.00  | 755.0     |
| 38.00  | 1598.0    |
| 39.10  | 6928.0    |
| 40.00  | 771.0     |
| 41.00  | 335.0     |
| 44.00  | 846.0     |
| 45.00  | 207.0     |
| 49.00  | 953.0     |
| 50.10  | 26664.0   |
| 51.10  | 88064.0   |
| 52.00  | 4623.0    |
| 55.00  | 564.0     |
| 56.00  | 1962.0    |
| 57.00  | 4671.0    |
| 58.00  | 185.0     |
| 61.00  | 836.0     |
| 62.00  | 1204.0    |
| 63.10  | 2808.0    |
| 64.00  | 451.0     |
| 65.10  | 1126.0    |
| 66.00  | 217.0     |
| 68.00  | 1149.0    |
| 69.00  | 67280.0   |
| 70.00  | 280.0     |
| 73.10  | 1068.0    |
| 74.10  | 9140.0    |
| 75.00  | 11791.0   |
| 76.10  | 4451.0    |
| 77.10  | 83728.0   |
| 78.10  | 6035.0    |
| 79.00  | 4773.0    |
| 80.00  | 3661.0    |
| 81.00  | 5861.0    |
| 82.00  | 1567.0    |
| 83.00  | 732.0     |
| 84.00  | 327.0     |
| 85.10  | 575.0     |
| 86.00  | 1546.0    |
| 87.00  | 868.0     |
| 88.10  | 343.0     |
| 91.00  | 1165.0    |
| 92.00  | 1435.0    |
| 93.00  | 10898.0   |
| 94.10  | 600.0     |
| 96.00  | 576.0     |
| 96.90  | 203.0     |
| 98.00  | 7709.0    |
| 99.00  | 5534.0    |
| 100.00 | 418.0     |
| 101.00 | 2599.0    |
| 101.90 | 211.0     |
| 103.00 | 590.0     |
| 104.00 | 1778.0    |
| 105.00 | 1464.0    |
| 106.00 | 428.0     |
| 107.00 | 20608.0   |
| 108.00 | 3148.0    |
| 109.00 | 471.0     |
| 110.00 | 27008.0   |
| 111.00 | 4800.0    |
| 111.90 | 747.0     |
| 112.90 | 205.0     |
| 116.00 | 1133.0    |
| 117.00 | 27848.0   |
| 118.00 | 1943.0    |
| 119.10 | 188.0     |
| 120.90 | 180.0     |
| 122.00 | 1213.0    |
| 123.00 | 2210.0    |
| 123.90 | 1006.0    |
| 125.10 | 828.0     |
| 126.10 | 246.0     |
| 127.00 | 49984.0   |
| 128.10 | 3657.0    |
| 129.00 | 21800.0   |
| 130.00 | 1947.0    |
| 131.00 | 386.0     |
| 132.00 | 178.0     |
| 133.00 | 226.0     |
| 133.90 | 599.0     |
| 135.00 | 1660.0    |
| 136.00 | 851.0     |
| 137.00 | 1275.0    |
| 138.10 | 206.0     |
| 139.00 | 156.0     |
| 139.80 | 320.0     |
| 141.00 | 2843.0    |
| 142.00 | 921.0     |
| 143.00 | 727.0     |
| 144.00 | 268.0     |
| 144.90 | 174.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |         |
|--------|---------|
| 146.00 | 815.0   |
| 147.00 | 1638.0  |
| 148.00 | 5576.0  |
| 149.00 | 900.0   |
| 150.00 | 319.0   |
| 151.10 | 458.0   |
| 151.80 | 197.0   |
| 153.00 | 1143.0  |
| 154.00 | 598.0   |
| 155.00 | 1571.0  |
| 156.00 | 2416.0  |
| 156.90 | 398.0   |
| 157.90 | 615.0   |
| 158.90 | 515.0   |
| 160.00 | 1045.0  |
| 161.00 | 1830.0  |
| 162.00 | 491.0   |
| 164.90 | 1156.0  |
| 166.00 | 1026.0  |
| 167.00 | 9428.0  |
| 168.00 | 5903.0  |
| 168.90 | 716.0   |
| 170.90 | 201.0   |
| 172.00 | 665.0   |
| 173.10 | 792.0   |
| 174.00 | 1331.0  |
| 175.00 | 2502.0  |
| 176.10 | 641.0   |
| 177.00 | 1096.0  |
| 177.90 | 443.0   |
| 179.00 | 5194.0  |
| 180.00 | 3090.0  |
| 181.00 | 1449.0  |
| 182.00 | 208.0   |
| 183.10 | 225.0   |
| 183.90 | 447.0   |
| 185.00 | 1974.0  |
| 186.00 | 14968.0 |
| 187.00 | 4301.0  |
| 188.10 | 588.0   |
| 188.90 | 1072.0  |
| 190.00 | 187.0   |
| 191.00 | 730.0   |
| 192.00 | 1779.0  |
| 193.00 | 1989.0  |
| 194.00 | 364.0   |
| 194.90 | 196.0   |
| 196.00 | 2937.0  |
| 197.00 | 637.0   |
| 198.00 | 88304.0 |
| 199.00 | 5998.0  |
| 199.90 | 646.0   |
| 201.30 | 226.0   |
| 201.70 | 333.0   |
| 203.00 | 891.0   |
| 204.10 | 4460.0  |
| 205.00 | 7783.0  |
| 206.10 | 27608.0 |
| 207.10 | 4792.0  |
| 208.00 | 1239.0  |
| 208.90 | 472.0   |
| 210.00 | 507.0   |
| 211.00 | 1429.0  |
| 211.90 | 252.0   |
| 214.90 | 376.0   |
| 216.00 | 698.0   |
| 217.00 | 8800.0  |
| 218.00 | 1187.0  |
| 221.00 | 2824.0  |
| 222.00 | 1047.0  |
| 223.00 | 1874.0  |
| 224.10 | 15940.0 |
| 225.00 | 4122.0  |
| 226.00 | 456.0   |
| 227.00 | 8112.0  |
| 228.00 | 1121.0  |
| 229.00 | 1611.0  |
| 229.80 | 175.0   |
| 231.00 | 630.0   |
| 233.90 | 298.0   |
| 234.90 | 605.0   |
| 235.90 | 432.0   |
| 237.00 | 955.0   |
| 238.00 | 228.0   |
| 239.00 | 334.0   |
| 240.00 | 341.0   |
| 241.00 | 467.0   |
| 242.00 | 1079.0  |
| 243.10 | 938.0   |
| 244.10 | 11577.0 |
| 245.00 | 1679.0  |
| 246.00 | 2828.0  |
| 246.90 | 355.0   |
| 249.00 | 424.0   |
| 251.90 | 154.0   |
| 253.00 | 256.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |         |
|--------|---------|
| 254.00 | 419.0   |
| 255.00 | 57280.0 |
| 256.10 | 7940.0  |
| 257.00 | 885.0   |
| 258.00 | 4134.0  |
| 259.00 | 731.0   |
| 265.00 | 1856.0  |
| 265.90 | 332.0   |
| 272.00 | 150.0   |
| 273.00 | 1650.0  |
| 274.00 | 4785.0  |
| 275.00 | 21440.0 |
| 276.00 | 3161.0  |
| 277.00 | 2250.0  |
| 278.00 | 425.0   |
| 281.00 | 239.0   |
| 282.00 | 227.0   |
| 283.00 | 285.0   |
| 285.00 | 357.0   |
| 293.00 | 684.0   |
| 293.90 | 166.0   |
| 295.00 | 172.0   |
| 296.00 | 8407.0  |
| 297.00 | 1405.0  |
| 303.00 | 868.0   |
| 304.10 | 358.0   |
| 314.00 | 351.0   |
| 314.90 | 1120.0  |
| 316.00 | 560.0   |
| 321.00 | 250.0   |
| 323.00 | 2090.0  |
| 323.90 | 309.0   |
| 327.00 | 456.0   |
| 332.80 | 247.0   |
| 334.00 | 1361.0  |
| 334.90 | 297.0   |
| 341.00 | 274.0   |
| 346.00 | 592.0   |
| 352.00 | 521.0   |
| 353.00 | 350.0   |
| 354.10 | 569.0   |
| 365.00 | 3234.0  |
| 366.00 | 373.0   |
| 370.90 | 159.0   |
| 372.00 | 1045.0  |
| 373.10 | 165.0   |
| 383.00 | 271.0   |
| 402.00 | 358.0   |
| 403.00 | 554.0   |
| 404.00 | 265.0   |
| 421.10 | 317.0   |
| 421.90 | 434.0   |
| 423.10 | 2835.0  |
| 424.00 | 1261.0  |
| 425.00 | 158.0   |
| 441.10 | 7456.0  |
| 442.10 | 48448.0 |
| 443.10 | 10204.0 |
| 444.00 | 1078.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

| m/z    | Abundance |
|--------|-----------|
| 35.90  | 172.0     |
| 37.00  | 674.0     |
| 38.00  | 1902.0    |
| 39.10  | 9330.0    |
| 40.00  | 874.0     |
| 41.10  | 358.0     |
| 44.00  | 1004.0    |
| 45.00  | 386.0     |
| 49.00  | 1229.0    |
| 50.10  | 37744.0   |
| 51.10  | 123768.0  |
| 52.00  | 6244.0    |
| 53.00  | 257.0     |
| 55.00  | 652.0     |
| 56.00  | 3450.0    |
| 57.00  | 7621.0    |
| 58.00  | 405.0     |
| 61.00  | 1259.0    |
| 62.00  | 1574.0    |
| 63.10  | 4367.0    |
| 64.00  | 613.0     |
| 65.00  | 1988.0    |
| 66.00  | 225.0     |
| 66.90  | 157.0     |
| 68.00  | 1623.0    |
| 69.00  | 97712.0   |
| 70.00  | 506.0     |
| 73.00  | 1580.0    |
| 74.00  | 12923.0   |
| 75.00  | 17656.0   |
| 76.10  | 6685.0    |
| 77.10  | 122704.0  |
| 78.10  | 8468.0    |
| 79.00  | 7502.0    |
| 80.00  | 5941.0    |
| 81.00  | 7942.0    |
| 82.00  | 2137.0    |
| 83.00  | 1697.0    |
| 83.90  | 375.0     |
| 85.00  | 973.0     |
| 86.00  | 2420.0    |
| 87.00  | 940.0     |
| 88.00  | 308.0     |
| 89.00  | 289.0     |
| 91.00  | 1325.0    |
| 92.00  | 1908.0    |
| 93.00  | 16333.0   |
| 94.00  | 1015.0    |
| 95.00  | 239.0     |
| 96.00  | 660.0     |
| 97.00  | 328.0     |
| 98.00  | 12277.0   |
| 99.00  | 8119.0    |
| 100.00 | 929.0     |
| 101.00 | 3743.0    |
| 102.00 | 180.0     |
| 102.90 | 1266.0    |
| 104.00 | 2453.0    |
| 105.00 | 2611.0    |
| 106.00 | 740.0     |
| 107.00 | 31792.0   |
| 108.00 | 4205.0    |
| 109.00 | 680.0     |
| 110.00 | 43856.0   |
| 111.00 | 7923.0    |
| 112.00 | 1112.0    |
| 113.10 | 440.0     |
| 115.10 | 215.0     |
| 116.00 | 1614.0    |
| 117.00 | 45056.0   |
| 118.00 | 3540.0    |
| 119.00 | 413.0     |
| 119.90 | 359.0     |
| 122.00 | 1995.0    |
| 123.00 | 2959.0    |
| 124.00 | 1417.0    |
| 125.00 | 1372.0    |
| 126.10 | 287.0     |
| 127.00 | 82944.0   |
| 128.00 | 6498.0    |
| 129.00 | 38640.0   |
| 130.00 | 3421.0    |
| 131.00 | 673.0     |
| 132.00 | 169.0     |
| 132.90 | 261.0     |
| 134.00 | 979.0     |
| 135.00 | 2968.0    |
| 136.00 | 1016.0    |
| 137.00 | 1852.0    |
| 138.00 | 278.0     |
| 139.00 | 187.0     |
| 139.90 | 620.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |          |
|--------|----------|
| 141.00 | 5294.0   |
| 142.00 | 1763.0   |
| 142.90 | 1062.0   |
| 144.10 | 346.0    |
| 145.00 | 208.0    |
| 146.00 | 1309.0   |
| 147.00 | 3098.0   |
| 148.00 | 9756.0   |
| 149.00 | 1512.0   |
| 150.00 | 323.0    |
| 151.00 | 550.0    |
| 151.90 | 354.0    |
| 153.00 | 1707.0   |
| 154.00 | 1044.0   |
| 155.00 | 3127.0   |
| 156.00 | 3956.0   |
| 157.00 | 772.0    |
| 158.00 | 911.0    |
| 158.90 | 768.0    |
| 160.00 | 1576.0   |
| 161.00 | 2537.0   |
| 162.00 | 617.0    |
| 163.00 | 215.0    |
| 163.90 | 180.0    |
| 165.00 | 2198.0   |
| 166.00 | 1526.0   |
| 167.00 | 16904.0  |
| 168.00 | 9981.0   |
| 169.00 | 1015.0   |
| 170.00 | 384.0    |
| 171.00 | 329.0    |
| 172.00 | 987.0    |
| 173.00 | 1138.0   |
| 174.00 | 2603.0   |
| 175.00 | 4257.0   |
| 176.10 | 914.0    |
| 177.00 | 2041.0   |
| 178.00 | 749.0    |
| 179.00 | 8968.0   |
| 180.00 | 5571.0   |
| 181.00 | 2452.0   |
| 182.00 | 441.0    |
| 182.70 | 170.0    |
| 184.00 | 453.0    |
| 185.00 | 3934.0   |
| 186.00 | 26848.0  |
| 187.00 | 8338.0   |
| 188.10 | 912.0    |
| 189.00 | 1972.0   |
| 190.10 | 496.0    |
| 191.00 | 1231.0   |
| 192.00 | 2602.0   |
| 193.00 | 3354.0   |
| 194.00 | 705.0    |
| 195.10 | 251.0    |
| 196.00 | 5387.0   |
| 197.10 | 927.0    |
| 198.00 | 160576.0 |
| 199.00 | 10613.0  |
| 200.00 | 978.0    |
| 201.50 | 832.0    |
| 203.00 | 1811.0   |
| 204.00 | 8253.0   |
| 205.10 | 13523.0  |
| 206.10 | 51824.0  |
| 207.10 | 7914.0   |
| 208.00 | 2327.0   |
| 209.00 | 946.0    |
| 210.00 | 993.0    |
| 211.00 | 2117.0   |
| 213.10 | 186.0    |
| 215.00 | 626.0    |
| 216.00 | 1575.0   |
| 217.00 | 17168.0  |
| 218.00 | 2274.0   |
| 220.00 | 230.0    |
| 221.00 | 5725.0   |
| 221.90 | 1840.0   |
| 223.00 | 3633.0   |
| 224.10 | 30688.0  |
| 225.10 | 8299.0   |
| 226.10 | 866.0    |
| 227.00 | 15247.0  |
| 228.00 | 2418.0   |
| 229.00 | 3018.0   |
| 230.00 | 382.0    |
| 231.00 | 958.0    |
| 233.00 | 329.0    |
| 234.00 | 1055.0   |
| 234.90 | 1069.0   |
| 236.00 | 907.0    |
| 237.10 | 1380.0   |
| 237.90 | 265.0    |
| 239.00 | 538.0    |
| 240.00 | 448.0    |
| 241.00 | 676.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |          |
|--------|----------|
| 242.10 | 1844.0   |
| 243.00 | 1979.0   |
| 244.10 | 22920.0  |
| 245.00 | 3082.0   |
| 246.00 | 5358.0   |
| 247.00 | 901.0    |
| 247.90 | 273.0    |
| 249.00 | 731.0    |
| 252.00 | 196.0    |
| 253.00 | 591.0    |
| 254.00 | 590.0    |
| 255.00 | 112904.0 |
| 256.00 | 16568.0  |
| 257.00 | 1191.0   |
| 258.00 | 8390.0   |
| 259.00 | 1043.0   |
| 260.00 | 215.0    |
| 260.90 | 162.0    |
| 264.00 | 152.0    |
| 265.00 | 3365.0   |
| 266.00 | 482.0    |
| 270.00 | 202.0    |
| 271.10 | 268.0    |
| 272.00 | 299.0    |
| 273.00 | 3633.0   |
| 274.00 | 8352.0   |
| 275.00 | 42184.0  |
| 276.00 | 5634.0   |
| 277.00 | 5228.0   |
| 278.00 | 1025.0   |
| 279.00 | 252.0    |
| 281.00 | 230.0    |
| 281.90 | 164.0    |
| 283.00 | 545.0    |
| 284.10 | 269.0    |
| 285.00 | 582.0    |
| 292.10 | 166.0    |
| 293.00 | 1346.0   |
| 294.00 | 284.0    |
| 295.00 | 274.0    |
| 296.00 | 17584.0  |
| 297.00 | 2364.0   |
| 301.00 | 244.0    |
| 302.00 | 268.0    |
| 303.10 | 1608.0   |
| 304.10 | 446.0    |
| 307.90 | 243.0    |
| 309.90 | 158.0    |
| 313.10 | 167.0    |
| 314.00 | 837.0    |
| 315.00 | 2131.0   |
| 316.00 | 999.0    |
| 317.00 | 177.0    |
| 321.00 | 443.0    |
| 322.10 | 180.0    |
| 323.10 | 4640.0   |
| 324.10 | 874.0    |
| 327.00 | 1023.0   |
| 328.10 | 470.0    |
| 332.00 | 433.0    |
| 333.00 | 429.0    |
| 334.00 | 2615.0   |
| 335.10 | 787.0    |
| 341.00 | 546.0    |
| 346.00 | 996.0    |
| 347.00 | 278.0    |
| 352.10 | 1461.0   |
| 353.10 | 822.0    |
| 354.00 | 1133.0   |
| 354.90 | 262.0    |
| 365.10 | 7066.0   |
| 366.00 | 1002.0   |
| 371.00 | 316.0    |
| 372.10 | 2071.0   |
| 373.00 | 500.0    |
| 383.00 | 651.0    |
| 390.10 | 232.0    |
| 402.00 | 928.0    |
| 403.00 | 1249.0   |
| 404.10 | 299.0    |
| 421.00 | 703.0    |
| 422.00 | 1089.0   |
| 423.10 | 5968.0   |
| 424.10 | 1935.0   |
| 425.00 | 208.0    |
| 441.10 | 17960.0  |
| 442.10 | 109648.0 |
| 443.10 | 21288.0  |
| 444.10 | 2238.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 417.0     |
| 38.10  | 1208.0    |
| 39.10  | 5788.0    |
| 40.00  | 725.0     |
| 43.00  | 157.0     |
| 44.00  | 811.0     |
| 45.00  | 243.0     |
| 49.00  | 990.0     |
| 50.10  | 24488.0   |
| 51.10  | 79248.0   |
| 52.00  | 3673.0    |
| 55.00  | 435.0     |
| 56.00  | 1987.0    |
| 57.00  | 4471.0    |
| 58.10  | 169.0     |
| 61.00  | 861.0     |
| 62.00  | 990.0     |
| 63.00  | 2581.0    |
| 64.00  | 367.0     |
| 65.00  | 1121.0    |
| 65.90  | 270.0     |
| 68.00  | 1092.0    |
| 69.00  | 62736.0   |
| 70.00  | 276.0     |
| 73.00  | 981.0     |
| 74.10  | 8228.0    |
| 75.00  | 11214.0   |
| 76.00  | 4719.0    |
| 77.10  | 82072.0   |
| 78.00  | 5296.0    |
| 79.00  | 5016.0    |
| 80.00  | 4066.0    |
| 81.00  | 5074.0    |
| 82.00  | 1282.0    |
| 83.00  | 1125.0    |
| 84.00  | 269.0     |
| 85.00  | 761.0     |
| 86.00  | 1851.0    |
| 87.00  | 686.0     |
| 88.10  | 190.0     |
| 91.00  | 1264.0    |
| 92.00  | 1381.0    |
| 93.00  | 9914.0    |
| 93.90  | 748.0     |
| 94.80  | 194.0     |
| 96.00  | 405.0     |
| 98.10  | 8326.0    |
| 99.00  | 5464.0    |
| 99.90  | 388.0     |
| 101.00 | 2708.0    |
| 103.00 | 865.0     |
| 104.00 | 1646.0    |
| 105.00 | 1746.0    |
| 105.90 | 512.0     |
| 107.00 | 21056.0   |
| 108.00 | 2616.0    |
| 108.90 | 547.0     |
| 110.00 | 28488.0   |
| 111.00 | 5384.0    |
| 112.00 | 913.0     |
| 116.00 | 1332.0    |
| 117.00 | 30456.0   |
| 118.00 | 1879.0    |
| 118.90 | 211.0     |
| 120.00 | 312.0     |
| 122.00 | 1519.0    |
| 122.90 | 2036.0    |
| 124.00 | 1218.0    |
| 125.00 | 1054.0    |
| 126.10 | 236.0     |
| 127.00 | 56080.0   |
| 128.00 | 4267.0    |
| 129.00 | 25368.0   |
| 130.00 | 2170.0    |
| 131.00 | 459.0     |
| 132.00 | 272.0     |
| 134.00 | 903.0     |
| 135.00 | 1899.0    |
| 135.90 | 745.0     |
| 137.10 | 1232.0    |
| 140.00 | 322.0     |
| 141.00 | 3356.0    |
| 142.00 | 968.0     |
| 143.00 | 1012.0    |
| 144.00 | 164.0     |
| 145.00 | 296.0     |
| 146.00 | 779.0     |
| 147.00 | 1865.0    |
| 148.00 | 7498.0    |
| 149.00 | 1351.0    |
| 149.90 | 320.0     |
| 151.10 | 378.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |          |
|--------|----------|
| 151.90 | 214.0    |
| 152.90 | 1264.0   |
| 154.00 | 904.0    |
| 155.00 | 2050.0   |
| 156.00 | 2858.0   |
| 157.00 | 462.0    |
| 158.00 | 759.0    |
| 159.00 | 467.0    |
| 160.00 | 1109.0   |
| 161.00 | 1654.0   |
| 162.00 | 518.0    |
| 162.90 | 216.0    |
| 163.90 | 207.0    |
| 165.00 | 1524.0   |
| 166.10 | 1359.0   |
| 167.00 | 12485.0  |
| 168.00 | 6766.0   |
| 169.00 | 824.0    |
| 169.90 | 193.0    |
| 171.00 | 366.0    |
| 171.90 | 890.0    |
| 173.10 | 1130.0   |
| 174.00 | 1623.0   |
| 175.00 | 3030.0   |
| 176.00 | 591.0    |
| 177.00 | 1254.0   |
| 178.00 | 525.0    |
| 179.00 | 6059.0   |
| 180.00 | 4047.0   |
| 181.00 | 2093.0   |
| 182.00 | 244.0    |
| 183.90 | 538.0    |
| 185.00 | 2763.0   |
| 186.00 | 20552.0  |
| 187.00 | 5611.0   |
| 188.00 | 557.0    |
| 189.00 | 1480.0   |
| 189.80 | 212.0    |
| 191.00 | 881.0    |
| 192.00 | 2405.0   |
| 193.00 | 2427.0   |
| 193.90 | 411.0    |
| 195.00 | 288.0    |
| 196.00 | 3726.0   |
| 196.90 | 691.0    |
| 198.00 | 124616.0 |
| 199.00 | 8521.0   |
| 200.00 | 803.0    |
| 201.40 | 480.0    |
| 202.90 | 1260.0   |
| 204.00 | 6638.0   |
| 205.10 | 10697.0  |
| 206.10 | 39456.0  |
| 207.10 | 7022.0   |
| 208.10 | 1921.0   |
| 209.00 | 731.0    |
| 210.10 | 901.0    |
| 211.00 | 1877.0   |
| 211.90 | 196.0    |
| 215.00 | 560.0    |
| 216.00 | 1044.0   |
| 217.00 | 12727.0  |
| 218.00 | 1481.0   |
| 220.10 | 152.0    |
| 221.00 | 4887.0   |
| 222.00 | 1586.0   |
| 223.00 | 3106.0   |
| 224.10 | 25040.0  |
| 225.00 | 6330.0   |
| 226.00 | 792.0    |
| 227.00 | 11544.0  |
| 228.00 | 1425.0   |
| 229.00 | 2188.0   |
| 230.10 | 304.0    |
| 231.10 | 1063.0   |
| 231.90 | 158.0    |
| 233.00 | 199.0    |
| 234.00 | 889.0    |
| 235.00 | 771.0    |
| 236.00 | 664.0    |
| 237.00 | 1059.0   |
| 239.00 | 402.0    |
| 240.00 | 422.0    |
| 241.00 | 560.0    |
| 242.10 | 1423.0   |
| 243.00 | 1488.0   |
| 244.10 | 19168.0  |
| 245.10 | 2689.0   |
| 246.00 | 4485.0   |
| 247.00 | 1008.0   |
| 249.00 | 723.0    |
| 250.10 | 197.0    |
| 251.00 | 225.0    |
| 252.10 | 183.0    |
| 253.00 | 491.0    |
| 254.00 | 592.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182

|        |          |
|--------|----------|
| 255.00 | 94440.0  |
| 256.00 | 14018.0  |
| 257.00 | 1014.0   |
| 258.00 | 7491.0   |
| 259.00 | 1172.0   |
| 260.00 | 252.0    |
| 261.00 | 195.0    |
| 264.00 | 177.0    |
| 265.00 | 3365.0   |
| 266.00 | 477.0    |
| 267.00 | 157.0    |
| 269.90 | 150.0    |
| 271.00 | 216.0    |
| 272.00 | 223.0    |
| 273.00 | 2974.0   |
| 274.00 | 7616.0   |
| 275.10 | 37192.0  |
| 276.00 | 4788.0   |
| 277.00 | 4311.0   |
| 278.00 | 693.0    |
| 281.00 | 218.0    |
| 283.00 | 515.0    |
| 284.10 | 254.0    |
| 285.00 | 559.0    |
| 286.00 | 153.0    |
| 289.00 | 188.0    |
| 292.10 | 198.0    |
| 293.00 | 591.0    |
| 294.00 | 249.0    |
| 295.00 | 206.0    |
| 296.00 | 15203.0  |
| 297.00 | 2004.0   |
| 300.90 | 223.0    |
| 302.10 | 210.0    |
| 303.00 | 1772.0   |
| 304.10 | 393.0    |
| 307.90 | 227.0    |
| 309.00 | 178.0    |
| 310.00 | 179.0    |
| 314.10 | 767.0    |
| 315.00 | 1799.0   |
| 316.10 | 850.0    |
| 316.90 | 198.0    |
| 321.10 | 479.0    |
| 322.10 | 168.0    |
| 323.10 | 4113.0   |
| 324.10 | 730.0    |
| 327.00 | 816.0    |
| 328.00 | 449.0    |
| 332.00 | 383.0    |
| 333.00 | 407.0    |
| 334.00 | 2843.0   |
| 335.00 | 839.0    |
| 341.00 | 445.0    |
| 346.00 | 1191.0   |
| 347.00 | 268.0    |
| 352.10 | 1400.0   |
| 353.00 | 664.0    |
| 354.10 | 1143.0   |
| 355.10 | 300.0    |
| 365.00 | 6936.0   |
| 366.00 | 862.0    |
| 371.00 | 205.0    |
| 372.10 | 1967.0   |
| 373.10 | 489.0    |
| 383.10 | 373.0    |
| 384.00 | 166.0    |
| 390.10 | 306.0    |
| 391.00 | 207.0    |
| 402.00 | 770.0    |
| 403.10 | 1126.0   |
| 404.10 | 489.0    |
| 421.10 | 1035.0   |
| 422.00 | 959.0    |
| 423.10 | 5895.0   |
| 424.10 | 1946.0   |
| 425.10 | 247.0    |
| 441.10 | 19952.0  |
| 442.10 | 128424.0 |
| 443.10 | 22928.0  |
| 444.10 | 2150.0   |
| 445.00 | 216.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP182