

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111820\
 Data File : BM027829.D
 Acq On : 18 Nov 2020 14:05
 Operator : JU/CG
 Sample : SST00502
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST005002

Quant Time: Nov 18 16:12:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111820.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 16:04:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.33	152	41540	20.00	ng/ul	0.00
20) Naphthalene-d8	11.17	136	181669	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.95	164	120362	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.67	188	257661	20.00	ng/ul	0.00
79) Chrysene-d12	21.77	240	211917	20.00	ng/ul	0.00
88) Perylene-d12	24.20	264	205693	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.50	96	1839	1.75	ng/uL	0.00
4) Pyridine-d5	0.00	84	0d	0.00	ng/ul	
7) Phenol-d5	0.00	99	0d	0.00	ng/ul	
9) Bis-(2-Chloroethyl)ether-d	0.00	67	0d	0.00	ng/ul	
11) 2-Chlorophenol-d4	7.86	132	13840	5.25	ng/ul	0.00
15) 4-Methylphenol-d8	0.00	113	0d	0.00	ng/ul	
21) Nitrobenzene-d5	9.51	128	7160	4.97	ng/ul	0.00
24) 2-Nitrophenol-d4	10.24	143	7904	5.08	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.78	165	15731	4.92	ng/ul	0.00
31) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul	
46) Dimethylphthalate-d6	14.35	166	48962	5.06	ng/ul	0.00
49) Acenaphthylene-d8	14.65	160	57666	4.91	ng/ul	0.00
54) 4-Nitrophenol-d4	0.00	143	0d	0.00	ng/ul	
60) Fluorene-d10	15.93	176	41967	4.73	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	0.00	200	0d	0.00	ng/ul	
73) Anthracene-d10	17.76	188	63901	4.97	ng/ul	0.00
81) Pyrene-d10	20.02	212	64791	5.42	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.04	264	55054	4.74	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.53	88	2106	1.794	ng/uL# 90
12) 2-Chlorophenol	7.89	128	14578	5.297	ng/ul 97
17) N-Nitroso-di-n-propylamine	9.15	70	15091	6.821	ng/ul 94
19) Hexachloroethane	9.45	117	7157	5.233	ng/ul 94
22) Nitrobenzene	9.56	77	21884	4.988	ng/ul 98
23) Isophorone	10.08	82	40756	5.876	ng/ul 99
25) 2-Nitrophenol	10.27	139	8401	5.091	ng/ul 97
26) 2,4-Dimethylphenol	10.32	107	19596	4.950	ng/ul 97
27) Bis(2-Chloroethoxy)methane	10.56	93	24180	6.265	ng/ul 99
29) 2,4-Dichlorophenol	10.80	162	15061	4.864	ng/ul 96
30) Naphthalene	11.23	128	49875	4.888	ng/ul 99
33) Hexachlorobutadiene	11.50	225	10973	4.278	ng/ul 94
35) 4-Chloro-3-methylphenol	12.41	107	18828	5.955	ng/ul# 95
36) 2-Methylnaphthalene	12.81	142	36599	5.547	ng/ul 98
37) 1-Methylnaphthalene	13.02	142	35147	5.326	ng/ul 91
39) 1,2,4,5-Tetrachlorobenzene	13.17	216	19665	4.416	ng/ul 99
41) 2,4,6-Trichlorophenol	13.40	196	12845	4.674	ng/ul 94
42) 2,4,5-Trichlorophenol	13.47	196	13911	4.696	ng/ul 98
43) 1,1'-Biphenyl	13.79	154	48890	4.814	ng/ul 99
44) 2-Chloronaphthalene	13.85	162	38461	4.772	ng/ul 97
45) 2-Nitroaniline	14.03	65	12882	4.857	ng/ul 91
47) Dimethylphthalate	14.39	163	51054	5.131	ng/ul 98

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111820\
 Data File : BM027829.D
 Acq On : 18 Nov 2020 14:05
 Operator : JU/CG
 Sample : SSTD00502
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005002

Quant Time: Nov 18 16:12:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111820.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 16:04:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
48) 2,6-Dinitrotoluene	14.51	165	9549	4.856	ng/ul	92
50) Acenaphthylene	14.68	152	56303	4.872	ng/ul	98
52) Acenaphthene	15.02	153	39769	4.809	ng/ul	100
56) Dibenzofuran	15.35	168	57162	4.804	ng/ul	98
57) 2,4-Dinitrotoluene	15.29	165	13270	4.675	ng/ul#	90
58) 2,3,4,6-Tetrachlorophenol	15.56	232	11755	4.607	ng/ul#	89
59) Diethylphthalate	15.73	149	52797	5.220	ng/ul	99
61) Fluorene	15.99	166	47397	4.745	ng/ul	98
62) 4-Chlorophenyl-phenylether	15.97	204	23887	4.393	ng/ul	95
67) N-Nitrosodiphenylamine	16.18	169	39421	5.088	ng/ul	98
68) 4-Bromophenyl-phenylether	16.86	248	14526	4.818	ng/ul	99
69) Hexachlorobenzene	16.98	284	16003	4.709	ng/ul	99
72) Phenanthrene	17.71	178	75008	4.860	ng/ul	98
74) Anthracene	17.80	178	77772	5.068	ng/ul	99
75) 1,2,3,4-Tetrachlorobenzene	13.77	216	19303	4.609	ng/uL	95
76) Pentachlorobenzene	15.27	250	19509	4.845	ng/uL	97
78) Di-n-butylphthalate	18.59	149	88040	5.842	ng/ul	98
80) Fluoranthene	19.69	202	84457	5.523	ng/ul	100
82) Pyrene	20.04	202	87373	5.558	ng/ul	99
83) Butylbenzylphthalate	20.90	149	37366	6.566	ng/ul	98
85) Benzo(a)anthracene	21.76	228	74364	5.029	ng/ul	98
86) Bis(2-ethylhexyl)phthalate	21.66	149	51350	6.681	ng/ul	95
87) Chrysene	21.81	228	70622	4.951	ng/ul	98
90) Benzo(b)fluoranthene	23.46	252	69529	4.331	ng/ul	98
91) Benzo(k)fluoranthene	23.51	252	65098	4.346	ng/ul	99
93) Benzo(a)pyrene	24.09	252	61205	4.753	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.70	276	71148	4.359	ng/ul	98
95) Dibenzo(a,h)anthracene	26.71	278	58074	4.276	ng/ul	99
96) Benzo(a,h,i)perylene	27.47	276	57537	4.251	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111820\
Data File : BM027829.D
Acq On : 18 Nov 2020 14:05
Operator : JU/CG
Sample : SSTD00502
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD005002

Quant Time: Nov 18 16:12:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111820.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 18 16:04:33 2020
Response via : Initial Calibration

