


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM111820\
 Data File : BM027828.D
 Acq On : 18 Nov 2020 13:29
 Operator : JU/CG
 Sample : DFTPP93
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP931

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM111820.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 18 17:45:46 2020



AutoFind: Scans 2564, 2565, 2566; Background Corrected with Scan 2558

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.7	130559	PASS
68	69	0.00	2	1.7	2410	PASS
69	198	0.00	100	47.5	142056	PASS
70	69	0.00	2	0.4	623	PASS
127	198	10	80	48.9	146197	PASS
197	198	0.00	2	0.6	1663	PASS
198	198	100	100	100.0	298794	PASS
199	198	5	9	6.7	19941	PASS
275	198	10	60	26.6	79445	PASS
365	198	1	100	3.6	10608	PASS
441	198	0.01	100	12.8	38146	PASS
442	198	50	100	82.9	247829	PASS
443	442	15	24	19.4	48096	PASS

m/z	Abundance
39.90	150.0
41.00	177.0
44.00	553.0
51.00	153.0
72.90	300.0
207.00	728.0
208.10	178.0
281.00	281.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

m/z	Abundance
37.00	483.0
38.10	1487.0
39.10	9660.0
40.10	659.0
41.10	454.0
42.90	222.0
44.00	492.0
45.00	392.0
49.10	865.0
50.10	34784.0
51.10	129640.0
52.10	6661.0
53.00	230.0
55.00	690.0
56.00	4100.0
57.00	9135.0
58.00	438.0
59.10	195.0
61.00	1726.0
62.00	1768.0
63.10	5669.0
64.10	708.0
65.10	2542.0
66.00	220.0
67.00	241.0
68.10	2265.0
69.00	138816.0
70.00	625.0
73.00	1250.0
74.00	13846.0
75.00	22168.0
76.00	7026.0
77.10	140160.0
78.10	9607.0
79.00	10059.0
80.00	7168.0
81.00	11085.0
82.00	2760.0
83.00	2187.0
84.00	330.0
85.00	1638.0
86.00	2930.0
87.00	1270.0
87.90	510.0
88.90	243.0
91.00	2341.0
92.00	2626.0
93.00	17680.0
94.00	1013.0
94.90	434.0
96.00	810.0
97.10	376.0
98.00	12811.0
99.00	9850.0
100.00	871.0
101.00	5411.0
102.00	265.0
103.00	1923.0
104.00	3674.0
105.00	3556.0
106.10	1040.0
107.00	43312.0
108.00	6788.0
109.00	1454.0
110.00	73272.0
111.00	11292.0
112.00	1215.0
112.90	448.0
115.10	156.0
116.00	2168.0
117.00	35304.0
118.00	2418.0
118.90	251.0
119.90	482.0
122.00	2712.0
123.00	4305.0
124.00	1896.0
125.00	1608.0
126.20	928.0
127.00	134400.0
128.00	10451.0
129.00	54232.0
130.00	4940.0
131.00	982.0
132.10	398.0
132.90	274.0
134.00	1858.0
135.00	4483.0
136.00	1702.0
137.00	2073.0
138.00	378.0
138.90	373.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

140.00	770.0
141.00	6869.0
142.00	2118.0
143.00	1370.0
144.00	362.0
144.90	423.0
146.00	1145.0
147.00	3596.0
148.00	8458.0
149.00	1640.0
149.90	392.0
151.00	710.0
152.00	608.0
153.00	2250.0
154.00	1619.0
155.00	3546.0
156.00	5823.0
157.10	1056.0
157.90	1109.0
158.90	1015.0
160.00	2557.0
161.00	3190.0
162.00	815.0
163.00	331.0
164.10	434.0
165.00	3152.0
166.00	2182.0
167.00	12641.0
168.00	6458.0
169.00	1194.0
170.00	615.0
171.00	573.0
171.90	1222.0
173.00	1576.0
174.00	2859.0
175.00	5489.0
176.00	1525.0
177.00	2695.0
178.00	841.0
179.00	9697.0
180.00	7086.0
180.90	3066.0
181.90	503.0
182.90	408.0
184.00	734.0
185.00	4936.0
186.00	37096.0
187.00	10890.0
188.00	1169.0
189.00	2501.0
189.90	409.0
190.90	1300.0
192.00	3746.0
193.00	3856.0
193.90	726.0
195.10	376.0
196.00	7180.0
198.00	257216.0
199.00	18048.0
200.00	1489.0
201.40	1054.0
203.00	1887.0
204.00	10249.0
205.00	16600.0
206.00	71480.0
207.00	10430.0
208.00	2500.0
209.00	895.0
210.00	1111.0
211.00	2623.0
212.00	388.0
213.00	191.0
215.00	847.0
216.00	1428.0
217.00	20320.0
218.00	2791.0
219.00	212.0
221.00	9361.0
222.00	2503.0
223.00	4181.0
224.00	36616.0
225.10	9707.0
226.00	1004.0
227.00	18352.0
228.00	2725.0
229.00	3373.0
229.90	465.0
231.00	1369.0
232.00	152.0
233.00	166.0
233.90	1011.0
235.00	1219.0
236.00	680.0
237.00	1617.0
238.00	218.0
238.90	637.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

240.00	476.0
241.00	908.0
242.00	1801.0
243.00	2160.0
244.00	27936.0
245.10	3671.0
246.00	6194.0
247.00	1205.0
248.00	303.0
249.00	908.0
250.00	182.0
250.80	189.0
252.00	271.0
253.00	621.0
255.00	151808.0
256.00	22144.0
257.00	1667.0
258.00	9225.0
259.00	1741.0
260.00	308.0
261.00	224.0
264.00	231.0
265.00	3607.0
266.00	564.0
270.90	351.0
272.00	388.0
273.00	4494.0
274.00	11331.0
275.00	63856.0
276.00	8448.0
277.00	5410.0
278.00	850.0
279.00	201.0
281.90	224.0
283.00	784.0
284.00	444.0
285.00	1059.0
285.90	218.0
289.00	236.0
290.00	203.0
292.00	302.0
293.00	1047.0
293.90	357.0
295.00	482.0
296.00	17648.0
297.00	2624.0
298.10	159.0
300.90	281.0
302.00	372.0
302.90	1901.0
304.00	624.0
307.90	282.0
309.90	276.0
314.00	728.0
315.00	1879.0
316.00	1169.0
317.00	234.0
321.00	558.0
322.00	371.0
323.00	5199.0
324.00	1044.0
325.90	209.0
326.90	932.0
327.90	634.0
332.00	495.0
333.00	635.0
334.00	3721.0
334.90	738.0
341.00	717.0
342.00	253.0
346.00	1404.0
346.90	221.0
352.00	2124.0
353.00	1060.0
354.00	1562.0
354.90	288.0
359.00	164.0
365.00	7732.0
365.90	1068.0
371.00	449.0
372.00	2780.0
372.90	571.0
383.00	781.0
383.90	260.0
390.00	308.0
391.00	252.0
392.10	208.0
401.90	848.0
403.00	1502.0
404.00	631.0
420.90	1283.0
422.00	1282.0
423.00	8091.0
424.00	1993.0
440.10	272.0
441.00	25256.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

442.00	165312.0
443.00	32928.0
444.00	3095.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

m/z	Abundance
37.00	615.0
38.10	1794.0
39.10	11255.0
40.00	586.0
41.00	384.0
44.00	669.0
45.00	420.0
49.10	1086.0
50.10	39640.0
51.10	152192.0
52.10	7882.0
53.00	278.0
55.00	812.0
56.00	4632.0
57.00	11499.0
58.00	431.0
58.90	243.0
61.00	2064.0
62.00	2374.0
63.00	6900.0
64.00	1006.0
65.10	2869.0
66.00	253.0
67.10	299.0
68.00	2895.0
69.00	165760.0
70.00	738.0
73.00	1530.0
74.00	16260.0
75.00	26560.0
76.10	8116.0
77.10	172288.0
78.10	11083.0
79.00	12699.0
80.00	10259.0
81.00	13236.0
82.00	3408.0
83.00	2738.0
84.00	396.0
85.00	2364.0
86.00	3420.0
87.00	1806.0
88.00	539.0
89.00	270.0
91.00	2875.0
92.00	3086.0
93.00	21872.0
94.00	1449.0
94.90	305.0
96.00	952.0
98.00	15836.0
99.00	11385.0
100.00	1098.0
101.00	6871.0
101.90	456.0
103.00	2017.0
104.00	5371.0
105.00	4241.0
106.10	1520.0
107.00	55576.0
108.00	7727.0
109.10	1594.0
110.00	92312.0
111.00	14219.0
112.00	1693.0
112.90	417.0
115.00	232.0
116.00	3111.0
117.00	43880.0
118.00	3403.0
119.00	454.0
120.00	651.0
121.00	317.0
122.00	3690.0
123.00	5699.0
124.00	2527.0
125.00	2367.0
127.00	174208.0
128.00	13127.0
129.00	69208.0
130.00	5529.0
131.00	1132.0
132.00	545.0
132.90	280.0
134.00	2049.0
135.00	5449.0
136.00	2011.0
137.00	2835.0
138.00	595.0
139.00	359.0
140.00	745.0
141.00	9393.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

142.00	2916.0
143.00	2187.0
143.90	492.0
144.90	293.0
146.00	1549.0
147.00	5019.0
148.00	10744.0
149.00	2079.0
149.90	536.0
151.10	974.0
151.90	731.0
153.00	2924.0
154.00	1983.0
155.00	5379.0
156.00	7442.0
157.10	1422.0
158.00	1639.0
159.00	1455.0
160.00	2779.0
161.00	3978.0
161.90	844.0
163.00	300.0
163.90	402.0
165.00	3379.0
166.00	2566.0
167.00	18496.0
168.00	7714.0
169.00	1460.0
170.00	702.0
171.00	977.0
171.90	1763.0
173.00	2364.0
174.00	4029.0
175.00	7249.0
176.00	1695.0
177.00	3098.0
178.00	1114.0
179.00	14153.0
180.00	9694.0
181.00	4230.0
182.00	749.0
182.80	439.0
184.00	1003.0
185.00	6450.0
186.00	49912.0
187.00	14499.0
188.00	1413.0
189.00	3638.0
189.90	580.0
191.00	1721.0
192.00	4334.0
193.00	4869.0
194.00	1092.0
195.00	474.0
196.00	9862.0
197.10	2659.0
198.00	352064.0
199.00	23624.0
200.00	2162.0
200.80	475.0
201.50	1497.0
203.00	2633.0
204.00	13884.0
205.00	22120.0
206.00	94616.0
207.00	13177.0
208.00	3890.0
208.90	1138.0
210.00	1403.0
211.00	3744.0
211.90	607.0
212.90	258.0
214.00	166.0
215.00	1228.0
216.00	2054.0
217.00	26544.0
218.00	3748.0
219.00	345.0
221.00	12833.0
222.00	3012.0
223.00	5832.0
224.00	52336.0
225.00	13414.0
226.00	1482.0
227.00	22680.0
228.00	3246.0
229.00	4913.0
229.90	531.0
231.00	1932.0
232.10	235.0
233.00	544.0
233.90	1816.0
235.00	1456.0
236.00	1000.0
237.00	1960.0
237.80	213.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

239.00	779.0
240.00	760.0
241.00	1374.0
242.00	2562.0
243.10	2949.0
244.00	39136.0
245.00	5207.0
246.00	8473.0
247.00	1528.0
247.90	367.0
249.00	1453.0
250.00	253.0
251.00	338.0
252.00	408.0
253.00	1072.0
254.20	2538.0
255.00	213568.0
256.00	31600.0
257.00	2264.0
258.00	12829.0
259.00	2062.0
260.10	271.0
260.90	352.0
264.00	350.0
265.00	4817.0
266.00	501.0
271.00	469.0
272.00	625.0
273.00	6174.0
274.00	16106.0
275.00	92320.0
276.00	11844.0
277.00	8154.0
277.90	1155.0
278.90	366.0
280.90	328.0
281.90	292.0
283.00	850.0
284.00	569.0
285.00	1370.0
286.00	338.0
289.00	344.0
289.90	355.0
290.90	194.0
292.00	337.0
292.90	2051.0
294.00	352.0
295.10	780.0
296.00	26688.0
297.00	3622.0
297.90	172.0
300.90	235.0
302.00	435.0
303.00	2749.0
304.00	653.0
308.00	372.0
308.90	297.0
309.90	503.0
313.00	210.0
314.00	1114.0
315.00	2661.0
316.00	1624.0
317.00	300.0
321.00	848.0
322.00	473.0
323.00	7838.0
324.00	1116.0
324.90	192.0
327.00	1560.0
328.00	730.0
328.90	242.0
332.00	521.0
333.00	876.0
334.00	5282.0
335.00	1380.0
336.00	203.0
338.90	197.0
341.00	808.0
341.90	200.0
346.00	1699.0
346.90	257.0
350.80	156.0
352.00	2205.0
353.00	1894.0
354.00	2494.0
354.90	518.0
359.00	181.0
365.00	12057.0
366.00	1909.0
370.00	234.0
371.00	590.0
372.00	4234.0
373.00	891.0
383.00	1246.0
383.90	472.0
390.00	675.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

391.00	423.0
391.90	259.0
401.00	287.0
402.00	1657.0
403.00	2697.0
404.00	924.0
404.90	154.0
421.00	2001.0
422.00	2134.0
423.00	15265.0
424.00	3481.0
425.00	372.0
440.10	362.0
441.00	43400.0
442.00	281856.0
443.00	54784.0
444.00	5421.0
445.00	339.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

m/z	Abundance
37.00	415.0
38.10	1431.0
39.10	7980.0
40.00	621.0
41.00	252.0
43.10	188.0
44.00	674.0
45.00	317.0
49.00	640.0
50.10	28872.0
51.10	110304.0
52.00	5343.0
53.00	318.0
55.00	795.0
56.00	3551.0
57.00	8064.0
57.90	459.0
61.00	1340.0
62.00	1969.0
63.10	4694.0
64.10	615.0
65.00	1909.0
67.00	181.0
68.10	2070.0
69.00	121592.0
70.00	508.0
73.00	1280.0
74.00	12183.0
75.00	18976.0
76.10	5857.0
77.10	119960.0
78.00	7874.0
79.00	9217.0
80.00	7211.0
81.00	9689.0
82.00	2438.0
83.00	2292.0
83.90	165.0
85.00	1458.0
86.00	2938.0
87.10	1115.0
87.90	390.0
91.00	2141.0
92.00	2547.0
93.00	16656.0
94.00	1188.0
95.00	179.0
96.00	614.0
96.90	230.0
98.00	12527.0
99.00	8862.0
100.00	642.0
101.00	5327.0
102.10	338.0
103.00	1577.0
104.00	3689.0
105.00	3461.0
106.10	852.0
107.00	40864.0
108.00	5923.0
109.10	1423.0
110.00	67472.0
111.00	10406.0
111.90	1223.0
113.00	526.0
115.00	192.0
116.00	2387.0
117.00	33160.0
118.00	2180.0
119.00	373.0
120.00	525.0
120.90	190.0
122.00	2918.0
123.00	4070.0
124.00	2001.0
125.00	1870.0
126.20	823.0
127.00	129984.0
128.00	9762.0
129.00	54952.0
130.00	4136.0
130.90	788.0
131.90	460.0
133.00	387.0
134.00	1272.0
135.00	4384.0
136.00	1789.0
137.00	2169.0
137.80	425.0
138.80	398.0
140.00	855.0
141.00	7478.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

142.00	1608.0
143.00	1368.0
144.00	241.0
145.00	296.0
146.00	1271.0
147.00	3789.0
148.00	8330.0
149.00	1432.0
150.00	378.0
151.10	972.0
151.80	598.0
153.00	2582.0
154.00	1893.0
155.00	4018.0
156.00	5627.0
157.00	993.0
158.00	1194.0
158.90	1066.0
160.00	2559.0
161.00	3521.0
161.90	1088.0
162.90	305.0
163.90	266.0
165.00	2763.0
166.00	2122.0
167.00	14289.0
168.00	6108.0
169.00	1136.0
169.90	441.0
170.90	716.0
172.00	1081.0
173.00	1883.0
174.00	3382.0
175.00	5499.0
176.00	1551.0
177.00	2481.0
178.00	968.0
179.00	10920.0
180.00	7806.0
181.00	3786.0
182.00	343.0
182.80	270.0
183.90	920.0
185.00	5178.0
186.00	39904.0
187.00	11794.0
188.00	1154.0
189.00	3317.0
189.90	366.0
191.00	1508.0
192.00	3796.0
193.00	3820.0
194.00	824.0
195.00	311.0
196.00	7568.0
197.10	2330.0
198.00	287104.0
199.00	18152.0
200.00	1680.0
201.30	1140.0
203.00	2322.0
204.00	11552.0
205.00	18616.0
206.00	77752.0
207.00	10563.0
208.00	2800.0
209.00	1105.0
210.00	1240.0
211.00	3331.0
212.10	523.0
212.90	260.0
215.00	704.0
216.00	2120.0
217.00	22648.0
218.00	2648.0
219.00	307.0
221.00	10316.0
222.00	2426.0
223.00	5294.0
224.00	45128.0
225.00	10992.0
226.00	1368.0
227.00	20080.0
228.00	2651.0
229.00	4466.0
230.00	441.0
231.00	1542.0
232.10	176.0
233.00	380.0
234.00	1216.0
234.90	1239.0
235.90	908.0
237.00	1713.0
237.90	210.0
239.00	725.0
240.00	613.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

241.00	1052.0
242.00	2085.0
243.00	2564.0
244.00	32472.0
245.00	4362.0
246.00	6880.0
247.00	1498.0
247.90	325.0
249.00	1269.0
250.00	222.0
250.90	309.0
251.90	442.0
252.90	933.0
254.10	2035.0
255.00	183104.0
256.00	25376.0
257.00	2220.0
258.00	10499.0
259.00	1514.0
259.90	342.0
261.00	398.0
262.90	172.0
264.00	232.0
265.00	4921.0
266.00	778.0
271.00	332.0
271.90	484.0
273.00	5566.0
274.00	14689.0
275.00	82160.0
276.00	10354.0
277.00	7064.0
277.90	1057.0
278.90	261.0
280.90	193.0
281.90	254.0
282.90	650.0
284.00	621.0
285.00	1172.0
286.00	211.0
289.00	282.0
290.00	223.0
291.90	377.0
293.00	1425.0
294.00	425.0
295.00	607.0
296.00	22792.0
296.90	2627.0
297.90	256.0
300.90	290.0
302.00	394.0
303.00	3074.0
304.00	693.0
308.00	353.0
308.90	180.0
310.00	364.0
313.00	264.0
314.00	1073.0
315.00	2629.0
316.00	1499.0
317.00	274.0
321.00	783.0
322.00	311.0
323.00	8018.0
324.00	1386.0
325.90	153.0
326.90	1272.0
328.00	687.0
331.90	583.0
332.90	561.0
334.00	4721.0
335.00	1011.0
341.00	866.0
342.00	246.0
346.00	1858.0
346.90	332.0
352.00	2441.0
353.00	1627.0
354.00	2581.0
355.00	448.0
358.90	343.0
365.00	12036.0
365.90	1560.0
369.90	274.0
371.00	572.0
372.00	4070.0
373.00	1147.0
383.00	1187.0
384.00	222.0
390.00	714.0
390.90	508.0
392.00	301.0
400.90	230.0
402.00	1679.0
403.00	2618.0
404.00	968.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931

421.00	2630.0
422.00	2358.0
423.00	15417.0
424.00	3161.0
424.90	423.0
439.20	163.0
441.00	45784.0
442.00	296320.0
443.00	56576.0
444.00	5010.0
444.90	401.0

Instrument :
BNA_M
ClientSampleId :
DFTPP931