

Data File : BM023817.D
Acq On : 20 Nov 2019 03:54
Operator : JU
Sample : K5847-13
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFPW4

Quant Time: Nov 20 07:50:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 20 06:55:09 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	427731	20.00	ng/ul	0.00
18) Naphthalene-d8	10.29	136	1553468	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.17	164	843947	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.93	188	1575134	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	1681542	20.00	ng/ul	0.00
86) Perylene-d12	23.31	264	2160198	20.00	ng/ul	0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.05	96	282	0.03	ng/uL	-0.01
5) Phenol-d5	6.72	99	301	0.01	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.85	67	2105	0.09	ng/ul	-0.02
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	8.25	113	486	0.02	ng/ul	0.01
19) Nitrobenzene-d5	8.61	128	268	0.02	ng/ul	-0.06
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	9.94	165	1233	0.05	ng/ul	0.01
29) 4-Chloroaniline-d4	10.46	131	552	0.02	ng/ul	0.02
44) Dimethylphthalate-d6	13.60	166	582	0.01	ng/ul	0.00
47) Acenaphthylene-d8	13.87	160	646	0.01	ng/ul	0.00
52) 4-Nitrophenol-d4	14.39	143	750	0.07	ng/ul	-0.01
58) Fluorene-d10	15.17	176	662	0.01	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.29	200	2050	0.22	ng/ul	-0.03
71) Anthracene-d10	16.93	188	1575464	21.08	ng/ul	-0.10
79) Pyrene-d10	19.34	212	6331	0.07	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.17	264	2681	0.02	ng/ul	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM023817.D
Acq On : 20 Nov 2019 03:54
Operator : JU
Sample : K5847-13
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFPW4

Quant Time: Nov 20 07:50:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 20 06:55:09 2019
Response via : Initial Calibration

