

Data File : BM023818.D
 Acq On : 20 Nov 2019 04:30
 Operator : JU
 Sample : K5847-14
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BFPW5

Quant Time: Nov 20 07:50:28 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 20 06:55:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	437260	20.00	ng/ul	0.00
18) Naphthalene-d8	10.29	136	1607597	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.17	164	852014	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.93	188	1560388	20.00	ng/ul	0.00
78) Chrysene-d12	21.14	240	1650046	20.00	ng/ul	0.00
86) Perylene-d12	23.30	264	2145269	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.06	96	262	0.02	ng/uL	0.00
5) Phenol-d5	6.75	99	411	0.01	ng/ul	0.04
7) Bis-(2-Chloroethyl)ether-d	6.85	67	1019	0.04	ng/ul	-0.02
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	8.25	113	239	0.01	ng/ul	0.01
19) Nitrobenzene-d5	8.67	128	116	0.01	ng/ul	0.00
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	9.94	165	959	0.03	ng/ul	0.01
29) 4-Chloroaniline-d4	10.46	131	132	0.00	ng/ul	0.02
44) Dimethylphthalate-d6	13.60	166	274	0.00	ng/ul	0.00
47) Acenaphthylene-d8	13.86	160	142	0.00	ng/ul	0.00
52) 4-Nitrophenol-d4	14.39	143	910	0.08	ng/ul	-0.01
58) Fluorene-d10	15.16	176	295	0.01	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.32	200	234	0.03	ng/ul	0.00
71) Anthracene-d10	16.93	188	1560423	21.08	ng/ul	-0.10
79) Pyrene-d10	19.34	212	5458	0.06	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.07	264	4831	0.04	ng/ul	-0.09

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM023818.D
Acq On : 20 Nov 2019 04:30
Operator : JU
Sample : K5847-14
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFPW5

Quant Time: Nov 20 07:50:28 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 20 06:55:09 2019
Response via : Initial Calibration

