

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111919\
 Data File : BM023793.D
 Acq On : 19 Nov 2019 12:49
 Operator : JU
 Sample : PB124677BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK77

Quant Time: Nov 19 13:38:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 19 12:27:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	336428	20.00	ng/ul	0.00
18) Naphthalene-d8	10.29	136	1378937	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.17	164	839974	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.93	188	1811788	20.00	ng/ul	0.00
78) Chrysene-d12	21.13	240	1796925	20.00	ng/ul	0.00
86) Perylene-d12	23.29	264	2195664	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.06	96	56801	6.96	ng/uL	0.00
5) Phenol-d5	6.71	99	960212	31.31	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.87	67	615604	34.93	ng/ul	0.00
9) 2-Chlorophenol-d4	7.06	132	753616	33.92	ng/ul	0.00
13) 4-Methylphenol-d8	8.23	113	751195	30.71	ng/ul	0.00
19) Nitrobenzene-d5	8.67	128	362704	36.75	ng/ul	0.00
22) 2-Nitrophenol-d4	9.39	143	394977	38.07	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.92	165	738717	31.23	ng/ul	0.00
29) 4-Chloroaniline-d4	10.44	131	1026403	39.51	ng/ul	0.00
44) Dimethylphthalate-d6	13.60	166	2328043	35.54	ng/ul	0.00
47) Acenaphthylene-d8	13.87	160	2620763	35.06	ng/ul	0.00
52) 4-Nitrophenol-d4	14.40	143	348173	31.14	ng/ul	0.00
58) Fluorene-d10	15.18	176	1998530	34.55	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.31	200	305120	29.11	ng/ul	0.00
71) Anthracene-d10	17.03	188	3128412	36.40	ng/ul	0.00
79) Pyrene-d10	19.33	212	3496524	37.32	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.15	264	4228756	36.60	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111919\
Data File : BM023793.D
Acq On : 19 Nov 2019 12:49
Operator : JU
Sample : PB124677BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK77

Quant Time: Nov 19 13:38:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 19 12:27:28 2019
Response via : Initial Calibration

