

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM111921\  
 Data File : BM033178.D  
 Acq On : 19 Nov 2021 16:29  
 Operator : CG/JU  
 Sample : PB140818BL  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SBLK818

Quant Time: Nov 20 01:03:47 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM111921.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Nov 19 15:41:12 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| -----                       |        |      |          |       |         |          |
| Internal Standards          |        |      |          |       |         |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.963  | 152  | 1848     | 0.400 | ng/ul   | 0.00     |
| 4) Naphthalene-d8           | 10.760 | 136  | 5582     | 0.400 | ng/ul # | 0.00     |
| 9) Acenaphthene-d10         | 14.580 | 164  | 3633     | 0.400 | ng/ul   | 0.00     |
| 13) Phenanthrene-d10        | 17.320 | 188  | 7729     | 0.400 | ng/ul   | 0.00     |
| 17) Chrysene-d12            | 21.475 | 240  | 6956     | 0.400 | ng/ul   | 0.00     |
| 23) Perylene-d12            | 23.818 | 264  | 6215     | 0.400 | ng/ul # | 0.00     |
|                             |        |      |          |       |         |          |
| System Monitoring Compounds |        |      |          |       |         |          |
| 3) 1,4-Dioxane-d8           | 3.413  | 96   | 12498    | 7.192 | ng/ul   | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.343 | 152  | 3400     | 0.440 | ng/ul   | 0.00     |
| 18) Fluoranthene-d10        | 19.331 | 212  | 8869     | 0.440 | ng/ul   | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM111921\  
Data File : BM033178.D  
Acq On : 19 Nov 2021 16:29  
Operator : CG/JU  
Sample : PB140818BL  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SBLK818

Quant Time: Nov 20 01:03:47 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM111921.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Nov 19 15:41:12 2021  
Response via : Initial Calibration

