

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112118\  
 Data File : BM017736.D  
 Acq On : 21 Nov 2018 18:10  
 Operator : MJ/SJ  
 Sample : J6023-03  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 111918-JETT-E-D-S

Manual Integrations  
 APPROVED

mohammad  
 11/23/2018 11:51:20 AM

Quant Time: Nov 23 03:15:57 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM111218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Nov 12 18:37:25 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.59  | 152  | 186278   | 20.00 | ng    | -0.02     |
| 21) Naphthalene-d8          | 10.36 | 136  | 797931m  | 20.00 | ng    | -0.02     |
| 39) Acenaphthene-d10        | 14.24 | 164  | 492893   | 20.00 | ng    | -0.02     |
| 64) Phenanthrene-d10        | 17.00 | 188  | 1158140  | 20.00 | ng    | -0.02     |
| 76) Chrysene-d12            | 21.20 | 240  | 1249493  | 20.00 | ng    | -0.02     |
| 87) Perylene-d12            | 23.39 | 264  | 1145634  | 20.00 | ng    | -0.02     |
| System Monitoring Compounds |       |      |          |       |       |           |
| 5) 2-Fluorophenol           | 5.22  | 112  | 400608   | 36.29 | ng    | -0.02     |
| 7) Phenol-d6                | 6.78  | 99   | 319147   | 20.70 | ng    | -0.02     |
| 23) Nitrobenzene-d5         | 8.75  | 82   | 1360989  | 88.05 | ng    | -0.02     |
| 42) 2,4,6-Tribromophenol    | 15.74 | 330  | 648409   | 91.56 | ng    | -0.02     |
| 45) 2-Fluorobiphenyl        | 12.86 | 172  | 3414065  | 89.86 | ng    | -0.02     |
| 79) Terphenyl-d14           | 19.64 | 244  | 4833852  | 87.68 | ng    | -0.01     |
| Target Compounds            |       |      |          |       |       |           |
| 10) Phenol                  | 6.81  | 94   | 61040    | 3.772 | ng    | Qvalue 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112118\  
Data File : BM017736.D  
Acq On : 21 Nov 2018 18:10  
Operator : MJ/SJ  
Sample : J6023-03  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
111918-JETT-E-D-S

Manual Integrations  
APPROVED

mohammad  
11/23/2018 11:51:20 AM

Quant Time: Nov 23 03:15:57 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM11218.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Nov 12 18:37:25 2018  
Response via : Initial Calibration

