


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112123\
Data File : BM042935.D
Acq On : 21 Nov 2023 15:41
Operator : MA/JU
Sample : DFTPP181
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP181

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Nov 22 10:35:26 2023



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	71.9	54701	PASS
68	69	0.00	2	1.8	772	PASS
69	198	0.00	100	56.2	42773	PASS
70	69	0.00	2	0.9	395	PASS
127	198	10	80	47.5	36195	PASS
197	198	0.00	2	0.7	526	PASS
198	198	100	100	100.0	76128	PASS
199	198	5	9	7.1	5426	PASS
275	198	10	60	27.9	21242	PASS
365	198	1	100	4.6	3536	PASS
441	198	0.01	100	15.1	11485	PASS
442	198	50	100	91.2	69437	PASS
443	442	15	24	18.3	12733	PASS

m/z	Abundance
36.00	707.0
37.00	1598.0
38.10	169.0
39.10	183.0
40.00	438.0
41.00	587.0
42.00	352.0
43.00	239.0
44.00	1396.0
47.00	4025.0
48.00	1936.0
49.00	16364.0
50.00	874.0
51.10	4704.0
73.10	208.0
82.00	180.0
82.90	235.0
84.00	8493.0
84.90	322.0
86.00	5479.0
88.00	874.0
193.00	160.0
207.00	580.0
208.00	188.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
36.00	860.0
37.00	1982.0
38.00	1337.0
39.10	5911.0
40.10	703.0
41.00	778.0
41.90	403.0
42.90	351.0
44.00	1607.0
45.00	300.0
47.00	4980.0
48.00	2120.0
49.00	20640.0
50.10	21688.0
51.10	74784.0
52.10	3741.0
53.10	251.0
55.00	597.0
56.00	1974.0
57.00	4127.0
58.00	243.0
60.10	191.0
61.00	844.0
62.10	905.0
63.00	1941.0
64.10	357.0
65.00	977.0
68.00	964.0
69.00	53176.0
70.00	432.0
73.00	925.0
74.10	7632.0
75.00	8844.0
76.10	3478.0
77.10	67024.0
78.10	4891.0
79.00	4281.0
80.00	2842.0
81.00	5028.0
82.00	1276.0
83.00	1474.0
84.00	10161.0
84.90	615.0
86.00	7687.0
86.90	575.0
88.00	1226.0
91.00	1055.0
92.00	1251.0
93.00	8023.0
94.00	637.0
96.10	518.0
98.00	6632.0
99.00	4257.0
99.90	338.0
101.00	2068.0
102.00	181.0
103.00	794.0
104.00	1522.0
105.00	1475.0
105.90	375.0
107.00	16912.0
108.00	2147.0
109.10	302.0
110.00	22856.0
111.00	4308.0
112.00	627.0
112.90	207.0
116.00	861.0
117.00	22864.0
118.00	1690.0
122.00	1394.0
123.00	1474.0
124.00	661.0
125.00	741.0
126.10	175.0
127.00	44368.0
128.00	3547.0
129.00	20216.0
130.00	1659.0
131.00	307.0
132.00	217.0
133.00	210.0
133.90	523.0
135.00	1495.0
136.10	644.0
137.00	1082.0
139.10	188.0
140.00	259.0
141.00	2461.0
142.10	1004.0
143.00	823.0
146.00	533.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

147.00	1573.0
148.00	5120.0
149.00	884.0
150.00	178.0
151.00	294.0
151.90	197.0
153.00	832.0
154.00	732.0
155.00	1687.0
156.00	2170.0
157.00	404.0
157.90	385.0
159.00	439.0
160.00	977.0
161.00	1391.0
161.90	309.0
165.00	1023.0
166.10	787.0
167.00	8167.0
168.00	4666.0
169.00	775.0
170.00	256.0
170.80	226.0
172.00	610.0
173.00	622.0
173.90	1076.0
175.00	2424.0
176.00	486.0
177.00	1223.0
178.00	317.0
179.00	4954.0
180.00	3368.0
180.90	935.0
182.00	229.0
183.90	374.0
185.00	2310.0
186.00	13730.0
187.00	4423.0
188.00	415.0
189.00	1116.0
190.90	606.0
192.00	1389.0
193.00	1854.0
194.10	388.0
195.00	151.0
196.00	2320.0
197.10	653.0
198.00	84712.0
199.00	5978.0
200.00	434.0
201.30	218.0
201.90	191.0
203.00	951.0
204.00	4670.0
205.10	7278.0
206.10	25776.0
207.00	4094.0
208.00	1177.0
209.00	554.0
210.10	457.0
211.00	1422.0
211.80	343.0
215.00	366.0
216.00	764.0
217.00	9034.0
218.00	1052.0
219.00	176.0
220.10	187.0
221.00	3324.0
222.00	1146.0
223.00	1927.0
224.10	15591.0
225.10	3944.0
226.00	564.0
227.00	8107.0
228.00	1335.0
229.00	1614.0
229.90	202.0
231.00	566.0
234.00	509.0
234.90	483.0
236.00	537.0
237.00	680.0
238.90	297.0
239.90	289.0
240.90	465.0
242.00	1000.0
243.00	1042.0
244.10	12642.0
245.10	1796.0
246.00	2971.0
246.90	556.0
249.00	387.0
251.90	153.0
253.00	344.0
253.90	367.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

255.00	58056.0
256.00	8418.0
256.90	614.0
258.00	4028.0
258.90	482.0
265.00	1628.0
266.00	330.0
267.00	164.0
270.90	177.0
272.00	224.0
273.00	1573.0
274.00	4569.0
275.00	21368.0
276.00	2982.0
277.00	2198.0
277.90	636.0
281.00	172.0
282.10	153.0
283.00	346.0
284.10	239.0
285.00	455.0
292.10	169.0
293.00	479.0
294.00	190.0
296.00	8367.0
297.00	1104.0
303.10	972.0
303.90	225.0
314.10	201.0
315.00	895.0
316.00	515.0
321.00	184.0
323.00	2442.0
324.00	285.0
327.00	607.0
327.90	233.0
332.10	175.0
333.10	199.0
334.00	1227.0
334.90	230.0
341.10	267.0
346.00	507.0
352.10	832.0
353.10	561.0
354.00	606.0
365.00	3185.0
365.90	427.0
370.90	178.0
372.10	1154.0
372.90	206.0
383.00	358.0
402.00	397.0
403.10	446.0
404.00	275.0
421.00	509.0
422.00	436.0
423.10	2919.0
424.00	872.0
425.20	197.0
441.10	9437.0
442.10	56752.0
443.10	10532.0
444.10	997.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
36.00	774.0
37.00	1904.0
38.00	1211.0
39.10	5409.0
40.00	1011.0
41.00	681.0
42.00	404.0
43.00	416.0
44.00	1555.0
45.00	326.0
47.00	5072.0
48.00	2332.0
49.00	19536.0
50.10	19776.0
51.10	69392.0
52.10	3344.0
53.10	232.0
55.00	549.0
56.10	1663.0
57.00	3643.0
61.00	680.0
62.10	972.0
63.00	2130.0
64.00	358.0
65.00	793.0
67.00	159.0
68.00	889.0
69.00	51480.0
70.00	487.0
71.90	253.0
73.00	809.0
74.10	6857.0
75.00	8709.0
76.10	3525.0
77.10	63048.0
78.10	4230.0
79.00	4248.0
80.00	2983.0
81.00	4238.0
82.00	1057.0
83.00	1193.0
84.00	11014.0
85.00	1049.0
86.00	8257.0
87.00	616.0
88.00	1244.0
91.00	1126.0
92.00	936.0
93.00	8287.0
94.00	528.0
96.00	263.0
98.00	6148.0
99.00	4375.0
99.90	434.0
101.00	2150.0
103.00	645.0
104.00	1443.0
105.00	1484.0
106.10	368.0
107.00	16504.0
108.00	2318.0
109.00	380.0
110.00	22824.0
111.00	3885.0
112.00	616.0
113.10	310.0
116.00	815.0
117.00	22328.0
118.00	1673.0
118.90	154.0
119.90	160.0
122.00	1160.0
123.00	1861.0
124.00	884.0
124.90	574.0
126.20	323.0
127.00	42472.0
128.10	3563.0
129.00	20272.0
130.00	1806.0
131.10	284.0
133.90	620.0
135.00	1785.0
136.00	747.0
137.00	939.0
137.90	188.0
140.00	384.0
141.00	2791.0
142.00	880.0
142.90	596.0
146.00	486.0
147.00	1830.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

148.00	5673.0
149.00	865.0
150.00	247.0
150.90	302.0
152.10	196.0
153.00	1124.0
153.90	758.0
155.00	1662.0
156.00	1967.0
157.10	387.0
158.00	490.0
158.90	453.0
160.00	1207.0
161.00	1261.0
162.00	357.0
163.00	156.0
165.00	1118.0
165.90	810.0
167.00	8462.0
168.00	4811.0
169.00	679.0
171.00	324.0
171.90	780.0
173.00	843.0
174.00	1107.0
175.00	2596.0
176.00	473.0
177.10	1201.0
178.00	448.0
179.00	5138.0
180.00	2956.0
181.00	1543.0
181.90	218.0
183.90	391.0
185.00	2368.0
186.00	15733.0
187.00	3947.0
188.00	429.0
189.00	1288.0
190.00	192.0
191.00	492.0
192.00	1542.0
193.00	2154.0
194.00	466.0
195.00	219.0
196.00	3020.0
197.00	624.0
198.00	92984.0
199.00	6876.0
200.00	617.0
200.90	164.0
201.60	333.0
203.00	971.0
204.00	4636.0
205.00	8079.0
206.10	31048.0
207.10	4018.0
208.10	1264.0
209.00	590.0
210.10	637.0
211.00	1267.0
212.00	171.0
215.10	484.0
216.00	981.0
217.00	9285.0
218.00	1236.0
219.00	215.0
220.10	251.0
221.00	3573.0
222.00	1286.0
223.00	1938.0
224.10	18616.0
225.00	3805.0
226.00	428.0
227.00	9146.0
228.00	1384.0
229.00	1912.0
230.00	214.0
230.90	484.0
234.10	492.0
235.00	484.0
236.00	460.0
237.00	711.0
239.00	386.0
240.00	249.0
241.00	568.0
242.00	1273.0
243.00	1174.0
244.10	13397.0
245.00	1933.0
246.00	3447.0
246.90	635.0
249.10	308.0
251.00	159.0
252.10	157.0
253.00	249.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

254.00	433.0
255.00	68592.0
256.00	10057.0
257.00	685.0
258.00	4623.0
259.00	745.0
260.00	195.0
265.00	2330.0
265.90	269.0
271.00	310.0
272.00	371.0
273.00	2191.0
274.00	5833.0
275.00	27048.0
276.00	3205.0
277.00	3084.0
278.00	747.0
283.00	333.0
284.10	214.0
285.10	522.0
293.00	584.0
293.90	249.0
295.00	219.0
296.00	10246.0
297.00	1361.0
298.10	177.0
302.20	191.0
303.00	1241.0
303.90	239.0
307.90	230.0
314.00	517.0
315.00	1393.0
316.00	577.0
321.00	364.0
321.90	168.0
323.10	2834.0
324.00	534.0
327.00	698.0
328.00	330.0
332.00	326.0
333.10	301.0
334.00	2016.0
335.00	522.0
341.00	443.0
346.10	770.0
352.00	800.0
353.10	604.0
354.10	993.0
355.10	168.0
365.00	4135.0
366.00	674.0
370.80	151.0
372.10	1345.0
373.10	330.0
383.00	380.0
390.10	252.0
402.00	584.0
403.00	760.0
404.10	263.0
421.00	675.0
422.10	848.0
423.10	4353.0
424.10	1065.0
441.10	13887.0
442.10	86784.0
443.10	15962.0
444.10	1553.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

m/z	Abundance
36.00	750.0
37.00	1954.0
38.00	643.0
39.10	2332.0
40.00	568.0
41.00	661.0
42.00	384.0
43.00	394.0
44.00	1506.0
45.00	183.0
47.00	4591.0
48.00	2161.0
49.00	18104.0
50.10	9585.0
51.10	34040.0
52.00	1618.0
55.00	372.0
56.00	1020.0
57.00	1778.0
61.00	364.0
62.00	421.0
63.00	1346.0
64.00	278.0
64.90	428.0
68.00	462.0
69.00	23664.0
70.00	265.0
73.00	755.0
74.10	3263.0
75.10	4380.0
76.00	1695.0
77.10	30616.0
78.10	2038.0
79.00	1872.0
80.00	1534.0
81.00	2114.0
81.90	501.0
83.00	654.0
84.00	9417.0
85.00	604.0
86.00	6510.0
86.90	396.0
88.00	999.0
90.90	431.0
91.90	433.0
93.00	3807.0
94.00	328.0
96.00	298.0
98.10	3159.0
99.00	2214.0
101.00	1010.0
103.00	320.0
103.90	721.0
104.90	580.0
107.00	8513.0
108.00	1486.0
109.10	205.0
110.00	11945.0
111.00	1928.0
112.00	332.0
113.00	202.0
116.00	496.0
117.00	11225.0
118.00	855.0
119.90	152.0
122.00	539.0
123.00	1009.0
123.90	416.0
124.90	281.0
127.00	21744.0
128.00	1445.0
129.00	9574.0
130.00	965.0
131.00	252.0
132.90	174.0
134.00	306.0
135.00	937.0
136.00	311.0
137.00	535.0
137.90	207.0
140.00	178.0
141.00	1443.0
141.90	501.0
143.00	307.0
145.20	179.0
146.00	341.0
147.00	1249.0
148.00	2650.0
149.00	591.0
149.90	174.0
151.00	170.0
153.00	469.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

154.00	316.0
155.00	972.0
156.10	1287.0
157.10	269.0
157.90	262.0
158.90	235.0
160.00	513.0
161.00	791.0
162.00	213.0
164.90	536.0
166.00	469.0
167.00	4818.0
168.00	2514.0
168.90	308.0
171.00	151.0
172.00	264.0
172.90	416.0
174.00	700.0
175.10	1427.0
175.90	362.0
176.90	603.0
178.10	164.0
179.00	2562.0
180.00	1538.0
180.90	774.0
181.90	187.0
184.00	210.0
185.00	1188.0
186.00	8403.0
187.00	2214.0
188.00	284.0
189.00	719.0
191.00	269.0
191.90	818.0
193.10	1344.0
194.00	154.0
196.10	1449.0
197.00	302.0
198.00	50688.0
199.00	3423.0
199.90	409.0
201.30	214.0
203.00	431.0
204.00	2473.0
205.10	4385.0
206.10	15711.0
207.10	3027.0
208.00	913.0
209.10	441.0
210.00	350.0
211.10	733.0
215.00	331.0
216.00	553.0
217.00	5120.0
218.00	819.0
221.00	1887.0
221.90	618.0
223.00	1204.0
224.10	10541.0
225.10	2628.0
226.00	273.0
227.00	5332.0
228.00	695.0
229.00	904.0
230.90	280.0
233.90	360.0
235.00	470.0
236.00	217.0
237.10	417.0
239.10	282.0
240.00	161.0
241.00	322.0
242.00	678.0
243.00	589.0
244.00	7797.0
245.00	1261.0
246.00	2145.0
246.90	365.0
249.00	267.0
253.00	257.0
254.00	189.0
255.00	40280.0
256.00	5648.0
257.00	396.0
258.00	3170.0
259.00	475.0
265.00	1348.0
273.00	1552.0
274.00	3119.0
275.00	15311.0
276.00	2067.0
277.00	1733.0
278.00	380.0
281.00	165.0
283.00	244.0
284.00	186.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181

285.00	340.0
293.10	483.0
296.00	6289.0
297.00	946.0
303.00	749.0
304.20	281.0
314.00	268.0
315.00	750.0
315.90	323.0
323.00	1773.0
324.10	220.0
326.90	355.0
328.00	272.0
333.00	270.0
334.00	1439.0
335.00	330.0
341.10	294.0
345.90	428.0
352.10	621.0
353.00	470.0
354.00	511.0
365.00	3289.0
366.00	624.0
372.00	969.0
373.00	268.0
383.10	301.0
390.00	164.0
402.00	384.0
403.10	566.0
421.00	581.0
422.10	548.0
423.10	3040.0
424.10	826.0
441.10	11132.0
442.10	64776.0
443.10	11706.0
444.10	1246.0

Instrument :
BNA_M
ClientSampleId :
DFTPP181