


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112318\
Data File : BM017752.D
Acq On : 23 Nov 2018 09:30
Operator : MJ/SJ
Sample : DFTPP66
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP66

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
Title : SVOA CALIBRATION
Last Update : Fri Nov 23 21:57:24 2018



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.4	188949	PASS
68	69	0.00	2	1.7	3800	PASS
69	198	0.00	100	40.9	218325	PASS
70	69	0.00	2	0.5	1145	PASS
127	198	10	80	46.1	246229	PASS
197	198	0.00	2	0.6	3091	PASS
198	198	100	100	100.0	534080	PASS
199	198	5	9	6.7	35738	PASS
275	198	10	60	29.1	155648	PASS
365	198	1	100	4.4	23485	PASS
441	198	0.01	100	11.9	63640	PASS
442	198	50	100	78.4	418602	PASS
443	442	15	24	19.3	80824	PASS

m/z	Abundance
37.00	1343.0
38.10	2624.0
39.10	15514.0
40.00	1232.0
41.10	581.0
44.00	3162.0
44.90	492.0
49.10	1862.0
50.00	55656.0
51.10	203776.0
52.10	9951.0
53.00	391.0
55.00	1188.0
56.00	5997.0
57.00	13223.0
57.90	675.0
61.00	2578.0
62.00	2595.0
63.00	8633.0
63.90	995.0
65.00	4172.0
66.20	379.0
68.00	3922.0
69.00	230656.0
70.00	1060.0
73.10	2530.0
74.00	24208.0
75.00	35808.0
76.10	11038.0
77.10	239296.0
78.10	15706.0
79.00	18472.0
80.00	14912.0
81.00	19808.0
82.00	4523.0
83.00	3654.0
83.80	469.0
85.00	3299.0
86.00	5923.0
87.00	2173.0
88.00	857.0
91.00	4262.0
92.00	5184.0
93.00	32288.0
93.90	1933.0
95.00	526.0
96.00	1179.0
97.00	468.0
98.00	24776.0
99.00	19880.0
100.00	1784.0
101.00	10411.0
102.10	373.0
103.00	3454.0
104.00	7298.0
105.00	6422.0
106.10	1722.0
107.00	77600.0
108.00	11101.0
109.00	2694.0
110.00	137536.0
111.00	21856.0
112.00	2570.0
113.00	1128.0
114.90	457.0
116.00	4036.0
117.00	73112.0
118.00	5637.0
119.00	384.0
120.10	1082.0
121.00	302.0
122.00	5220.0
123.00	7679.0
124.00	3346.0
125.00	3859.0
126.20	1441.0
127.00	238080.0
128.00	18472.0
129.00	99504.0
130.00	8940.0
130.90	1827.0
132.00	932.0
133.10	629.0
134.00	3652.0
135.00	8611.0
136.00	3623.0
137.10	3519.0
138.00	944.0
138.80	515.0
140.00	1532.0
141.00	13639.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

141.90	4351.0
142.90	2888.0
144.10	922.0
145.00	954.0
146.00	2305.0
147.00	6584.0
148.00	16760.0
149.00	3214.0
149.90	1164.0
151.10	2335.0
152.10	713.0
152.90	4189.0
153.90	4016.0
155.00	7372.0
156.00	11354.0
157.10	2313.0
158.00	2701.0
159.00	2450.0
160.00	4752.0
161.00	6823.0
161.90	1469.0
163.00	352.0
163.90	787.0
165.00	4845.0
166.00	3978.0
167.00	27968.0
168.00	15811.0
169.00	2129.0
170.10	1010.0
171.00	1318.0
172.00	2278.0
173.00	3233.0
174.00	4858.0
175.10	10463.0
176.00	2151.0
177.00	4848.0
178.10	1911.0
179.00	20872.0
180.00	15666.0
181.00	5945.0
182.00	965.0
182.80	1059.0
184.00	1507.0
185.10	9215.0
186.00	72112.0
187.00	20704.0
188.10	2351.0
189.00	5023.0
190.00	806.0
191.00	2379.0
192.00	6292.0
193.00	7345.0
194.00	1790.0
195.10	703.0
196.00	12805.0
197.10	2866.0
198.00	497984.0
199.00	34328.0
199.90	2694.0
201.40	1781.0
203.00	4194.0
204.00	20312.0
205.00	32568.0
206.00	136256.0
207.00	19408.0
208.00	5029.0
209.00	2218.0
210.20	2220.0
211.00	5655.0
211.90	1080.0
214.90	1643.0
216.00	3825.0
217.00	39400.0
218.00	5320.0
219.00	680.0
221.10	19168.0
221.90	5037.0
223.00	7577.0
224.10	74760.0
225.00	19376.0
226.10	2481.0
227.00	33264.0
228.00	5830.0
229.00	7246.0
229.90	891.0
231.00	3282.0
232.00	631.0
233.00	460.0
234.00	1795.0
235.00	2544.0
236.00	1452.0
237.00	3401.0
238.00	832.0
238.90	1244.0
240.00	1144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

240.90	2192.0
242.00	4168.0
243.00	3878.0
244.10	53288.0
245.10	8513.0
246.00	13319.0
247.00	2447.0
247.90	626.0
249.00	2108.0
251.00	484.0
252.00	556.0
253.00	1458.0
254.00	2314.0
255.00	299456.0
256.00	46648.0
257.00	4458.0
258.00	20248.0
259.00	2935.0
259.90	418.0
260.90	533.0
264.00	647.0
265.00	8306.0
266.00	1216.0
270.00	439.0
271.00	915.0
272.00	945.0
273.00	8165.0
274.00	22544.0
275.00	132480.0
276.00	17456.0
277.00	12686.0
278.00	2638.0
281.00	619.0
282.20	404.0
283.00	1473.0
283.90	850.0
285.00	1840.0
286.00	729.0
288.90	557.0
292.00	501.0
293.00	2907.0
294.00	1114.0
295.00	992.0
296.00	39704.0
297.10	5709.0
298.00	512.0
302.10	827.0
303.10	4176.0
304.10	1274.0
308.00	332.0
310.00	391.0
314.10	2232.0
315.00	5556.0
316.00	2320.0
316.90	594.0
321.10	978.0
321.80	494.0
323.00	12460.0
324.10	2312.0
326.90	2576.0
328.00	1290.0
332.10	934.0
333.00	1091.0
334.00	6948.0
335.10	2445.0
336.00	442.0
341.10	1442.0
342.00	581.0
346.00	2632.0
347.00	735.0
352.10	3116.0
353.00	2895.0
354.10	3701.0
355.10	1000.0
359.00	319.0
365.00	18632.0
366.00	2078.0
370.90	947.0
372.00	5639.0
373.10	1199.0
383.00	1345.0
383.80	597.0
390.10	893.0
391.10	689.0
391.90	440.0
401.00	452.0
402.00	3104.0
403.10	3670.0
404.00	1478.0
405.00	449.0
421.00	2639.0
422.10	2717.0
423.10	17912.0
424.00	5741.0
425.00	964.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

439.90	362.0
441.10	44176.0
442.10	296512.0
443.10	60304.0
444.00	5709.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
37.10	1006.0
38.10	3145.0
39.10	17072.0
40.00	1005.0
41.00	492.0
44.00	3582.0
45.00	630.0
49.00	1805.0
50.10	62712.0
51.10	225856.0
52.00	11008.0
53.00	553.0
55.00	1067.0
56.00	7227.0
57.00	15367.0
57.90	748.0
60.00	323.0
61.00	2551.0
62.00	3392.0
63.00	9721.0
64.00	1196.0
65.10	4566.0
66.00	318.0
67.20	404.0
68.10	4544.0
69.00	257856.0
70.00	1375.0
73.10	2808.0
74.00	28056.0
75.00	44128.0
76.10	13850.0
77.10	277824.0
78.10	18056.0
79.00	22496.0
80.00	15526.0
81.00	20208.0
82.00	5143.0
83.00	4165.0
84.10	481.0
85.00	3342.0
86.00	6587.0
87.00	3028.0
88.00	828.0
89.00	761.0
91.00	4678.0
92.00	6174.0
93.00	38624.0
94.00	3100.0
95.10	650.0
96.10	1267.0
98.00	30912.0
99.00	21304.0
100.00	2087.0
101.00	12202.0
102.10	731.0
103.00	4173.0
104.00	8263.0
105.00	7567.0
106.00	2103.0
107.00	89256.0
108.00	13630.0
109.10	3289.0
110.00	154752.0
111.00	23448.0
112.00	2966.0
113.00	1062.0
114.90	433.0
116.00	5444.0
117.00	89472.0
118.00	6695.0
118.90	525.0
120.00	1259.0
121.00	549.0
122.00	6864.0
123.00	9826.0
124.00	5168.0
125.00	3820.0
127.00	295424.0
128.00	22904.0
129.00	124360.0
130.00	10158.0
131.00	1974.0
132.00	1073.0
132.80	372.0
134.00	3761.0
135.00	9571.0
136.10	4071.0
137.00	5358.0
138.00	1319.0
138.90	687.0
140.00	1877.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

141.00	16896.0
142.00	5556.0
143.00	2916.0
144.10	707.0
145.00	889.0
146.00	3253.0
147.00	9076.0
148.00	20632.0
149.00	3681.0
150.10	1022.0
151.00	1678.0
151.90	1382.0
153.00	5999.0
154.00	3822.0
155.00	8973.0
156.00	12649.0
157.00	2019.0
158.00	3541.0
159.00	2385.0
160.00	5366.0
161.00	7486.0
162.00	2542.0
163.10	789.0
164.00	929.0
165.00	6546.0
166.00	4737.0
167.00	35760.0
168.00	19808.0
169.00	2758.0
170.20	1162.0
170.90	1182.0
172.00	3582.0
173.00	2626.0
174.00	7131.0
175.00	13435.0
176.00	3543.0
177.00	6440.0
178.00	2318.0
179.00	26408.0
180.00	17992.0
181.00	8669.0
182.00	1193.0
182.90	631.0
183.90	2049.0
185.00	12108.0
186.00	94440.0
187.00	28704.0
188.00	2375.0
189.00	6312.0
189.90	1043.0
191.00	3511.0
192.00	8493.0
193.00	10017.0
194.00	2060.0
195.10	980.0
196.00	17424.0
197.10	3912.0
198.00	636992.0
199.00	43680.0
200.00	3728.0
201.40	2177.0
203.00	5069.0
204.00	25648.0
205.10	42128.0
206.10	170240.0
207.00	23112.0
208.00	6525.0
209.00	2655.0
210.20	2808.0
211.10	6166.0
212.10	979.0
212.90	599.0
215.00	2093.0
216.00	4444.0
217.00	52472.0
218.00	5957.0
218.90	572.0
220.30	1150.0
221.00	24416.0
222.00	7281.0
223.00	10234.0
224.10	94944.0
225.00	25488.0
226.10	3154.0
227.00	46688.0
228.00	5852.0
229.00	8456.0
229.90	1232.0
231.00	3389.0
232.00	889.0
232.90	792.0
234.00	3236.0
235.00	3105.0
236.00	2417.0
237.10	4831.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

238.00	744.0
239.00	1921.0
240.00	2119.0
241.00	2640.0
242.00	5221.0
243.10	5731.0
244.10	71416.0
245.10	8904.0
246.00	17616.0
247.00	2957.0
248.10	775.0
249.00	2557.0
250.00	634.0
251.00	695.0
252.00	629.0
253.10	1771.0
254.10	2996.0
255.00	408768.0
256.00	59128.0
257.00	5206.0
258.00	24312.0
259.00	4107.0
260.00	689.0
261.00	1009.0
264.20	533.0
265.00	11228.0
266.00	1298.0
269.90	581.0
271.00	924.0
271.90	844.0
273.00	11153.0
274.00	32536.0
275.00	185920.0
276.00	23368.0
277.00	17024.0
278.00	2551.0
279.10	433.0
281.00	595.0
282.20	305.0
283.00	1669.0
283.90	1228.0
285.00	3062.0
286.00	942.0
289.00	686.0
290.00	734.0
291.90	1410.0
293.10	3645.0
294.00	1180.0
295.00	1116.0
296.00	57224.0
297.00	8338.0
298.00	592.0
301.10	795.0
302.10	967.0
303.00	6260.0
304.10	2138.0
308.00	1016.0
309.00	512.0
310.10	910.0
312.00	320.0
314.10	3273.0
315.00	6515.0
316.00	3346.0
317.00	407.0
321.00	1122.0
322.00	1112.0
323.10	15065.0
324.10	2915.0
326.00	391.0
327.00	3678.0
328.00	1613.0
329.10	667.0
332.00	1432.0
333.00	1472.0
334.00	11213.0
335.00	2864.0
336.00	511.0
338.90	308.0
340.00	314.0
341.10	1786.0
341.90	410.0
346.00	3505.0
346.90	686.0
352.10	4570.0
353.00	4052.0
354.10	5348.0
355.10	959.0
359.10	472.0
365.00	26816.0
366.00	4003.0
369.90	730.0
371.20	1210.0
372.10	9436.0
373.00	2896.0
383.10	2193.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

384.00	860.0
385.20	334.0
390.00	702.0
391.00	483.0
392.00	792.0
401.10	456.0
402.00	3411.0
403.00	5251.0
404.00	2346.0
405.00	491.0
421.10	3745.0
422.00	4538.0
423.10	27696.0
424.10	9586.0
425.10	999.0
439.70	323.0
440.20	650.0
441.10	76664.0
442.10	486528.0
443.10	92656.0
444.10	8526.0
445.00	348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
36.80	635.0
38.00	1761.0
39.10	10309.0
40.00	1064.0
44.00	3467.0
45.00	382.0
49.00	1250.0
50.10	39192.0
51.10	137216.0
52.00	6605.0
53.00	305.0
54.90	903.0
56.00	4201.0
57.00	9817.0
57.90	365.0
60.90	2224.0
62.00	2250.0
63.10	6505.0
64.10	1162.0
65.00	2825.0
68.10	2936.0
69.00	166464.0
70.00	1002.0
73.00	2324.0
74.10	18248.0
75.00	25648.0
76.10	8911.0
77.10	182528.0
78.10	11759.0
79.00	13601.0
80.00	10170.0
81.00	13831.0
82.00	3593.0
83.00	2453.0
84.00	357.0
85.00	2572.0
86.00	3996.0
87.00	1835.0
87.90	571.0
89.00	305.0
91.00	2699.0
92.00	3982.0
93.00	24280.0
94.00	1284.0
95.00	377.0
96.00	1079.0
97.20	342.0
98.00	20072.0
99.00	13406.0
99.90	1239.0
101.00	7276.0
101.90	557.0
103.00	2940.0
103.90	5375.0
105.00	4895.0
106.00	1566.0
107.00	59400.0
108.00	8250.0
109.10	1789.0
110.00	104144.0
111.00	16552.0
112.00	1486.0
113.10	703.0
116.00	3042.0
117.00	59944.0
118.00	3731.0
119.10	450.0
120.00	812.0
122.00	4280.0
123.00	5767.0
124.00	3141.0
125.00	2699.0
125.90	929.0
126.20	1064.0
127.00	205184.0
128.00	14399.0
129.00	83232.0
130.00	7125.0
131.00	1408.0
132.00	836.0
132.90	482.0
134.00	2442.0
135.00	6824.0
136.00	2668.0
137.10	3868.0
137.90	746.0
139.00	591.0
140.00	993.0
141.00	11946.0
142.00	3459.0
142.90	2776.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

144.00	955.0
145.00	764.0
146.00	1870.0
147.00	5942.0
148.00	14375.0
149.10	2990.0
150.00	970.0
151.00	1540.0
151.70	836.0
153.00	4499.0
154.00	3344.0
155.00	6135.0
156.00	9379.0
157.00	1998.0
157.90	2285.0
159.00	1876.0
160.00	3945.0
161.00	5330.0
162.00	1455.0
162.90	631.0
164.00	504.0
165.00	4989.0
166.00	3298.0
167.00	25144.0
168.00	10710.0
169.00	1800.0
169.90	615.0
170.90	1156.0
172.00	2250.0
173.00	2780.0
174.00	4662.0
175.10	9552.0
176.00	2730.0
177.00	5496.0
177.90	1607.0
179.00	19968.0
180.00	14011.0
181.00	5542.0
182.00	985.0
182.80	355.0
183.90	1743.0
185.00	9453.0
186.00	65472.0
187.00	18120.0
188.00	2268.0
189.00	4758.0
190.00	762.0
191.00	2570.0
192.00	7498.0
193.00	8089.0
194.10	1399.0
195.00	626.0
196.00	14313.0
196.70	1983.0
197.10	2496.0
198.00	467264.0
199.00	29208.0
200.00	2508.0
201.60	2440.0
203.00	3724.0
204.00	17848.0
205.00	30976.0
206.00	130736.0
207.10	18144.0
208.10	4374.0
209.00	1904.0
210.30	2150.0
211.00	4360.0
211.70	1022.0
215.00	1593.0
216.10	2898.0
217.00	37768.0
218.00	4426.0
218.90	434.0
220.00	844.0
221.00	16696.0
221.90	4367.0
223.00	7445.0
224.00	72824.0
225.10	18824.0
226.00	2776.0
227.00	34720.0
228.00	5265.0
229.00	6666.0
230.00	1113.0
231.00	3301.0
231.90	646.0
232.90	749.0
234.00	2374.0
235.00	2386.0
235.90	2048.0
237.00	3587.0
238.10	489.0
239.00	1397.0
239.90	998.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

241.00	2386.0
242.00	4328.0
243.00	4394.0
244.10	59560.0
245.00	7915.0
246.00	13331.0
246.90	2726.0
247.90	390.0
248.90	1822.0
250.00	462.0
250.90	584.0
251.90	558.0
252.90	1006.0
253.90	1782.0
255.00	321600.0
256.00	45384.0
257.00	3353.0
258.00	20176.0
259.00	3108.0
260.00	400.0
260.90	731.0
263.90	550.0
265.00	7941.0
266.00	1323.0
267.00	316.0
269.90	492.0
270.90	773.0
272.10	572.0
273.00	8878.0
274.00	25992.0
275.00	148544.0
276.00	18696.0
277.00	14034.0
278.00	3033.0
281.10	503.0
282.00	471.0
283.00	1495.0
284.00	1086.0
285.10	2503.0
286.00	497.0
288.90	792.0
290.00	518.0
290.80	373.0
291.90	614.0
293.00	2726.0
294.00	653.0
295.00	1112.0
296.00	46176.0
297.00	6096.0
298.00	518.0
301.00	764.0
302.00	752.0
303.10	5293.0
304.00	1431.0
305.10	318.0
307.90	515.0
309.00	474.0
310.00	615.0
313.00	496.0
314.10	2261.0
315.00	4350.0
315.90	2504.0
317.10	567.0
321.00	1743.0
322.00	880.0
323.00	14837.0
324.10	2411.0
326.00	448.0
326.90	2364.0
328.00	1474.0
329.00	417.0
332.10	1173.0
333.10	1315.0
334.00	9338.0
335.00	2757.0
336.00	316.0
340.10	380.0
341.10	1699.0
342.10	497.0
345.90	3102.0
346.90	631.0
352.00	4442.0
353.00	3227.0
354.00	5028.0
355.10	1122.0
365.00	25008.0
366.00	3316.0
370.00	535.0
370.90	1173.0
372.10	8419.0
373.00	1834.0
383.00	2393.0
384.00	752.0
390.10	934.0
391.10	1019.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

391.90	465.0
400.90	691.0
402.00	3654.0
403.00	5051.0
404.00	1513.0
405.10	361.0
415.10	337.0
421.00	3696.0
422.10	3862.0
423.00	27584.0
424.10	7221.0
425.10	772.0
440.20	646.0
441.10	70080.0
442.10	472768.0
443.10	89512.0
444.10	7382.0
445.00	414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

Scan 2451 (17.404 min): BM017752.D
DFTPP66

m/z	Abundance
40.00	661.0
44.00	2901.0
73.10	666.0
207.10	1190.0
208.10	306.0
281.00	363.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66