

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112322\
 Data File : BM037682.D
 Acq On : 23 Nov 2022 19:48
 Operator : CG/JU
 Sample : SSTD00590
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040

Quant Time: Nov 24 00:41:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 00:36:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	194818	20.000	ng/u1	0.00
20) Naphthalene-d8	10.939	136	932439	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	622717	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.474	188	1333567	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	1389701	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1402484	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	9150	2.070	ng/uL	0.00
4) Pyridine-d5	3.881	84	55365	4.197	ng/u1	0.01
7) Phenol-d5	7.257	99	80657	4.757	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	57930	5.568	ng/u1	0.00
11) 2-Chlorophenol-d4	7.640	132	58357	4.505	ng/u1	0.00
15) 4-Methylphenol-d8	8.816	113	62342	4.503	ng/u1	0.00
21) Nitrobenzene-d5	9.281	128	31332	4.498	ng/u1	0.00
24) 2-Nitrophenol-d4	10.010	143	29696	3.914	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	56899	4.068	ng/u1	0.00
31) 4-Chloroaniline-d4	11.069	131	99022	4.632	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	198986	4.622	ng/u1	0.00
49) Acenaphthylene-d8	14.433	160	223661	4.321	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	37459	4.353	ng/u1	0.00
60) Fluorene-d10	15.727	176	174941	4.556	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.833	200	21185	2.508	ng/u1	0.00
73) Anthracene-d10	17.574	188	269346	4.386	ng/u1	0.00
81) Pyrene-d10	19.851	212	301173	4.147	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	304333	4.216	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	10214	2.200	ng/uL#	31
5) Pyridine	3.905	79	65957	4.991	ng/u1	94
6) Benzaldehyde	7.245	77	49240	6.278	ng/u1	98
8) Phenol	7.287	94	86007	4.821	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.528	93	74392	5.198	ng/u1	98
12) 2-Chlorophenol	7.675	128	65047	4.641	ng/u1	97
13) 2-Methylphenol	8.551	108	63536	4.668	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.640	45	132423	6.515	ng/u1	98
16) Acetophenone	8.945	105	112689	5.075	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.922	70	65768	5.963	ng/u1	97
18) 4-Methylphenol	8.881	108	70001	4.669	ng/u1	96
19) Hexachloroethane	9.210	117	29127	5.236	ng/u1	98
22) Nitrobenzene	9.328	77	87678	5.215	ng/u1	100
23) Isophorone	9.857	82	178132	5.494	ng/u1	98
25) 2-Nitrophenol	10.045	139	33947	3.915	ng/u1	97
26) 2,4-Dimethylphenol	10.098	107	78078	4.788	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.339	93	106943	5.305	ng/u1	98
29) 2,4-Dichlorophenol	10.575	162	58190	4.042	ng/u1	98
30) Naphthalene	10.986	128	237172	4.767	ng/u1	100
32) 4-Chloroaniline	11.092	127	103455	4.671	ng/u1	99
33) Hexachlorobutadiene	11.275	225	32475	3.663	ng/u1	98
34) Caprolactam	11.857	113	22790	4.811	ng/u1#	76
35) 4-Chloro-3-methylphenol	12.198	107	73904	4.950	ng/u1	96
36) 2-Methylnaphthalene	12.586	142	162956	4.743	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112322\
 Data File : BM037682.D
 Acq On : 23 Nov 2022 19:48
 Operator : CG/JU
 Sample : SSTD00590
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Quant Time: Nov 24 00:41:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 00:36:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	163395	4.726	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.945	216	64769	3.466	ng/ul	97
40) Hexachlorocyclopentadiene	12.927	237	34264	3.279	ng/ul	95
41) 2,4,6-Trichlorophenol	13.180	196	47102	3.904	ng/ul	96
42) 2,4,5-Trichlorophenol	13.245	196	49041	3.775	ng/ul	99
43) 1,1'-Biphenyl	13.580	154	217664	4.703	ng/ul	99
44) 2-Chloronaphthalene	13.627	162	164017	4.410	ng/ul	99
45) 2-Nitroaniline	13.822	65	56864	5.799	ng/ul	97
47) Dimethylphthalate	14.192	163	213113	4.647	ng/ul	98
48) 2,6-Dinitrotoluene	14.310	165	35394	3.941	ng/ul	88
50) Acenaphthylene	14.463	152	256097	4.526	ng/ul	99
51) 3-Nitroaniline	14.639	138	40968	4.228	ng/ul#	95
52) Acenaphthene	14.804	153	188718	4.688	ng/ul	98
53) 2,4-Dinitrophenol	14.839	184	15169	2.516	ng/ul	98
55) 4-Nitrophenol	14.927	109	35687	5.639	ng/ul	98
56) Dibenzofuran	15.133	168	251972	4.635	ng/ul	98
57) 2,4-Dinitrotoluene	15.092	165	52040	3.955	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.357	232	40519	3.583	ng/ul	99
59) Diethylphthalate	15.545	149	233894	5.120	ng/ul	99
61) Fluorene	15.780	166	218791	4.758	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.774	204	92165	4.079	ng/ul	98
63) 4-Nitroaniline	15.792	138	39685	4.405	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.845	198	23345	2.602	ng/ul#	89
67) N-Nitrosodiphenylamine	15.980	169	179979	4.553	ng/ul	98
68) 4-Bromophenyl-phenylether	16.663	248	52465	3.698	ng/ul	98
69) Hexachlorobenzene	16.780	284	61288	3.540	ng/ul	98
70) Atrazine	16.927	200	58935	4.211	ng/ul	99
71) Pentachlorophenol	17.121	266	32894	3.105	ng/ul	96
72) Phenanthrene	17.515	178	351361	4.712	ng/ul	98
74) Anthracene	17.610	178	349179	4.654	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.545	216	65922	3.411	ng/ul	99
76) Pentachlorobenzene	15.057	250	73172	3.480	ng/ul	97
77) Carbazole	17.874	167	332555	4.913	ng/ul	99
78) Di-n-butylphthalate	18.427	149	670176	8.594	ng/ul	99
80) Fluoranthene	19.521	202	387192	4.333	ng/ul	98
82) Pyrene	19.880	202	409370	4.348	ng/ul	95
83) Butylbenzylphthalate	20.762	149	187411	5.126	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.550	252	111746	3.470	ng/ul	95
85) Benzo(a)anthracene	21.621	228	406165	4.554	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.539	149	288036	5.650	ng/ul	99
87) Chrysene	21.680	228	392240	4.674	ng/ul	98
89) Di-n-octyl phthalate	22.492	149	483865	4.742	ng/ul	100
90) Benzo(b)fluoranthene	23.368	252	415853	4.262	ng/ul	100
91) Benzo(k)fluoranthene	23.415	252	405109	4.241	ng/ul	98
93) Benzo(a)pyrene	24.021	252	361328	4.428	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.738	276	412784	4.727	ng/ul	99
95) Dibenzo(a,h)anthracene	26.756	278	366134	4.662	ng/ul	99
96) Benzo(g,h,i)perylene	27.538	276	363162	4.854	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 24 00:41:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 24 00:36:57 2022
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 2.200 ng/uL

RT: 3.475 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :





BNA_M

ClientSampleId :

SSTD005040

Tgt Ion: 88 Resp: 10214

Ion	Ratio	Lower	Upper
88	100		
43	158.9	42.8	64.2#
58	91.0	62.6	94.0



#5

Pyridine

Concen: 4.991 ng/uL

RT: 3.905 min Scan# 88




Delta R.T. 0.012 min


Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Tgt Ion: 79 Resp: 65957

Ion	Ratio	Lower	Upper
79	100		
52	89.7	68.2	102.4
51	45.4	32.4	48.6









#6
Benzaldehyde
Concen: 6.278 ng/ul
RT: 7.245 min Scan# 656
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 77 Resp: 49240
Ion Ratio Lower Upper
77 100
105 92.3 73.4 110.0
106 84.5 70.2 105.2



#8
Phenol
Concen: 4.821 ng/ul
RT: 7.287 min Scan# 663
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion: 94 Resp: 86007
Ion Ratio Lower Upper
94 100
65 30.0 25.0 37.4
66 37.8 31.0 46.6









#10
Bis(2-Chloroethyl)ether
Concen: 5.198 ng/ul
RT: 7.528 min Scan# 704
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 93 Resp: 74392
Ion Ratio Lower Upper
93 100
63 87.6 68.6 103.0
95 32.3 27.1 40.7



#12
2-Chlorophenol
Concen: 4.641 ng/ul
RT: 7.675 min Scan# 729
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:128 Resp: 65047
Ion Ratio Lower Upper
128 100
64 51.1 42.2 63.4
130 35.0 26.1 39.1





#13

2-Methylphenol

Concen: 4.668 ng/ul

RT: 8.551 min Scan# 81

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :





SSTD005040

Tgt Ion:108 Resp: 63536

Ion Ratio Lower Upper

108 100

107 89.9 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 6.515 ng/ul

RT: 8.640 min Scan# 893

Delta R.T. -0.012 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion: 45 Resp: 132423


Ion Ratio Lower Upper

45 100

77 11.5 9.4 14.0

79 7.6 7.3 10.9









#16
Acetophenone
Concen: 5.075 ng/ul
RT: 8.945 min Scan# 945
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:105 Resp: 112689
Ion Ratio Lower Upper
105 100
77 84.4 66.6 99.8
51 35.7 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 5.963 ng/ul
RT: 8.922 min Scan# 941
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion: 70 Resp: 65768
Ion Ratio Lower Upper
70 100
42 74.1 57.1 85.7
101 8.8 7.0 10.6
130 16.9 14.8 22.2









#18
4-Methylphenol
Concen: 4.669 ng/ul
RT: 8.881 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:108 Resp: 70001
Ion Ratio Lower Upper
108 100
107 108.9 90.4 135.6



#19
Hexachloroethane
Concen: 5.236 ng/ul
RT: 9.210 min Scan# 990
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:117 Resp: 29127
Ion Ratio Lower Upper
117 100
201 74.9 60.4 90.6
199 49.6 38.1 57.1









#22
Nitrobenzene
Concen: 5.215 ng/ul
RT: 9.328 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 77 Resp: 87678
Ion Ratio Lower Upper
77 100
123 40.4 32.4 48.6
65 13.7 11.4 17.0



#23
Isophorone
Concen: 5.494 ng/ul
RT: 9.857 min Scan# 1100
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion: 82 Resp: 178132
Ion Ratio Lower Upper
82 100
95 6.3 5.3 7.9
138 14.7 12.5 18.7









#25
2-Nitrophenol
Concen: 3.915 ng/ul
RT: 10.045 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:139 Resp: 33947
Ion Ratio Lower Upper
139 100
65 65.3 49.6 74.4
109 33.1 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 4.788 ng/ul
RT: 10.098 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:107 Resp: 78078
Ion Ratio Lower Upper
107 100
121 48.6 39.6 59.4
122 86.7 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 5.305 ng/ul
 RT: 10.339 min Scan# 1182
 Delta R.T. 0.000 min
 Lab File: BM037682.D
 Acq: 23 Nov 2022 19:48




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion: 93 Resp: 106943
 Ion Ratio Lower Upper
 93 100
 95 31.5 26.4 39.6
 123 10.3 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 4.042 ng/ul
 RT: 10.575 min Scan# 1222
 Delta R.T. -0.006 min
 Lab File: BM037682.D
 Acq: 23 Nov 2022 19:48

Tgt Ion: 162 Resp: 58190
 Ion Ratio Lower Upper
 162 100
 164 67.1 52.6 79.0
 98 41.2 33.6 50.4









#30
Naphthalene
Concen: 4.767 ng/ul
RT: 10.986 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:128 Resp: 237172
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 13.3 10.4 15.6



#32
4-Chloroaniline
Concen: 4.671 ng/ul
RT: 11.092 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:127 Resp: 103455
Ion Ratio Lower Upper
127 100
129 33.2 26.1 39.1









#33
Hexachlorobutadiene
Concen: 3.663 ng/ul
RT: 11.275 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:225 Resp: 32475
Ion Ratio Lower Upper
225 100
223 62.9 51.1 76.7
227 64.9 50.7 76.1



#34
Caprolactam
Concen: 4.811 ng/ul
RT: 11.857 min Scan# 1440
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:113 Resp: 22790
Ion Ratio Lower Upper
113 100
55 240.0 163.3 244.9
56 182.8 120.3 180.5#









#35
4-Chloro-3-methylphenol
Concen: 4.950 ng/ul
RT: 12.198 min Scan# 1499
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:107 Resp: 73904
Ion Ratio Lower Upper
107 100
144 23.6 20.2 30.4
142 81.9 62.6 93.8



#36
2-Methylnaphthalene
Concen: 4.743 ng/ul
RT: 12.586 min Scan# 1564
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:142 Resp: 162956
Ion Ratio Lower Upper
142 100
141 87.4 72.2 108.4









#37
1-Methylnaphthalene
Concen: 4.726 ng/ul
RT: 12.804 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:142 Resp: 163395
Ion Ratio Lower Upper
142 100
141 88.0 73.3 109.9
116 3.9 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 3.466 ng/ul
RT: 12.945 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:216 Resp: 64769
Ion Ratio Lower Upper
216 100
214 78.1 60.2 90.4
179 23.2 18.6 27.8
108 25.1 18.3 27.5









#40
Hexachlorocyclopentadiene
Concen: 3.279 ng/ul
RT: 12.927 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:237 Resp: 34264
Ion Ratio Lower Upper
237 100
235 65.5 49.1 73.7
272 14.1 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 3.904 ng/ul
RT: 13.180 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:196 Resp: 47102
Ion Ratio Lower Upper
196 100
198 91.9 76.9 115.3
200 29.1 24.1 36.1









#42
2,4,5-Trichlorophenol
Concen: 3.775 ng/ul
RT: 13.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:196 Resp: 49041
Ion Ratio Lower Upper
196 100
198 93.3 75.4 113.2
200 29.6 23.7 35.5



#43
1,1'-Biphenyl
Concen: 4.703 ng/ul
RT: 13.580 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:154 Resp: 217664
Ion Ratio Lower Upper
154 100
153 40.6 32.1 48.1
76 14.4 12.4 18.6









#44
2-Chloronaphthalene
Concen: 4.410 ng/ul
RT: 13.627 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:162 Resp: 164017
Ion Ratio Lower Upper
162 100
164 33.5 26.5 39.7
127 41.1 32.4 48.6



#45
2-Nitroaniline
Concen: 5.799 ng/ul
RT: 13.822 min Scan# 1774
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion: 65 Resp: 56864
Ion Ratio Lower Upper
65 100
92 63.1 48.6 73.0
138 86.4 67.0 100.4





#47

Dimethylphthalate

Concen: 4.647 ng/ul

RT: 14.192 min Scan# 1837

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:163 Resp: 213113

Ion Ratio Lower Upper

163 100

194 3.2 2.9 4.3

164 9.3 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 3.941 ng/ul

RT: 14.310 min Scan# 1857

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:165 Resp: 35394


Ion Ratio Lower Upper

165 100

89 80.9 56.4 84.6

121 27.7 18.9 28.3









#50
Acenaphthylene
Concen: 4.526 ng/ul
RT: 14.463 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:152 Resp: 256097
Ion Ratio Lower Upper
152 100
151 20.7 16.3 24.5
153 11.9 10.3 15.5



#51
3-Nitroaniline
Concen: 4.228 ng/ul
RT: 14.639 min Scan# 1913
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:138 Resp: 40968
Ion Ratio Lower Upper
138 100
108 12.9 8.4 12.6#
92 124.5 104.0 156.0





#52

Acenaphthene

Concen: 4.688 ng/ul

RT: 14.804 min Scan# 1941

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:153 Resp: 188718

Ion Ratio Lower Upper

153 100

152 46.8 38.4 57.6

154 90.9 70.7 106.1



#53

2,4-Dinitrophenol

Concen: 2.516 ng/ul

RT: 14.839 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:184 Resp: 15169


Ion Ratio Lower Upper

184 100

63 91.1 74.9 112.3

154 71.6 57.8 86.6









#55
4-Nitrophenol
Concen: 5.639 ng/ul
RT: 14.927 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:109 Resp: 35687
Ion Ratio Lower Upper
109 100
139 124.7 102.4 153.6
65 147.2 119.6 179.4



#56
Dibenzofuran
Concen: 4.635 ng/ul
RT: 15.133 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:168 Resp: 251972
Ion Ratio Lower Upper
168 100
139 41.1 32.2 48.2









#57
2,4-Dinitrotoluene
Concen: 3.955 ng/ul
RT: 15.092 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:165 Resp: 52040
Ion Ratio Lower Upper
165 100
63 56.6 46.6 70.0
182 2.4 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.583 ng/ul
RT: 15.357 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:232 Resp: 40519
Ion Ratio Lower Upper
232 100
131 62.1 50.2 75.4
130 3.6 2.8 4.2
166 37.5 31.2 46.8









#59
Diethylphthalate
Concen: 5.120 ng/ul
RT: 15.545 min Scan# 2067
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:149 Resp: 233894
Ion Ratio Lower Upper
149 100
177 19.6 16.1 24.1
150 11.5 9.4 14.0



#61
Fluorene
Concen: 4.758 ng/ul
RT: 15.780 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:166 Resp: 218791
Ion Ratio Lower Upper
166 100
165 95.7 79.3 118.9
167 12.8 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 4.079 ng/ul
RT: 15.774 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:204 Resp: 92165
Ion Ratio Lower Upper
204 100
206 31.8 25.8 38.8
141 69.5 54.3 81.5



#63
4-Nitroaniline
Concen: 4.405 ng/ul
RT: 15.792 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:138 Resp: 39685
Ion Ratio Lower Upper
138 100
92 58.6 43.7 65.5
108 83.2 68.1 102.1









#66
4,6-Dinitro-2-methylphenol
Concen: 2.602 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:198 Resp: 23345
Ion Ratio Lower Upper
198 100
182 7.0 4.0 6.0#
77 41.5 27.9 41.9



#67
N-Nitrosodiphenylamine
Concen: 4.553 ng/ul
RT: 15.980 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:169 Resp: 179979
Ion Ratio Lower Upper
169 100
168 68.2 54.0 81.0
167 38.4 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 3.698 ng/ul
RT: 16.663 min Scan# 2257
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:248 Resp: 52465
Ion Ratio Lower Upper
248 100
250 99.8 78.5 117.7
141 95.5 78.9 118.3



#69
Hexachlorobenzene
Concen: 3.540 ng/ul
RT: 16.780 min Scan# 2277
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:284 Resp: 61288
Ion Ratio Lower Upper
284 100
142 48.8 40.4 60.6
249 31.2 24.5 36.7





#70

Atrazine

Concen: 4.211 ng/ul

RT: 16.927 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:200 Resp: 58935

Ion Ratio Lower Upper

200 100

173 27.6 22.6 34.0

215 45.1 36.6 55.0



#71

Pentachlorophenol

Concen: 3.105 ng/ul

RT: 17.121 min Scan# 2335

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:266 Resp: 32894


Ion Ratio Lower Upper

266 100

264 61.6 52.0 78.0

268 62.2 52.2 78.2





#72

Phenanthrene

Concen: 4.712 ng/ul

RT: 17.515 min Scan# 2402

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:178 Resp: 351361

Ion Ratio Lower Upper

178 100

179 16.2 12.0 18.0

176 19.9 15.4 23.2



#74

Anthracene

Concen: 4.654 ng/ul

RT: 17.610 min Scan# 2418

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:178 Resp: 349179


Ion Ratio Lower Upper

178 100

179 15.7 12.6 18.8

176 18.9 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 3.411 ng/uL

RT: 13.545 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:216 Resp: 65922

Ion Ratio Lower Upper

216 100

214 79.6 63.3 94.9

218 49.7 38.6 58.0



#76

Pentachlorobenzene

Concen: 3.480 ng/uL

RT: 15.057 min Scan# 1984

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:250 Resp: 73172


Ion Ratio Lower Upper

250 100

248 62.2 50.1 75.1

252 58.6 49.9 74.9









#77
 Carbazole
 Concen: 4.913 ng/ul
 RT: 17.874 min Scan# 2463
 Delta R.T. 0.000 min
 Lab File: BM037682.D
 Acq: 23 Nov 2022 19:48




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion:167 Resp: 332555
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.5 26.3
 139 13.2 10.9 16.3



#78
 Di-n-butylphthalate
 Concen: 8.594 ng/ul
 RT: 18.427 min Scan# 2557
 Delta R.T. -0.006 min
 Lab File: BM037682.D
 Acq: 23 Nov 2022 19:48

Tgt Ion:149 Resp: 670176
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 5.9 4.4 6.6









#80
Fluoranthene
Concen: 4.333 ng/ul
RT: 19.521 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:202 Resp: 387192
Ion Ratio Lower Upper
202 100
101 13.8 10.6 16.0
100 10.9 8.2 12.4



#82
Pyrene
Concen: 4.348 ng/ul
RT: 19.880 min Scan# 2804
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:202 Resp: 409370
Ion Ratio Lower Upper
202 100
101 17.8 12.2 18.4
100 14.4 10.2 15.2









#83
Butylbenzylphthalate
Concen: 5.126 ng/ul
RT: 20.762 min Scan# 2954
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:149 Resp: 187411
Ion Ratio Lower Upper
149 100
91 84.7 65.9 98.9
206 16.1 13.4 20.2



#84
3,3'-Dichlorobenzidine
Concen: 3.470 ng/ul
RT: 21.550 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:252 Resp: 111746
Ion Ratio Lower Upper
252 100
254 68.1 51.0 76.4
126 15.2 12.2 18.4





#85

Benzo(a)anthracene

Concen: 4.554 ng/ul

RT: 21.621 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:228 Resp: 406165

Ion Ratio Lower Upper

228 100

229 19.9 15.6 23.4

226 27.3 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 5.650 ng/ul

RT: 21.539 min Scan# 3086

Delta R.T. 0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:149 Resp: 288036


Ion Ratio Lower Upper

149 100

167 25.9 20.2 30.4

279 3.0 2.2 3.4





#87

Chrysene

Concen: 4.674 ng/ul

RT: 21.680 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:228 Resp: 392240

Ion Ratio Lower Upper

228 100

226 30.4 24.2 36.4

229 18.5 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 4.742 ng/ul




RT: 22.492 min Scan# 3248


Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Tgt Ion:149 Resp: 483865





#90

Benzo(b)fluoranthene

Concen: 4.262 ng/ul

RT: 23.368 min Scan# 31

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:252 Resp: 415853

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.6

125 13.5 11.0 16.4



#91

Benzo(k)fluoranthene

Concen: 4.241 ng/ul

RT: 23.415 min Scan# 3405

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:252 Resp: 405109


Ion Ratio Lower Upper

252 100

253 21.4 17.5 26.3

125 14.4 10.6 16.0





#93

Benzo(a)pyrene

Concen: 4.428 ng/ul

RT: 24.021 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:252 Resp: 361328

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 14.4 12.0 18.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.727 ng/ul

RT: 26.738 min Scan# 3970

Delta R.T. -0.006 min

Lab File: BM037682.D

Acq: 23 Nov 2022 19:48




Tgt Ion:276 Resp: 412784


Ion Ratio Lower Upper

276 100

138 33.2 26.9 40.3

277 25.7 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 4.662 ng/ul
RT: 26.756 min Scan# 3974
Delta R.T. -0.006 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Instrument :
BNA_M
ClientSampleId :
SSTD005040

Tgt Ion:278 Resp: 366134
Ion Ratio Lower Upper
278 100
139 20.6 16.6 25.0
279 24.5 19.3 28.9



#96
Benzo(g,h,i)perylene
Concen: 4.854 ng/ul
RT: 27.538 min Scan# 4106
Delta R.T. -0.012 min
Lab File: BM037682.D
Acq: 23 Nov 2022 19:48

Tgt Ion:276 Resp: 363162
Ion Ratio Lower Upper
276 100
138 28.1 23.4 35.2
277 23.9 19.1 28.7

