

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112322\
 Data File : BM037685.D
 Acq On : 23 Nov 2022 21:39
 Operator : CG/JU
 Sample : SST04003
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040043

Quant Time: Nov 24 00:43:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 00:36:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	159117	20.000	ng/u1	0.00
20) Naphthalene-d8	10.940	136	808661	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.739	164	514377	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.474	188	1053150	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	1037753	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1012654	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	60962	16.889	ng/uL	0.00
4) Pyridine-d5	3.869	84	524746	48.700	ng/u1	0.00
7) Phenol-d5	7.263	99	647853	46.784	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	440205	51.804	ng/u1	0.00
11) 2-Chlorophenol-d4	7.640	132	463452	43.804	ng/u1	0.00
15) 4-Methylphenol-d8	8.822	113	525081	46.435	ng/u1	0.00
21) Nitrobenzene-d5	9.287	128	267213	44.236	ng/u1	0.00
24) 2-Nitrophenol-d4	10.016	143	264699	40.228	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	479490	39.526	ng/u1	0.00
31) 4-Chloroaniline-d4	11.075	131	772005	41.644	ng/u1	0.00
46) Dimethylphthalate-d6	14.145	166	1492823	41.974	ng/u1	0.00
49) Acenaphthylene-d8	14.439	160	1819129	42.545	ng/u1	0.00
54) 4-Nitrophenol-d4	14.922	143	335412	47.182	ng/u1	0.00
60) Fluorene-d10	15.727	176	1310581	41.323	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.839	200	221489	33.205	ng/u1	0.00
73) Anthracene-d10	17.574	188	2053566	42.339	ng/u1	0.00
81) Pyrene-d10	19.857	212	2232516	41.169	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	2199523	42.196	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	67224	17.726	ng/uL	95
5) Pyridine	3.893	79	529302	49.044	ng/u1	98
6) Benzaldehyde	7.246	77	388175	60.593	ng/u1	99
8) Phenol	7.293	94	685043	47.018	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.528	93	563028	48.169	ng/u1	98
12) 2-Chlorophenol	7.675	128	506598	44.251	ng/u1	100
13) 2-Methylphenol	8.557	108	530029	47.683	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.646	45	1025118	61.750	ng/u1	99
16) Acetophenone	8.946	105	883349	48.710	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.934	70	519277	57.648	ng/u1	100
18) 4-Methylphenol	8.887	108	588822	48.088	ng/u1	97
19) Hexachloroethane	9.210	117	217388	47.846	ng/u1	99
22) Nitrobenzene	9.328	77	723989	49.657	ng/u1	99
23) Isophorone	9.857	82	1448956	51.527	ng/u1	99
25) 2-Nitrophenol	10.045	139	296997	39.491	ng/u1	98
26) 2,4-Dimethylphenol	10.098	107	646150	45.694	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.340	93	819966	46.901	ng/u1	99
29) 2,4-Dichlorophenol	10.581	162	489437	39.205	ng/u1	97
30) Naphthalene	10.992	128	1807528	41.895	ng/u1	100
32) 4-Chloroaniline	11.098	127	811399	42.239	ng/u1	99
33) Hexachlorobutadiene	11.275	225	247635	32.210	ng/u1	99
34) Caprolactam	11.869	113	135413	32.962	ng/u1	93
35) 4-Chloro-3-methylphenol	12.204	107	625067	48.271	ng/u1	99
36) 2-Methylnaphthalene	12.586	142	1253908	42.084	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112322\
 Data File : BM037685.D
 Acq On : 23 Nov 2022 21:39
 Operator : CG/JU
 Sample : SST04093
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SST040043


Quant Time: Nov 24 00:43:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 24 00:36:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	1262993	42.126	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.951	216	523559	33.914	ng/ul	97
40) Hexachlorocyclopentadiene	12.928	237	298267	34.559	ng/ul	99
41) 2,4,6-Trichlorophenol	13.181	196	376475	37.777	ng/ul	99
42) 2,4,5-Trichlorophenol	13.251	196	409282	38.141	ng/ul	100
43) 1,1'-Biphenyl	13.581	154	1605152	41.986	ng/ul	99
44) 2-Chloronaphthalene	13.628	162	1268056	41.278	ng/ul	99
45) 2-Nitroaniline	13.828	65	499962	61.722	ng/ul	95
47) Dimethylphthalate	14.198	163	1597415	42.167	ng/ul	99
48) 2,6-Dinitrotoluene	14.316	165	323440	43.604	ng/ul	99
50) Acenaphthylene	14.469	152	2044528	43.748	ng/ul	100
51) 3-Nitroaniline	14.639	138	387814	48.449	ng/ul	99
52) Acenaphthene	14.804	153	1437662	43.233	ng/ul	99
53) 2,4-Dinitrophenol	14.839	184	170285	34.187	ng/ul	98
55) 4-Nitrophenol	14.939	109	298008	57.007	ng/ul	99
56) Dibenzofuran	15.139	168	1863070	41.492	ng/ul	100
57) 2,4-Dinitrotoluene	15.092	165	472854	43.506	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.357	232	333507	35.699	ng/ul	99
59) Diethylphthalate	15.551	149	1710184	45.324	ng/ul	99
61) Fluorene	15.780	166	1618864	42.623	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.775	204	697094	37.347	ng/ul	99
63) 4-Nitroaniline	15.798	138	382281	51.368	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.851	198	239814	33.843	ng/ul	96
67) N-Nitrosodiphenylamine	15.986	169	1361065	43.597	ng/ul	99
68) 4-Bromophenyl-phenylether	16.663	248	340765	30.415	ng/ul	99
69) Hexachlorobenzene	16.780	284	443185	32.412	ng/ul	99
70) Atrazine	16.927	200	450453	40.757	ng/ul	99
71) Pentachlorophenol	17.121	266	280628	33.538	ng/ul	97
72) Phenanthrene	17.521	178	2548620	43.281	ng/ul	100
74) Anthracene	17.610	178	2620944	44.237	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.551	216	529241	34.679	ng/ul	100
76) Pentachlorobenzene	15.057	250	552015	33.243	ng/ul	98
77) Carbazole	17.874	167	2456018	45.949	ng/ul	100
78) Di-n-butylphthalate	18.427	149	3176614	51.585	ng/ul	99
80) Fluoranthene	19.521	202	2871218	43.027	ng/ul	100
82) Pyrene	19.886	202	3061548	43.550	ng/ul	99
83) Butylbenzylphthalate	20.762	149	1461178	53.515	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.551	252	898331	37.351	ng/ul	98
85) Benzo(a)anthracene	21.621	228	2930189	43.997	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.533	149	2182161	57.316	ng/ul	98
87) Chrysene	21.680	228	2738347	43.698	ng/ul	99
89) Di-n-octyl phthalate	22.492	149	3727981	50.599	ng/ul	100
90) Benzo(b)fluoranthene	23.368	252	2853228	40.502	ng/ul	99
91) Benzo(k)fluoranthene	23.421	252	2836121	41.119	ng/ul	99
93) Benzo(a)pyrene	24.021	252	2511930	42.635	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.744	276	2903831	46.053	ng/ul	99
95) Dibenzo(a,h)anthracene	26.768	278	2607147	45.979	ng/ul	99
96) Benzo(g,h,i)perylene	27.556	276	2526145	46.759	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 24 00:43:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 24 00:36:57 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 17.726 ng/uL
RT: 3.475 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion: 88 Resp: 67224
Ion Ratio Lower Upper
88 100
43 46.6 42.8 64.2
58 80.5 62.6 94.0



#5
Pyridine
Concen: 49.044 ng/uL
RT: 3.893 min Scan# 86
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion: 79 Resp: 529302
Ion Ratio Lower Upper
79 100
52 83.3 68.2 102.4
51 41.2 32.4 48.6









#6
Benzaldehyde
Concen: 60.593 ng/ul
RT: 7.246 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion: 77 Resp: 388175
Ion Ratio Lower Upper
77 100
105 91.7 73.4 110.0
106 90.2 70.2 105.2



#8
Phenol
Concen: 47.018 ng/ul
RT: 7.293 min Scan# 664
Delta R.T. 0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion: 94 Resp: 685043
Ion Ratio Lower Upper
94 100
65 31.5 25.0 37.4
66 40.0 31.0 46.6









#10
Bis(2-Chloroethyl)ether
Concen: 48.169 ng/ul
RT: 7.528 min Scan# 704
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion: 93 Resp: 563028
Ion Ratio Lower Upper
93 100
63 84.5 68.6 103.0
95 32.0 27.1 40.7



#12
2-Chlorophenol
Concen: 44.251 ng/ul
RT: 7.675 min Scan# 729
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:128 Resp: 506598
Ion Ratio Lower Upper
128 100
64 53.0 42.2 63.4
130 32.8 26.1 39.1





#13

2-Methylphenol

Concen: 47.683 ng/ul

RT: 8.557 min Scan# 81

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :





SSTD040043

Tgt Ion:108 Resp: 530029

Ion Ratio Lower Upper

108 100

107 87.7 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 61.750 ng/ul

RT: 8.646 min Scan# 894

Delta R.T. -0.006 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion: 45 Resp: 1025118


Ion Ratio Lower Upper

45 100

77 12.1 9.4 14.0

79 8.6 7.3 10.9









#16
Acetophenone
Concen: 48.710 ng/ul
RT: 8.946 min Scan# 945
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:105 Resp: 883349
Ion Ratio Lower Upper
105 100
77 82.7 66.6 99.8
51 34.2 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 57.648 ng/ul
RT: 8.934 min Scan# 943
Delta R.T. 0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion: 70 Resp: 519277
Ion Ratio Lower Upper
70 100
42 71.5 57.1 85.7
101 8.5 7.0 10.6
130 18.4 14.8 22.2









#18
4-Methylphenol
Concen: 48.088 ng/ul
RT: 8.887 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:108 Resp: 588822
Ion Ratio Lower Upper
108 100
107 110.0 90.4 135.6



#19
Hexachloroethane
Concen: 47.846 ng/ul
RT: 9.210 min Scan# 990
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:117 Resp: 217388
Ion Ratio Lower Upper
117 100
201 76.6 60.4 90.6
199 48.1 38.1 57.1









#22
Nitrobenzene
Concen: 49.657 ng/ul
RT: 9.328 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion: 77 Resp: 723989
Ion Ratio Lower Upper
77 100
123 39.8 32.4 48.6
65 14.4 11.4 17.0



#23
Isophorone
Concen: 51.527 ng/ul
RT: 9.857 min Scan# 1100
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion: 82 Resp: 1448956
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 15.9 12.5 18.7









#25
2-Nitrophenol
Concen: 39.491 ng/ul
RT: 10.045 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:139 Resp: 296997
Ion Ratio Lower Upper
139 100
65 64.2 49.6 74.4
109 32.4 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 45.694 ng/ul
RT: 10.098 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:107 Resp: 646150
Ion Ratio Lower Upper
107 100
121 48.6 39.6 59.4
122 85.0 69.0 103.6





#27

Bis(2-Chloroethoxy)methane

Concen: 46.901 ng/ul

RT: 10.340 min Scan# 1182

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion: 93 Resp: 819966

Ion Ratio Lower Upper

93 100

95 32.4 26.4 39.6

123 10.7 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 39.205 ng/ul

RT: 10.581 min Scan# 1223

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:162 Resp: 489437


Ion Ratio Lower Upper

162 100

164 62.9 52.6 79.0

98 41.2 33.6 50.4









#30
Naphthalene
Concen: 41.895 ng/ul
RT: 10.992 min Scan# 1293
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:128 Resp: 1807528
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 42.239 ng/ul
RT: 11.098 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:127 Resp: 811399
Ion Ratio Lower Upper
127 100
129 33.1 26.1 39.1









#33
Hexachlorobutadiene
Concen: 32.210 ng/ul
RT: 11.275 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:225 Resp: 247635
Ion Ratio Lower Upper
225 100
223 63.6 51.1 76.7
227 64.8 50.7 76.1



#34
Caprolactam
Concen: 32.962 ng/ul
RT: 11.869 min Scan# 1442
Delta R.T. 0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:113 Resp: 135413
Ion Ratio Lower Upper
113 100
55 215.1 163.3 244.9
56 158.0 120.3 180.5









#35
4-Chloro-3-methylphenol
Concen: 48.271 ng/ul
RT: 12.204 min Scan# 1499
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:107 Resp: 625067
Ion Ratio Lower Upper
107 100
144 24.4 20.2 30.4
142 77.2 62.6 93.8



#36
2-Methylnaphthalene
Concen: 42.084 ng/ul
RT: 12.586 min Scan# 1564
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:142 Resp: 1253908
Ion Ratio Lower Upper
142 100
141 89.0 72.2 108.4





#37

1-Methylnaphthalene

Concen: 42.126 ng/ul

RT: 12.804 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:142 Resp: 1262993

Ion Ratio Lower Upper

142 100

141 92.6 73.3 109.9

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 33.914 ng/ul

RT: 12.951 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Tgt Ion:216 Resp: 523559




Ion Ratio Lower Upper


216 100

214 79.4 60.2 90.4

179 23.1 18.6 27.8

108 23.3 18.3 27.5









#40
Hexachlorocyclopentadiene
Concen: 34.559 ng/ul
RT: 12.928 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:237 Resp: 298267
Ion Ratio Lower Upper
237 100
235 62.6 49.1 73.7
272 11.8 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 37.777 ng/ul
RT: 13.181 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:196 Resp: 376475
Ion Ratio Lower Upper
196 100
198 95.1 76.9 115.3
200 30.2 24.1 36.1





#42

2,4,5-Trichlorophenol

Concen: 38.141 ng/ul

RT: 13.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:196 Resp: 409282

Ion Ratio Lower Upper

196 100

198 94.3 75.4 113.2

200 30.0 23.7 35.5



#43

1,1'-Biphenyl

Concen: 41.986 ng/ul

RT: 13.581 min Scan# 1733

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:154 Resp: 1605152


Ion Ratio Lower Upper

154 100

153 40.5 32.1 48.1

76 15.9 12.4 18.6









#44
2-Chloronaphthalene
Concen: 41.278 ng/ul
RT: 13.628 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:162 Resp: 1268056
Ion Ratio Lower Upper
162 100
164 32.3 26.5 39.7
127 41.1 32.4 48.6



#45
2-Nitroaniline
Concen: 61.722 ng/ul
RT: 13.828 min Scan# 1775
Delta R.T. 0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion: 65 Resp: 499962
Ion Ratio Lower Upper
65 100
92 62.5 48.6 73.0
138 90.5 67.0 100.4





#47

Dimethylphthalate

Concen: 42.167 ng/ul

RT: 14.198 min Scan# 1838

Delta R.T. 0.006 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:163 Resp: 1597415

Ion Ratio Lower Upper

163 100

194 3.9 2.9 4.3

164 9.7 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 43.604 ng/ul

RT: 14.316 min Scan# 1858

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:165 Resp: 323440


Ion Ratio Lower Upper

165 100

89 71.5 56.4 84.6

121 24.2 18.9 28.3









#50
Acenaphthylene
Concen: 43.748 ng/ul
RT: 14.469 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:152 Resp: 2044528
Ion Ratio Lower Upper
152 100
151 20.2 16.3 24.5
153 13.0 10.3 15.5



#51
3-Nitroaniline
Concen: 48.449 ng/ul
RT: 14.639 min Scan# 1913
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:138 Resp: 387814
Ion Ratio Lower Upper
138 100
108 10.4 8.4 12.6
92 128.9 104.0 156.0









#52
Acenaphthene
Concen: 43.233 ng/ul
RT: 14.804 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:153 Resp: 1437662
Ion Ratio Lower Upper
153 100
152 48.2 38.4 57.6
154 87.3 70.7 106.1



#53
2,4-Dinitrophenol
Concen: 34.187 ng/ul
RT: 14.839 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:184 Resp: 170285
Ion Ratio Lower Upper
184 100
63 95.6 74.9 112.3
154 70.3 57.8 86.6









#55
4-Nitrophenol
Concen: 57.007 ng/ul
RT: 14.939 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:109 Resp: 298008
Ion Ratio Lower Upper
109 100
139 127.7 102.4 153.6
65 146.4 119.6 179.4



#56
Dibenzofuran
Concen: 41.492 ng/ul
RT: 15.139 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:168 Resp: 1863070
Ion Ratio Lower Upper
168 100
139 40.2 32.2 48.2





#57

2,4-Dinitrotoluene

Concen: 43.506 ng/ul

RT: 15.092 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:165 Resp: 472854

Ion Ratio Lower Upper

165 100

63 59.7 46.6 70.0

182 2.4 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 35.699 ng/ul

RT: 15.357 min Scan# 2035

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Tgt Ion:232 Resp: 333507




Ion Ratio Lower Upper


232 100

131 61.3 50.2 75.4

130 3.3 2.8 4.2

166 38.9 31.2 46.8









#59
Diethylphthalate
Concen: 45.324 ng/ul
RT: 15.551 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:149 Resp: 1710184
Ion Ratio Lower Upper
149 100
177 20.4 16.1 24.1
150 12.0 9.4 14.0



#61
Fluorene
Concen: 42.623 ng/ul
RT: 15.780 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:166 Resp: 1618864
Ion Ratio Lower Upper
166 100
165 97.7 79.3 118.9
167 13.1 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 37.347 ng/ul
RT: 15.775 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:204 Resp: 697094
Ion Ratio Lower Upper
204 100
206 32.6 25.8 38.8
141 67.0 54.3 81.5



#63
4-Nitroaniline
Concen: 51.368 ng/ul
RT: 15.798 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:138 Resp: 382281
Ion Ratio Lower Upper
138 100
92 58.3 43.7 65.5
108 88.4 68.1 102.1









#66
4,6-Dinitro-2-methylphenol
Concen: 33.843 ng/ul
RT: 15.851 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:198 Resp: 239814
Ion Ratio Lower Upper
198 100
182 4.0 4.0 6.0
77 37.4 27.9 41.9



#67
N-Nitrosodiphenylamine
Concen: 43.597 ng/ul
RT: 15.986 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:169 Resp: 1361065
Ion Ratio Lower Upper
169 100
168 68.5 54.0 81.0
167 37.2 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 30.415 ng/ul
RT: 16.663 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:248 Resp: 340765
Ion Ratio Lower Upper
248 100
250 98.2 78.5 117.7
141 99.9 78.9 118.3



#69
Hexachlorobenzene
Concen: 32.412 ng/ul
RT: 16.780 min Scan# 2277
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:284 Resp: 443185
Ion Ratio Lower Upper
284 100
142 49.9 40.4 60.6
249 31.4 24.5 36.7





#70

Atrazine

Concen: 40.757 ng/ul

RT: 16.927 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:200 Resp: 450453

Ion Ratio Lower Upper

200 100

173 27.7 22.6 34.0

215 46.2 36.6 55.0



#71

Pentachlorophenol

Concen: 33.538 ng/ul

RT: 17.121 min Scan# 2335

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:266 Resp: 280628


Ion Ratio Lower Upper

266 100

264 61.5 52.0 78.0

268 63.9 52.2 78.2





#72

Phenanthrene

Concen: 43.281 ng/ul

RT: 17.521 min Scan# 2403

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:178 Resp: 2548620

Ion Ratio Lower Upper

178 100

179 15.0 12.0 18.0

176 19.6 15.4 23.2



#74

Anthracene

Concen: 44.237 ng/ul

RT: 17.610 min Scan# 2418

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:178 Resp: 2620944


Ion Ratio Lower Upper

178 100

179 15.6 12.6 18.8

176 19.3 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 34.679 ng/uL

RT: 13.551 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:216 Resp: 529241

Ion Ratio Lower Upper

216 100

214 78.6 63.3 94.9

218 48.4 38.6 58.0



#76

Pentachlorobenzene

Concen: 33.243 ng/uL

RT: 15.057 min Scan# 1984

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:250 Resp: 552015


Ion Ratio Lower Upper

250 100

248 63.4 50.1 75.1

252 64.0 49.9 74.9





#77

Carbazole

Concen: 45.949 ng/ul

RT: 17.874 min Scan# 2463

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:167 Resp: 2456018

Ion Ratio Lower Upper

167 100

166 21.9 17.5 26.3

139 13.7 10.9 16.3



#78

Di-n-butylphthalate

Concen: 51.585 ng/ul

RT: 18.427 min Scan# 2557

Delta R.T. -0.006 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:149 Resp: 3176614


Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.0

104 5.8 4.4 6.6





#80

Fluoranthene

Concen: 43.027 ng/ul

RT: 19.521 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:202 Resp: 2871218

Ion Ratio Lower Upper

202 100

101 13.4 10.6 16.0

100 10.5 8.2 12.4



#82

Pyrene

Concen: 43.550 ng/ul

RT: 19.886 min Scan# 2805

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:202 Resp: 3061548


Ion Ratio Lower Upper

202 100

101 14.8 12.2 18.4

100 12.2 10.2 15.2









#83
Butylbenzylphthalate
Concen: 53.515 ng/ul
RT: 20.762 min Scan# 2954
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:149 Resp: 1461178
Ion Ratio Lower Upper
149 100
91 81.9 65.9 98.9
206 17.0 13.4 20.2



#84
3,3'-Dichlorobenzidine
Concen: 37.351 ng/ul
RT: 21.551 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:252 Resp: 898331
Ion Ratio Lower Upper
252 100
254 65.5 51.0 76.4
126 15.3 12.2 18.4









#85
Benzo(a)anthracene
Concen: 43.997 ng/ul
RT: 21.621 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:228 Resp: 2930189
Ion Ratio Lower Upper
228 100
229 19.9 15.6 23.4
226 28.0 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 57.316 ng/ul
RT: 21.533 min Scan# 3085
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:149 Resp: 2182161
Ion Ratio Lower Upper
149 100
167 26.3 20.2 30.4
279 3.0 2.2 3.4





#87

Chrysene

Concen: 43.698 ng/ul

RT: 21.680 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:228 Resp: 2738347

Ion Ratio Lower Upper

228 100

226 30.7 24.2 36.4

229 19.8 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 50.599 ng/ul




RT: 22.492 min Scan# 3248


Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Tgt Ion:149 Resp: 3727981









#90
Benzo(b)fluoranthene
Concen: 40.502 ng/ul
RT: 23.368 min Scan# 3397
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39




Instrument :
BNA_M
ClientSampleId :
SSTD040043


Tgt Ion:252 Resp: 2853228
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 13.0 11.0 16.4



#91
Benzo(k)fluoranthene
Concen: 41.119 ng/ul
RT: 23.421 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BM037685.D
Acq: 23 Nov 2022 21:39

Tgt Ion:252 Resp: 2836121
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 12.7 10.6 16.0





#93

Benzo(a)pyrene

Concen: 42.635 ng/ul

RT: 24.021 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39

Instrument :

BNA_M

ClientSampleId :

SSTD040043





Tgt Ion:252 Resp: 2511930

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 14.9 12.0 18.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 46.053 ng/ul

RT: 26.744 min Scan# 3971

Delta R.T. 0.000 min

Lab File: BM037685.D

Acq: 23 Nov 2022 21:39




Tgt Ion:276 Resp: 2903831


Ion Ratio Lower Upper

276 100

138 32.8 26.9 40.3

277 24.7 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 45.979 ng/ul
 RT: 26.768 min Scan# 3974
 Delta R.T. 0.006 min
 Lab File: BM037685.D
 Acq: 23 Nov 2022 21:39

Instrument :
 BNA_M
 ClientSampleId :
 SST040043

Tgt Ion:278 Resp: 2607147
 Ion Ratio Lower Upper
 278 100
 139 20.5 16.6 25.0
 279 23.1 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 46.759 ng/ul
 RT: 27.556 min Scan# 4109
 Delta R.T. 0.006 min
 Lab File: BM037685.D
 Acq: 23 Nov 2022 21:39

Tgt Ion:276 Resp: 2526145
 Ion Ratio Lower Upper
 276 100
 138 28.3 23.4 35.2
 277 23.7 19.1 28.7

