



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112322\
 Data File : BM037681.D
 Acq On : 23 Nov 2022 19:11
 Operator : CG/JU
 Sample : DFTPP085
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP085

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM112322.M
 Title : SVOA CALIBRATION
 Last Update : Thu Nov 24 01:27:15 2022



AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.3	233298	PASS
68	69	0.00	2	1.5	3282	PASS
69	198	0.00	100	49.1	218917	PASS
70	69	0.00	2	0.5	1014	PASS
127	198	10	80	54.6	243307	PASS
197	198	0.00	2	0.4	1791	PASS
198	198	100	100	100.0	445803	PASS
199	198	5	9	6.8	30325	PASS
275	198	10	60	23.3	103733	PASS
365	198	1	100	3.1	13848	PASS
441	198	0.01	100	9.9	44077	PASS
442	198	50	100	62.1	276992	PASS
443	442	15	24	18.7	51763	PASS

m/z	Abundance
36.00	373.0
37.00	352.0
38.00	516.0
39.00	1458.0
40.00	3656.0
41.10	3960.0
42.00	3003.0
43.10	12372.0
44.00	19664.0
45.00	8518.0
46.10	247.0
47.00	324.0
48.90	815.0
50.00	400.0
51.10	643.0
52.00	354.0
53.10	410.0
54.10	262.0
55.10	1600.0
56.00	1812.0
57.00	1335.0
57.90	269.0
59.00	212.0
60.00	4666.0
61.00	328.0
62.90	272.0
64.00	155.0
65.10	470.0
66.00	203.0
67.00	343.0
68.00	380.0
69.00	496.0
69.90	280.0
71.00	518.0
73.10	1871.0
74.00	338.0
75.00	522.0
76.00	199.0
77.00	772.0
78.00	727.0
79.10	411.0
80.00	195.0
81.10	374.0
82.00	265.0
83.00	368.0
83.90	579.0
85.00	492.0
85.90	281.0
89.10	290.0
91.00	653.0
92.90	226.0
95.00	390.0
96.00	1238.0
97.00	435.0
98.00	198.0
99.00	284.0
101.20	172.0
102.00	195.0
103.00	329.0
105.10	533.0
106.10	159.0
109.10	290.0
111.00	256.0
112.00	175.0
114.90	262.0
116.00	207.0
116.90	238.0
118.90	373.0
120.90	202.0
122.00	173.0
123.10	517.0
123.90	275.0
125.00	158.0
125.90	273.0
128.00	361.0
128.90	256.0
130.80	186.0
133.00	534.0
133.80	155.0
135.10	408.0
137.00	158.0
140.90	241.0
143.00	171.0
144.90	198.0
147.00	416.0
148.10	160.0
149.10	432.0
150.00	183.0
151.00	192.0
152.00	182.0
155.10	211.0
164.90	176.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

167.10	229.0
178.10	230.0
178.90	237.0
179.90	206.0
181.00	164.0
190.90	529.0
192.90	633.0
194.00	204.0
207.00	4769.0
208.00	888.0
209.00	549.0
221.00	389.0
267.00	247.0
281.00	889.0
281.80	183.0
283.10	191.0
355.00	201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

m/z	Abundance
35.90	548.0
37.10	1740.0
38.10	3273.0
39.10	21992.0
40.00	4248.0
41.10	4307.0
42.10	2868.0
43.00	11140.0
44.00	17368.0
45.00	8912.0
45.90	397.0
47.00	349.0
47.90	245.0
49.00	2135.0
50.00	65760.0
51.10	257664.0
52.10	13966.0
53.00	681.0
53.90	345.0
55.10	2928.0
56.00	8956.0
57.00	18280.0
58.00	982.0
59.00	358.0
60.00	4646.0
61.00	3329.0
62.00	3744.0
63.10	9232.0
64.00	1585.0
65.10	4703.0
66.00	524.0
67.00	704.0
68.00	4257.0
69.00	236608.0
70.00	1315.0
71.10	836.0
72.00	153.0
73.10	3704.0
74.10	23368.0
75.00	38504.0
76.00	12767.0
77.10	273984.0
78.10	19992.0
79.00	17880.0
80.00	12924.0
81.00	19832.0
82.00	4861.0
83.00	4852.0
84.10	994.0
85.00	3633.0
86.00	5338.0
87.00	2680.0
87.90	1283.0
89.10	504.0
89.70	154.0
91.00	4239.0
92.00	4641.0
93.00	30344.0
94.00	2045.0
95.00	550.0
96.00	2064.0
97.10	747.0
98.00	23192.0
99.00	18488.0
100.00	1969.0
101.00	9998.0
101.90	717.0
103.00	3588.0
104.00	6146.0
105.00	6336.0
106.00	2030.0
107.00	77840.0
108.00	11361.0
109.00	1947.0
110.00	133760.0
111.00	20848.0
112.00	2636.0
112.90	742.0
114.00	273.0
115.00	795.0
116.00	4006.0
117.00	70704.0
118.00	4883.0
119.00	1048.0
120.00	1008.0
121.10	398.0
122.00	5394.0
123.00	8220.0
124.00	3543.0
125.00	3590.0
126.10	676.0
127.00	253440.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

128.00	18848.0
129.00	96888.0
130.00	8180.0
131.00	1806.0
132.00	1004.0
132.90	1046.0
134.00	2560.0
135.00	8136.0
136.00	3265.0
137.00	4824.0
138.00	977.0
139.00	487.0
140.00	1466.0
141.00	12646.0
142.10	4335.0
143.00	3139.0
144.00	649.0
145.00	856.0
146.00	2371.0
147.00	6425.0
148.00	16128.0
149.00	4457.0
150.00	1016.0
151.00	1474.0
152.10	912.0
153.00	4219.0
154.00	3268.0
155.10	7091.0
156.00	10282.0
157.00	1964.0
158.00	1867.0
159.00	1746.0
160.00	3708.0
161.00	5329.0
162.00	1854.0
163.00	989.0
164.00	712.0
165.00	4272.0
166.10	3355.0
167.00	26808.0
168.00	16656.0
169.00	2683.0
170.00	708.0
170.90	1078.0
171.90	2243.0
173.00	2984.0
174.00	5263.0
175.00	9403.0
176.10	2566.0
177.00	4832.0
178.00	1495.0
179.00	18200.0
180.00	14366.0
181.00	6158.0
182.00	765.0
182.90	438.0
184.00	1418.0
185.00	9074.0
186.00	63800.0
187.00	18136.0
188.00	2077.0
189.00	4560.0
190.00	786.0
191.00	2688.0
192.00	5288.0
193.00	6630.0
194.00	1614.0
195.00	711.0
196.00	11973.0
197.00	1981.0
198.00	426112.0
199.00	30376.0
200.00	2373.0
201.60	2001.0
203.00	2908.0
204.00	17520.0
205.00	27816.0
206.00	113792.0
207.00	20648.0
208.00	5225.0
209.00	1880.0
210.00	1912.0
211.00	4750.0
212.00	924.0
213.10	200.0
215.00	1710.0
216.00	2719.0
217.00	30568.0
218.00	4266.0
219.00	437.0
221.00	15193.0
222.00	4475.0
223.00	6699.0
224.00	58688.0
225.00	14558.0
226.00	1895.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

227.00	27600.0
228.00	3772.0
229.00	5713.0
229.90	636.0
231.00	1872.0
232.00	328.0
233.00	357.0
234.00	1615.0
235.00	1637.0
236.00	1377.0
237.00	2541.0
237.90	345.0
238.90	1044.0
240.00	848.0
241.00	1277.0
242.00	2774.0
243.00	3653.0
244.10	45536.0
245.10	6104.0
246.00	9423.0
247.00	1877.0
248.00	582.0
249.00	1858.0
249.80	491.0
251.00	536.0
251.90	360.0
253.00	860.0
254.00	1213.0
255.00	225216.0
256.00	33200.0
257.00	2895.0
258.00	14052.0
259.00	2232.0
260.00	496.0
260.90	412.0
263.90	308.0
265.00	5877.0
266.00	936.0
267.00	476.0
269.90	317.0
271.00	722.0
272.00	745.0
273.00	6995.0
274.00	17056.0
275.00	90112.0
276.00	12054.0
277.00	8130.0
278.00	1778.0
278.90	387.0
281.00	1300.0
282.10	347.0
283.00	1251.0
283.90	627.0
285.00	1325.0
286.00	269.0
288.10	176.0
289.00	223.0
290.00	240.0
291.00	280.0
292.00	703.0
293.00	1899.0
294.00	532.0
295.10	463.0
296.00	30072.0
297.00	4183.0
297.90	212.0
300.90	415.0
302.00	423.0
303.10	2632.0
303.90	706.0
305.10	229.0
308.00	504.0
309.00	214.0
309.90	277.0
310.90	152.0
313.00	311.0
314.00	928.0
315.00	3009.0
316.10	1490.0
316.90	291.0
321.00	807.0
322.00	371.0
323.00	7879.0
324.00	1378.0
325.00	273.0
326.90	1433.0
328.00	787.0
328.80	185.0
332.00	614.0
333.00	676.0
334.00	4405.0
335.00	1235.0
336.10	157.0
339.00	191.0
341.00	784.0
341.90	394.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

346.00	1416.0
347.00	287.0
352.00	2135.0
353.10	1454.0
354.00	1998.0
355.10	565.0
356.00	156.0
365.00	11635.0
366.00	1568.0
370.00	291.0
371.00	467.0
372.10	3797.0
373.00	1199.0
383.00	1258.0
384.00	401.0
390.00	466.0
391.00	311.0
392.10	327.0
401.00	273.0
402.10	1375.0
403.00	1562.0
404.00	680.0
405.00	198.0
421.00	1773.0
422.00	1412.0
423.10	10859.0
424.10	2997.0
425.00	347.0
441.10	32952.0
442.10	203136.0
443.10	38440.0
444.10	4057.0
445.00	196.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

m/z	Abundance
36.10	541.0
37.00	1245.0
38.10	3891.0
39.10	22848.0
40.00	4046.0
41.10	4052.0
42.00	2715.0
43.10	11621.0
44.00	17520.0
45.00	9004.0
46.00	412.0
46.90	366.0
47.90	198.0
49.00	2144.0
50.10	68904.0
51.10	275776.0
52.10	14917.0
53.10	1090.0
54.10	244.0
55.10	2745.0
56.00	9918.0
57.00	18888.0
58.00	1225.0
59.00	580.0
60.00	4911.0
61.00	3607.0
62.10	3714.0
63.00	9898.0
64.00	1760.0
65.00	5564.0
66.10	563.0
67.10	764.0
68.10	4145.0
69.00	258816.0
70.00	1458.0
71.00	623.0
71.90	244.0
73.10	3868.0
74.10	26560.0
75.00	41424.0
76.10	14490.0
77.10	294848.0
78.00	20728.0
79.00	19408.0
80.00	14945.0
81.00	21640.0
82.00	5543.0
83.00	5040.0
84.00	1353.0
85.00	3577.0
86.00	6459.0
87.00	2842.0
88.00	1195.0
88.90	730.0
89.80	156.0
91.00	4487.0
92.00	5424.0
93.00	34384.0
94.00	2584.0
94.90	735.0
96.00	2190.0
97.20	861.0
98.00	26232.0
99.00	21512.0
100.00	1576.0
101.00	12356.0
102.00	626.0
103.00	3316.0
104.00	6901.0
105.00	6473.0
106.10	1675.0
107.00	86904.0
108.00	13583.0
109.00	2175.0
110.00	153088.0
111.00	23656.0
112.00	3333.0
113.00	1245.0
114.00	423.0
114.90	829.0
116.00	4100.0
117.00	80568.0
118.00	6381.0
119.00	896.0
120.00	1109.0
120.90	451.0
122.00	5583.0
123.00	9328.0
124.00	4280.0
125.00	3637.0
126.10	1183.0
127.00	289600.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

128.00	21464.0
129.00	114400.0
130.00	9241.0
131.00	2047.0
132.00	737.0
133.00	1011.0
134.00	2976.0
135.00	9594.0
136.00	4414.0
137.00	4840.0
137.90	1262.0
139.00	795.0
140.10	1271.0
141.00	14558.0
142.00	4921.0
143.00	3455.0
144.00	845.0
145.00	908.0
146.00	2717.0
147.00	8002.0
148.00	19392.0
149.00	4810.0
150.00	969.0
151.10	1812.0
152.10	1167.0
153.00	4781.0
154.00	3470.0
155.00	8381.0
156.10	12020.0
157.00	1974.0
158.00	2830.0
159.00	2314.0
160.00	4651.0
161.00	7402.0
162.00	2166.0
162.90	606.0
164.00	760.0
165.00	5403.0
166.00	4348.0
167.00	32784.0
168.00	19752.0
169.00	2456.0
170.00	1192.0
171.00	1705.0
172.00	3172.0
173.00	3624.0
174.00	6199.0
175.00	11546.0
176.00	3391.0
177.00	5227.0
178.00	2053.0
179.00	22624.0
180.00	15533.0
181.00	6706.0
182.00	1388.0
183.00	698.0
184.00	1645.0
185.00	10596.0
186.00	79600.0
187.00	23544.0
188.00	2602.0
189.00	5646.0
190.00	916.0
191.00	2680.0
192.00	7501.0
193.00	7859.0
194.00	1833.0
195.00	814.0
196.00	14898.0
196.80	2326.0
198.00	542336.0
199.00	36904.0
200.00	2790.0
201.50	2559.0
203.00	4628.0
204.00	21760.0
205.00	35888.0
206.00	145792.0
207.00	24440.0
208.00	6086.0
209.00	2492.0
210.30	2155.0
211.00	6373.0
212.00	977.0
213.00	309.0
215.00	2062.0
216.00	3258.0
217.00	42904.0
218.00	4901.0
219.00	476.0
221.00	19080.0
222.00	4743.0
223.00	9198.0
224.00	82408.0
225.10	20176.0
226.10	1908.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

227.00	38080.0
228.00	6209.0
229.00	7352.0
230.00	967.0
231.00	3011.0
231.90	532.0
233.00	780.0
234.00	1901.0
235.00	2381.0
236.00	1597.0
237.00	3104.0
237.80	529.0
239.00	1794.0
240.00	1033.0
240.90	2076.0
242.00	4157.0
243.00	4662.0
244.00	61520.0
245.00	8425.0
246.00	12999.0
247.00	2615.0
248.00	619.0
249.00	2050.0
250.00	503.0
251.00	679.0
252.00	427.0
253.10	1382.0
254.10	1327.0
255.00	307072.0
256.00	44536.0
257.00	3828.0
258.00	20032.0
259.00	3080.0
259.90	534.0
260.90	434.0
261.90	207.0
263.00	168.0
264.00	537.0
265.00	8422.0
266.00	1041.0
267.00	403.0
268.10	188.0
270.00	357.0
271.00	929.0
272.00	725.0
273.00	9246.0
274.00	24104.0
275.00	126512.0
276.00	16584.0
277.00	11245.0
278.00	2342.0
278.90	399.0
281.00	1050.0
282.00	592.0
283.00	1412.0
284.00	773.0
285.00	1935.0
286.10	456.0
287.90	231.0
288.90	477.0
290.10	410.0
291.00	377.0
291.90	450.0
293.00	2818.0
294.00	732.0
295.00	752.0
296.00	42104.0
297.00	5161.0
297.90	524.0
299.00	167.0
300.90	467.0
302.00	678.0
303.10	3973.0
304.00	1214.0
308.00	335.0
308.90	255.0
310.00	470.0
313.00	440.0
314.00	2003.0
315.00	4300.0
316.10	2649.0
316.90	377.0
320.10	157.0
321.00	1261.0
322.10	508.0
323.00	10990.0
324.10	2376.0
325.00	317.0
325.90	294.0
327.00	2221.0
328.00	1014.0
328.90	242.0
332.00	878.0
333.00	806.0
334.00	6952.0
335.00	1848.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

335.90	174.0
339.10	183.0
341.10	1510.0
342.10	540.0
346.00	2641.0
347.00	513.0
350.80	191.0
352.10	3088.0
353.10	2322.0
354.10	3460.0
355.10	1092.0
358.90	258.0
365.00	16113.0
366.00	2455.0
367.00	212.0
370.00	430.0
371.10	905.0
372.00	5085.0
373.10	1162.0
374.20	164.0
383.00	1323.0
384.00	472.0
390.00	562.0
391.00	545.0
392.10	262.0
401.00	280.0
402.00	2341.0
403.10	3036.0
404.10	1097.0
405.00	243.0
414.90	160.0
421.00	2551.0
422.10	2199.0
423.10	16400.0
424.10	4402.0
425.00	514.0
441.10	53008.0
442.10	329280.0
443.10	61920.0
444.10	6058.0
445.10	272.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

m/z	Abundance
35.90	470.0
37.00	1210.0
38.00	2329.0
39.10	14478.0
40.00	4158.0
41.00	3721.0
42.00	2680.0
43.10	12227.0
44.00	19208.0
45.00	9263.0
46.00	379.0
47.00	329.0
49.00	1659.0
50.10	44336.0
51.10	168384.0
52.10	9072.0
53.10	853.0
54.00	340.0
55.00	2550.0
56.00	6608.0
57.00	13355.0
58.00	703.0
59.00	283.0
60.10	4832.0
61.00	2053.0
62.00	2282.0
63.10	7417.0
64.00	997.0
65.10	3761.0
66.00	363.0
67.00	551.0
68.00	2585.0
69.00	162816.0
70.00	1110.0
71.10	725.0
73.00	2967.0
74.10	16126.0
75.00	26600.0
76.10	7832.0
77.10	186688.0
78.10	13421.0
79.00	12558.0
80.00	8960.0
81.00	13369.0
82.00	3170.0
83.00	3017.0
83.90	668.0
85.00	2352.0
86.00	3872.0
87.00	1998.0
87.90	706.0
88.90	590.0
91.00	2981.0
92.00	3114.0
93.00	21144.0
94.10	1611.0
95.00	672.0
96.00	1479.0
97.10	722.0
98.00	16672.0
99.00	13085.0
100.10	1056.0
101.00	7160.0
101.90	261.0
103.00	2342.0
104.00	4279.0
105.00	4992.0
106.00	1556.0
107.00	55320.0
108.00	8345.0
109.00	1483.0
110.00	96976.0
111.00	14023.0
112.00	2201.0
113.10	704.0
113.90	393.0
115.00	872.0
116.00	2870.0
117.00	55152.0
118.00	3763.0
119.00	871.0
120.00	1052.0
121.10	412.0
122.00	4055.0
123.00	5989.0
124.00	2825.0
125.00	2406.0
126.00	616.0
127.00	186880.0
128.00	13436.0
129.00	75224.0
130.00	6808.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

131.00	1272.0
132.00	780.0
132.90	873.0
134.00	1990.0
135.00	5490.0
136.00	2493.0
137.00	3461.0
137.90	816.0
139.00	515.0
140.00	1327.0
141.00	9629.0
142.00	3567.0
143.00	2586.0
143.90	600.0
145.00	562.0
146.00	2103.0
147.00	5945.0
148.00	13628.0
149.00	3586.0
150.00	778.0
151.00	1100.0
151.90	651.0
153.00	3509.0
153.90	2431.0
155.00	5079.0
156.00	7974.0
157.10	1774.0
158.00	1416.0
159.00	1417.0
160.00	3570.0
161.00	4885.0
162.00	1960.0
163.00	521.0
164.00	555.0
165.00	3964.0
166.00	3029.0
167.00	23440.0
168.00	12664.0
169.00	2305.0
170.10	795.0
171.00	1043.0
172.00	2173.0
173.00	2508.0
174.00	4094.0
175.00	7955.0
176.00	2393.0
177.00	4009.0
178.00	1468.0
179.00	15241.0
180.00	10896.0
181.00	4945.0
182.10	770.0
182.90	461.0
184.00	1045.0
185.00	7399.0
186.00	55280.0
187.00	14758.0
188.00	1666.0
189.00	3328.0
190.00	574.0
191.00	2473.0
192.00	5231.0
193.00	6528.0
194.00	1546.0
195.00	963.0
196.00	11656.0
196.80	1067.0
198.00	368960.0
199.00	23696.0
200.00	2552.0
201.50	1535.0
203.00	2912.0
204.00	14126.0
205.00	25760.0
206.00	98352.0
207.00	17808.0
208.00	4867.0
209.00	1771.0
210.00	1751.0
211.00	4360.0
213.00	331.0
215.10	1064.0
216.00	2460.0
217.00	27968.0
218.00	3464.0
219.10	166.0
220.20	671.0
221.00	12501.0
221.90	3569.0
223.00	6082.0
224.00	57016.0
225.00	14565.0
226.00	1500.0
227.00	26648.0
228.00	3519.0
229.00	5409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

229.90	687.0
231.00	2038.0
232.00	368.0
233.00	623.0
234.00	1472.0
235.00	1836.0
236.10	1204.0
237.00	2347.0
238.00	357.0
239.00	1107.0
240.00	553.0
241.00	1315.0
242.00	3105.0
243.00	3392.0
244.00	44624.0
245.00	5480.0
246.00	9120.0
247.00	1678.0
247.90	363.0
248.90	1360.0
250.00	349.0
250.90	324.0
252.00	433.0
252.90	900.0
253.90	1104.0
255.00	222912.0
256.00	31384.0
257.00	2899.0
258.00	14993.0
259.10	2102.0
260.00	339.0
261.00	363.0
263.00	168.0
264.00	372.0
265.00	5886.0
266.00	858.0
266.90	439.0
269.90	252.0
271.00	829.0
272.10	553.0
273.00	6550.0
274.00	16968.0
275.00	94576.0
276.00	11917.0
277.00	9249.0
278.00	1856.0
279.00	320.0
281.00	1339.0
282.00	553.0
283.00	1391.0
284.00	624.0
285.00	1620.0
286.10	377.0
288.00	219.0
289.10	406.0
289.90	392.0
290.90	325.0
292.10	492.0
293.00	1939.0
294.00	641.0
295.00	800.0
296.00	30848.0
297.00	4215.0
301.00	373.0
302.00	538.0
303.10	3383.0
304.00	1128.0
305.00	158.0
308.00	333.0
308.90	280.0
310.10	293.0
312.90	279.0
314.00	1336.0
315.00	3372.0
316.00	1717.0
316.90	291.0
321.00	962.0
321.90	346.0
323.10	9197.0
324.00	1773.0
324.90	203.0
326.00	295.0
327.00	1585.0
328.10	790.0
332.00	602.0
333.00	856.0
334.10	5761.0
335.00	1481.0
336.00	280.0
341.00	1029.0
342.00	193.0
346.00	2212.0
347.00	439.0
352.10	2743.0
353.00	1954.0
354.00	2659.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085

355.10	619.0
359.10	157.0
365.00	13795.0
366.00	1820.0
369.90	253.0
371.00	776.0
372.10	5091.0
373.10	1071.0
374.00	215.0
383.00	1365.0
384.00	323.0
384.90	204.0
390.00	567.0
391.00	308.0
392.10	214.0
401.10	204.0
402.10	1925.0
403.10	2441.0
404.00	992.0
405.10	155.0
421.00	2637.0
422.10	2103.0
423.10	16552.0
424.10	3862.0
425.00	397.0
441.10	46272.0
442.10	298560.0
443.10	54928.0
444.10	5050.0
445.00	204.0

Instrument :
BNA_M
ClientSampleId :
DFTPP085