

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112423\
 Data File : BM043024.D
 Acq On : 25 Nov 2023 20:27
 Operator : MA/JU
 Sample : PB157192BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS192

Quant Time: Nov 27 02:44:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 27 02:36:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.880	152	581	0.400	ng/ul	0.00
4) Naphthalene-d8	10.644	136	1939	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.464	164	1156	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.233	188	2386	0.400	ng/ul	0.00
17) Chrysene-d12	21.415	240	1326	0.400	ng/ul	0.00
23) Perylene-d12	23.776	264	1520	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	987	1.107	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.244	152	1482	0.571	ng/ul	0.00
18) Fluoranthene-d10	19.256	212	2783	0.576	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.197	88	2548	2.661	ng/ul#	82
5) Naphthalene	10.693	128	2525	0.479	ng/ul	98
7) 2-Methylnaphthalene	12.321	142	1581	0.509	ng/ul	97
8) 1-Methylnaphthalene	12.513	142	1613	0.465	ng/ul	99
10) Acenaphthylene	14.200	152	2848	0.483	ng/ul	98
11) Acenaphthene	14.528	153	2212	0.495	ng/ul	98
12) Fluorene	15.532	166	2383	0.519	ng/ul	99
14) Pentachlorophenol	16.887	266	614	0.951	ng/ul	94
15) Phenanthrene	17.275	178	3284	0.482	ng/ul	100
16) Anthracene	17.385	178	3046	0.472	ng/ul	100
19) Fluoranthene	19.289	202	3910	0.520	ng/ul	99
20) Pyrene	19.656	202	4079	0.510	ng/ul	100
21) Benzo(a)anthracene	21.406	228	2199	0.511	ng/ul	98
22) Chrysene	21.453	228	3959	0.494	ng/ul	99
24) Benzo(b)fluoranthene	23.040	252	2961	0.577	ng/ul	99
25) Benzo(k)fluoranthene	23.095	252	4131	0.585	ng/ul	99
26) Benzo(a)pyrene	23.677	252	3134	0.560	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	26.239	276	4326	0.571	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.252	278	3367	0.574	ng/ul	93
29) Benzo(g,h,i)perylene	26.990	276	4143	0.562	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112423\
Data File : BM043024.D
Acq On : 25 Nov 2023 20:27
Operator : MA/JU
Sample : PB157192BS
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS192

Quant Time: Nov 27 02:44:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 27 02:36:19 2023
Response via : Initial Calibration

